



Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 1 0 6 2 8 4 6 4 1 2 0 6 2 3 D



D 5 1 0 1 1 3 7 0 8 4 1 4 1 0 7 6 D



D 6 1 1 3 1 1 5 1 5 5 0 7 4 1 2 3 D



D 7 0 8 2 1 2 3 0 7 8 1 0 5 1 0 0 D



D 8 1 6 6 0 7 4 1 1 1 3 2 6 4 4 1 D



D 9 3 3 0 4 4 7 1 8 0 2 8 0 2 4 0 D



D 0 3 6 0 3 1 7 D

Reference No.:		1300036412	
Lot No:		910628 (rev. 2)	
Expiry date:		2023-06-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,1-13,7
	Prothrombin %	%	84-141
	Prothrombin INR	INR	0,76-1,13
PT Liq	Prothrombin Time	sec	11,5-15,5
	Prothrombin %	%	74-123
	Prothrombin INR	INR	0,82-1,23
PT R	Prothrombin Time	sec	7,8-10,5
	Prothrombin %	%	100-166
	Prothrombin INR	INR	0,74-1,11
APTT	Partial Thromboplastin Time	sec	32,6-44,1
APTT Liq	Partial Thromboplastin Time	sec	33,0-44,7
FIB Cl	Fibrinogen	g/L	1,8-2,8
TT	Thrombin Time	sec	24,0-36,0* *NA for Yumizen G200 and G400 DDi
AT	Antithrombin	%	81-110
Important notice			
<p>The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.</p> <p>The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.</p>			



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4910628564120623D



D 5145217037061126D



D 6189164246035058D



D 7140209100150046D



D 8077098147455616D



D 9454615060150000D



D 0999369D

Reference No.:		1300036412	
Lot No:		910628 (rev. 2)	
Expiry date:		2023-06-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	14,5-21,7
	Prothrombin %	%	37-61
	Prothrombin INR	INR	1,26-1,89
PT Liq	Prothrombin Time	sec	16,4-24,6
	Prothrombin %	%	35-58
	Prothrombin INR	INR	1,40-2,09
PT R	Prothrombin Time	sec	10,0-15,0
	Prothrombin %	%	46-77
	Prothrombin INR	INR	0,98-1,47
APTT	Partial Thromboplastin Time	sec	45,5-61,6
APTT Liq	Partial Thromboplastin Time	sec	45,4-61,5
FIB Cl	Fibrinogen	g/L	0,6-1,5
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	40-66

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.