



# Yumizen G CTRL I (level 1) for Yumizen G Line



D4910628464120623D



D5101137084141076D



D6113100135080133D



D7078117078105100D



D8166074111326441D



D9330447180280240D



D0360315D

Reference No.:		1300036412	
Lot No:		910628 (rev. 3)	
Expiry date:		2023-06-30	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,1-13,7
	Prothrombin %	%	84-141
	Prothrombin INR	INR	0,76-1,13
PT Liq (*)	Prothrombin Time	sec	10,0-13,5 (*)
	Prothrombin %	%	80-133 (*)
	Prothrombin INR	INR	0,78-1,17 (*)
PT R	Prothrombin Time	sec	7,8-10,5
	Prothrombin %	%	100-166
	Prothrombin INR	INR	0,74-1,11
APTT	Partial Thromboplastin Time	sec	32,6-44,1
APTT Liq	Partial Thromboplastin Time	sec	33,0-44,7
FIB Cl	Fibrinogen	g/L	1,8-2,8
TT	Thrombin Time	sec	24,0-36,0** **Not applicable for G200 and G400 DDi
AT	Antithrombin	%	81-110

### Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.

(\*) Warning! This range has to be used only with Yumizen G PT Liq 4 lots : 920315 and higher except for the Yumizen G PT Liq 4 lots 920611 & 920705. For the other lots, including 920611 & 920705, please refer to the previous revision of values sheet.



# Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 1 0 6 2 8 5 6 4 1 2 0 6 2 3 D



D 5 1 4 5 2 1 7 0 3 7 0 6 1 1 2 6 D



D 6 1 8 9 1 3 7 2 0 5 0 3 4 0 5 7 D



D 7 1 2 2 1 8 4 1 0 0 1 5 0 0 4 6 D



D 8 0 7 7 0 9 8 1 4 7 4 5 5 6 1 6 D



D 9 4 5 4 6 1 5 0 6 0 1 5 0 0 0 0 D



D 0 9 9 9 3 6 4 D

Reference No.:		1300036412	
Lot No.:		910628 (rev. 3)	
Expiry date:		2023-06-30	
Optical Kit	Yumizen G Line Parameter	Unit	Range
PT	Prothrombin Time	sec	14,5-21,7
	Prothrombin %	%	37-61
	Prothrombin INR	INR	1,26-1,89
PT Liq (*)	Prothrombin Time	sec	13,7-20,5 (*)
	Prothrombin %	%	34-57 (*)
	Prothrombin INR	INR	1,22-1,84 (*)
PT R	Prothrombin Time	sec	10,0-15,0
	Prothrombin %	%	46-77
	Prothrombin INR	INR	0,98-1,47
APTT	Partial Thromboplastin Time	sec	45,5-61,6
APTT Liq	Partial Thromboplastin Time	sec	45,4-61,5
FIB Cl	Fibrinogen	g/L	0,6-1,5
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	40-66

### Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.

(\*) Warning! This range has to be used only with Yumizen G PT Liq 4 lots : 920315 and higher except for the Yumizen G PT Liq 4 lots 920611 & 920705. For the other lots, including 920611 & 920705, please refer to the previous revision of values sheet.