



Yumizen G CTRL I (level 1) for Yumizen G Line



D4910727464120723D



D5105142077128079D



D6119119161069115D



D7084126077104102D



D8170073110333451D



D9315426220320193D



D0290336D

Reference No.:		1300036412	
Lot No:		910727 (rev. 2)	
Expiry date:		2023-07-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	10,5-14,2
	Prothrombin %	%	77-128
	Prothrombin INR	INR	0,79-1,19
PT Liq	Prothrombin Time	sec	11,9-16,1
	Prothrombin %	%	69-115
	Prothrombin INR	INR	0,84-1,26
PT R	Prothrombin Time	sec	7,7-10,4
	Prothrombin %	%	102-170
	Prothrombin INR	INR	0,73-1,10
APTT	Partial Thromboplastin Time	sec	33,3-45,1
APTT Liq	Partial Thromboplastin Time	sec	31,5-42,6
FIB Cl	Fibrinogen	g/L	2,2-3,2
TT	Thrombin Time	sec	19,3-29,0
AT	Antithrombin	%	87-118

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.
 The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 1 0 7 2 7 5 6 4 1 2 0 7 2 3 D



D 5 1 3 9 2 0 9 0 3 9 0 6 4 1 2 0 D



D 6 1 8 1 1 7 1 2 5 7 0 3 1 0 5 2 D



D 7 1 4 3 2 1 4 0 9 7 1 4 5 0 4 7 D



D 8 0 7 8 0 9 9 1 4 8 5 0 5 6 8 3 D



D 9 5 0 3 6 8 0 0 8 0 1 9 0 0 0 0 D



D 0 9 9 9 3 8 0 D

Reference No.:		1300036412	
Lot No:		910727 (rev. 2)	
Expiry date:		2023-07-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	13,9-20,9
	Prothrombin %	%	39-64
	Prothrombin INR	INR	1,20-1,81
PT Liq	Prothrombin Time	sec	17,1-25,7
	Prothrombin %	%	31-52
	Prothrombin INR	INR	1,43-2,14
PT R	Prothrombin Time	sec	9,7-14,5
	Prothrombin %	%	47-78
	Prothrombin INR	INR	0,99-1,48
APTT	Partial Thromboplastin Time	sec	50,5-68,3
APTT Liq	Partial Thromboplastin Time	sec	50,3-68,0
FIB Cl	Fibrinogen	g/L	0,8-1,9
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	44-73

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.