

HORIBA ABX SAS

Parc Euromédecine

Rue du Caducée - BP 7290

34184 Montpellier Cedex 4 - FRANCE

Phone: +33 (0)4 67 14 15 16

Web: www.horiba.com

EU Declaration of Conformity

(N° dc90126ben)

WE THE MANUFACTURER

Name	HORIBA ABX SAS
 Address	Parc Euromédecine Rue du Caducée BP 7290 34184 Montpellier Cedex 4 FRANCE
Single Registration Number	FR-MF-000000320

TAKE SOLE RESPONSIBILITY FOR AND HEREBY DECLARE THAT THE PRODUCT(S)

Device category	Hematology Reagent
Product name	Minilysebio 0.4L
Models	0906020
Basic UDI-DI	361023minilysebioVG
Country of origin	FRANCE

Intended Use

Minilysebio 0.4L is a lysing solution intended for *in vitro* diagnostic use and designed for lysing erythrocytes (RBC) for leucocytes (WBC) counting and differentiation and for hemoglobin determination on HORIBA Medical blood cell counters.

Clinical laboratories use.

MEET(S) THE PROVISIONS OF THE FOLLOWING DIRECTIVES, REGULATIONS, STANDARDS AND COMMON SPECIFICATIONS

Regulations	Regulation (EU) 2017/746 on <i>in vitro</i> diagnostic medical devices Risk Class: A ☒ B ☐ C ☐ D ☐
IVDR conformity assessment route	☒ ANNEX II + ANNEX III + ANNEX IV (<i>Class A devices excluding sterile devices</i>)
Common Specifications	Not applicable

Montpellier, France
2022/09/12

Claire MALLIÉ
Quality & Regulatory Affairs Junior
Director / PRRC

