

**LOT** MX 438 L  
 Rev 1

**CONTROL** L

(Exp.)

**2023-02-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1	2.2			2.1					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.27	2.27			2.25					± 0.17
HB HGB	g/dl	5.8	5.9			5.6					± 0.5
	g/l	58	59			56					± 5
	mmol/l	3.60	3.65			3.47					± 0.31
HT HCT	%	16.1	16.3			16.4					± 2.0
	l/l	0.161	0.163			0.164					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	71	72			73					± 4
TGMH MCH	pg	25.6	25.9			24.8					± 2.0
	fmol	1.59	1.61			1.54					± 0.12
CCMH MCHC	g/dl	36.0	36.0			34.0					± 3.0
	g/l	360	360			340					± 30
	mmol/l	22.35	22.34			21.10					± 1.86
IDR RDW	%	15.5	14.0			15.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	81	85			81					± 22
VPM MPV	µm <sup>3</sup> ; fl	7.9	8.0			7.5					± 2.0
LYMPHO	#	1.43	1.47			1.42					± 0.40
	%	68.0	66.7			67.5					± 8.0
MONO	#	0.17	0.22			0.17					± 0.20
	%	8.0	10.0			8.0					± 6.0
GRANULO	#	0.50	0.52			0.51					± 0.40
	%	24.0	23.5			24.5					± 7.0

**LOT** MX 438 N  
Rev 1

**CONTROL** N

 (Exp.)

**2023-02-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 BP - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.6	7.6			7.5					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.61	4.59			4.57					± 0.26
HB HGB	g/dl	13.4	13.4			13.1					± 0.7
	g/l	134	134			131					± 7
	mmol/l	8.32	8.32			8.15					± 0.43
HT HCT	%	37.3	37.2			38.4					± 2.7
	l/l	0.373	0.372			0.384					± 0.027
VGM MCV	µm <sup>3</sup> ; fl	81	81			84					± 4
TGMH MCH	pg	29.1	29.2			28.7					± 2.0
	fmol	1.81	1.81			1.78					± 0.12
CCMH MCHC	g/dl	35.9	36.0			34.2					± 3.0
	g/l	359	360			342					± 30
	mmol/l	22.28	22.38			21.24					± 1.86
IDR RDW	%	13.5	13.0			14.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	244	257			248					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.2	8.3			7.6					± 2.0
LYMPHO	#	2.66	2.74			2.55					± 0.40
	%	35.0	36.0			34.0					± 5.0
MONO	#	0.46	0.51			0.49					± 0.30
	%	6.0	6.7			6.5					± 4.0
GRANULO	#	4.48	4.36			4.46					± 0.60
	%	59.0	57.3			59.5					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143

**LOT** MX 438 H  
 Rev 1

**CONTROL** H

(Exp.)

**2023-02-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	20.3	20.2			20.2					± 1.7
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	5.70	5.67			5.68					± 0.32
HB HGB	g/dl	18.6	18.2			18.0					± 0.9
	g/l	186	182			180					± 9
	mmol/l	11.55	11.30			11.19					± 0.56
HT HCT	%	50.7	50.5			51.7					± 3.4
	l/l	0.507	0.505			0.517					± 0.034
VGM MCV	µm <sup>3</sup> ; fl	89	89			91					± 4
TGMH MCH	pg	32.6	32.1			31.7					± 2.0
	fmol	2.03	1.99			1.97					± 0.12
CCMH MCHC	g/dl	36.7	36.1			34.9					± 3.0
	g/l	367	361			349					± 30
	mmol/l	22.77	22.40			21.65					± 1.86
IDR RDW	%	13.5	13.0			14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	524	536			528					± 65
VPM MPV	µm <sup>3</sup> ; fl	7.8	7.8			7.2					± 2.0
LYMPHO	#	3.45	3.43			3.43					± 1.00
	%	17.0	17.0			17.0					± 5.0
MONO	#	1.02	1.31			1.01					± 0.80
	%	5.0	6.5			5.0					± 4.0
GRANULO	#	15.83	15.44			15.76					± 1.40
	%	78.0	76.4			78.0					± 7.0