



Yumizen G CTRL I (level 1) for Yumizen G Line



D 4 9 1 1 2 2 5 4 6 4 1 2 1 2 2 3 D



D 5 1 1 9 1 6 1 0 6 2 1 0 4 0 8 7 D



D 6 1 3 1 1 1 1 5 0 6 1 1 0 2 D



D 7 0 8 8 1 3 2 0 9 2 1 2 4 0 6 9 D



D 8 1 1 5 0 8 3 1 2 4 3 3 6 4 5 5 D



D 9 3 2 2 4 3 6 2 0 0 3 0 0 1 8 6 D



D 0 2 7 8 3 1 8 D

Reference No.:		1300036412	
Lot No:		911225 (rev. 5)	
Expiry date:		2023-12-31	
Optical	Yumizen G Line		
Kit	Parameter	Unit	Range
PT	Prothrombin Time	sec	11,9-16,1
	Prothrombin %	%	62-104
	Prothrombin INR	INR	0,87-1,31
PT Liq (*)	Prothrombin Time	sec	11,1-15,0 (*)
	Prothrombin %	%	61-102 (*)
	Prothrombin INR	INR	0,88-1,32 (*)
PT R	Prothrombin Time	sec	9,2-12,4
	Prothrombin %	%	69-115
	Prothrombin INR	INR	0,83-1,24
APTT	Partial Thromboplastin Time	sec	33,6-45,5
APTT Liq	Partial Thromboplastin Time	sec	32,2-43,6
FIB Cl	Fibrinogen	g/L	2,0-3,0
TT	Thrombin Time	sec	18,6-27,8
AT	Antithrombin	%	86-116

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.

(*) Warning! This range has to be used only with Yumizen G PT Liq 4 lots : 920315 and higher except for the Yumizen G PT Liq 4 lots 920611 & 920705. For the other lots, including 920611 & 920705, please refer to the previous revision of values sheet.



Yumizen G CTRL II (level 2) for Yumizen G Line



D 4 9 1 1 2 2 5 5 6 4 1 2 1 2 2 3 D



D 5 1 4 3 2 1 4 0 4 1 0 6 9 1 1 4 D



D 6 1 7 1 1 4 8 2 2 1 0 3 0 0 5 0 D



D 7 1 3 4 2 0 1 1 1 1 1 6 7 0 4 2 D



D 8 0 7 0 1 0 6 1 5 9 5 1 7 7 0 0 D



D 9 4 9 6 6 7 1 0 8 0 1 9 0 0 0 0 D



D 0 9 9 9 3 2 2 D

Reference No.:		1300036412	
Lot No:		911225 (rev. 5)	
Expiry date:		2023-12-31	
Optical Kit	Yumizen G Line Parameter	Unit	Range
PT	Prothrombin Time	sec	14,3-21,4
	Prothrombin %	%	41-69
	Prothrombin INR	INR	1,14-1,71
PT Liq (*)	Prothrombin Time	sec	14,8-22,1 (*)
	Prothrombin %	%	30-50 (*)
	Prothrombin INR	INR	1,34-2,01 (*)
PT R	Prothrombin Time	sec	11,1-16,7
	Prothrombin %	%	42-70
	Prothrombin INR	INR	1,06-1,59
APTT	Partial Thromboplastin Time	sec	51,7-70,0
APTT Liq	Partial Thromboplastin Time	sec	49,6-67,1
FIB Cl	Fibrinogen	g/L	0,8-1,9
TT	Thrombin Time	sec	NA
AT	Antithrombin	%	37-62

Important notice

The Mean Normal Prothrombin Time (MNPT) depends on the population, race, gender, sampling tube, etc. Our value, that is identical with the 100% point of the calibration curve is for information only. According to the CLSI every laboratory should determine its own MNPT.

The % and INR dimension are linked to this MNPT. Then, the ranges of the controls need to be adjusted.

(*) Warning! This range has to be used only with Yumizen G PT Liq 4 lots : 920315 and higher except for the Yumizen G PT Liq 4 lots 920611 & 920705. For the other lots, including 920611 & 920705, please refer to the previous revision of values sheet.