

# Yumizen G DDi 2 for YUMIZEN G1500/G1550



<b>Reference No.:</b>		<b>1300036391</b>	
<b>Lot No.:</b>		<b>920709 (rev. 1)</b>	
<b>Expiry date:</b>		<b>2023-07-31</b>	
<b>Coagulometers</b>		<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G1500/G1550</b>	<b>µg(FEU)/mL</b>	<b>ΔOD</b>
		4,00	0,2475
		3,00	0,1855
		2,00	0,1190
		1,00	0,0555
		0,50	0,0265
		0,25	0,0130
<b>Slope value</b>		<b>0,07</b>	

# Yumizen G DDi 2 for YUMIZEN G800



<b>Reference No.:</b>		<b>1300036391</b>	
<b>Lot No.:</b>		<b>920709 (rev. 1)</b>	
<b>Expiry date:</b>		<b>2023-07-31</b>	
<b>Coagulometers</b>		<b>MASTER CURVE</b>	
<b>Optical</b>	<b>Yumizen G800</b>	<b>µg(FEU)/mL</b>	<b>ΔOD</b>
		4,00	0,2570
		3,00	0,1965
		2,00	0,1190
		1,00	0,0505
		0,50	0,0235
		0,25	0,0115
<b>Slope value</b>		<b>0,06</b>	



# Yumizen G DDi 2 for YUMIZEN G400 DDi/G200

<b>Reference No.:</b>		<b>1300036391</b>	
<b>Lot No.:</b>		<b>920709 (rev. 1)</b>	
<b>Expiry date:</b>		<b>2023-07-31</b>	
<b>Coagulometers</b>		<b>MASTER CURVE</b>	
		<b>µg(FEU)/mL</b>	<b>ΔOD</b>
<b>Optical</b>	<b>Yumizen G400 DDi/G200</b>	<b>4,00</b>	<b>0,1098</b>
		<b>3,00</b>	<b>0,0810</b>
		<b>2,00</b>	<b>0,0525</b>
		<b>1,00</b>	<b>0,0254</b>
		<b>0,50</b>	<b>0,0125</b>
		<b>0,25</b>	<b>0,0065</b>
		<b>Slope value</b>	<b>0,02</b>
<b>Important notice</b>			
<p>For Yumizen G400 DDi and Yumizen G200, make sure that the end time is set to 180 seconds in the D-Dimer setup and the curve type <b>Lin-Lin p-p</b> is selected in <b>Calibration &gt; D-Dimer &gt; ug FEU/ml</b>.          For more information, please contact your local HORIBA Medical representative.</p>			