

ABX Minocal



LOT CX 472
Rev 1

CAL

(Exp.) 2022-08-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	PENTRA	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P 60 18P					
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.0	8.8			9.2					± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.48	4.50			4.49					± 0.06
HB HGB	g/dl	13.6	13.4			13.6					± 0.2
	g/l	136	134			136					± 2
	mmol/l	8.45	8.34			8.45					± 0.12
HT HCT	%	38.5	38.9			39.7					± 1.0
	l/l	0.385	0.389			0.397					± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	236	248			250					± 10
VMP MPV	µm ³ ; fl	8.9	8.5			9.2					± 0.5

ABX Minocal



LOT CX 472

CAL

(Exp.) 2022-08-05
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	LyseVet								TOLERANCES TOLERANCE
		MICROS		MICROS						
		Scil Vet ABC	Scil Vet ABC Plus ESV 60							
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	8.7	8.8							± 0.2
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.57	4.50							± 0.06
HB HGB	g/dl	13.4	13.3							± 0.2
	g/l	134	133							± 2
	mmol/l	8.32	8.26							± 0.12
HT HCT	%	38.4	38.5							± 1.0
	l/l	0.384	0.385							± 0.010
PLA PLT	10 ³ /mm ³ ; 10 ⁹ /l	242	246							± 10
VMP MPV	μm ³ ; fl	N/A	9.3							± 0.5

Ref: TEMP-0387 Rev.48 FRONT / RECTO 1300105208