

ABX Minocal



LOT

CX 477
Rev 1

CAL

(Exp.) **2023-01-05**
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|---|--------------|--------------|--------------|--------------|----------------------|--------------|---------------|---------------------|--------------|-------------------------|
| | | MICROS | MICROS | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | PENTRA | |
| | | EMI CRP | LC-767G | MS CRP | XLR | DX NEXUS DF NEXUS | MS60 | DX120 - DF120 | 60 60 C+ ES60 | XL80 | |
| GB WBC | 10 ³ /mm ³ ;10 ⁹ /l | 9.2 | 9.1 | 9.2 | 9.0 | 9.2 | 8.9 | 9.2 | 9.1 | 9.0 | ± 0.2 |
| GR RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 4.58 | 4.54 | 4.54 | 4.56 | 4.58 | 4.51 | 4.58 | 4.58 | 4.56 | ± 0.06 |
| HB HGB | g/dl | 13.9 | 13.8 | 13.8 | 14.0 | 14.0 | 13.9 | 14.0 | 14.0 | 14.0 | ± 0.2 |
| | g/l | 139 | 138 | 138 | 140 | 140 | 139 | 140 | 140 | 140 | ± 2 |
| | mmol/l | 8.63 | 8.57 | 8.57 | 8.69 | 8.69 | 8.63 | 8.69 | 8.69 | 8.69 | ± 0.12 |
| HT HCT | % | 39.4 | 39.0 | 38.4 | 38.3 | 38.2 | 37.9 | 38.2 | 38.7 | 38.3 | ± 1.0 |
| | l/l | 0.394 | 0.390 | 0.384 | 0.383 | 0.382 | 0.379 | 0.382 | 0.387 | 0.383 | ± 0.010 |
| PLA PLT | 10 ³ /mm ³ ;10 ⁹ /l | 260 | 254 | 246 | 245 | 257 | 250 | 257 | 249 | 245 | ± 10 |
| VMP MPV | µm ³ ; fl | 8.9 | 8.9 | N/A | N/A | N/A | N/A | N/A | N/A | N/A | ± 0.5 |

Ref: TEMP-0387 Rev.48 BACK / VERSO 1300090308