

ABX Minocal


 (Exp.) 2022-10-05
 (YYYY - MM - DD)

LOT		CX 474				CAL		(Exp.) 2022-10-05 (YYYY - MM - DD)			
		Rev 1									
PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60	DX120 - DF120	60 60 C+ ES60	XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.1	9.2	9.2	9.0	9.3	9.1	9.3	9.0	9.0	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.44	4.46	4.40	4.48	4.51	4.47	4.51	4.48	4.48	± 0.06
HB HGB	g/dl	13.2	13.2	13.2	13.3	13.3	13.3	13.3	13.3	13.3	± 0.2
	g/l	132	132	132	133	133	133	133	133	133	± 2
	mmol/l	8.20	8.20	8.20	8.26	8.26	8.26	8.26	8.26	8.26	± 0.12
HT HCT	%	37.3	37.5	35.6	36.5	36.3	36.2	36.3	36.5	36.5	± 1.0
	l/l	0.373	0.375	0.356	0.365	0.363	0.362	0.363	0.365	0.365	± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	248	244	243	243	249	245	249	242	243	± 10
VMP MPV	µm ³ ; fl	8.7	8.3	N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.48 BACK / VERSO 1300090308