

ABX Minocal



LOT		CX 473		CAL		(Exp.) 2022-09-05 (YYYY - MM - DD)					
Rev 1		ABX Lysebio									TOLERANCES TOLERANCE
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	PENTRA	
		EMI CRP	LC-767G	MS CRP	XLR	DX NEXUS DF NEXUS	MS60	DX120 - DF120	60 60 C+ ES60	XL80	
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	9.2	9.4	9.3	9.1	9.4	9.2	9.4	9.1	9.1	± 0.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	4.40	4.43	4.37	4.44	4.49	4.45	4.49	4.44	4.44	± 0.06
HB HGB	g/dl	13.1	13.1	13.0	13.3	13.2	13.1	13.2	13.3	13.3	± 0.2
	g/l	131	131	130	133	132	131	132	133	133	± 2
	mmol/l	8.14	8.14	8.07	8.26	8.20	8.14	8.20	8.26	8.26	± 0.12
HT HCT	%	37.2	37.2	36.1	36.4	36.4	36.0	36.4	36.4	36.4	± 1.0
	l/l	0.372	0.372	0.361	0.364	0.364	0.360	0.364	0.364	0.364	± 0.010
PLA PLT	10 ³ /mm ³ ;10 ⁹ /l	250	248	240	245	254	248	254	248	245	± 10
VMP MPV	µm ³ ; fl	8.3	8.2	N/A	N/A	N/A	N/A	N/A	N/A	N/A	± 0.5

Ref: TEMP-0387 Rev.48 BACK / VERSO 1300090308