



LOT MX 437 L
 Rev 1

CONTROL L



2022-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.26	2.25			2.25					± 0.17
HB HGB	g/dl	5.9	5.9			5.9					± 0.5
	g/l	59	59			59					± 5
	mmol/l	3.66	3.66			3.66					± 0.31
HT HCT	%	15.8	15.8			15.8					± 2.0
	l/l	0.158	0.158			0.158					± 0.020
VGM MCV	µm ³ ; fl	70	70			70					± 4
TGMH MCH	pg	26.1	26.2			26.2					± 2.0
	fmol	1.62	1.63			1.63					± 0.12
CCMH MCHC	g/dl	37.3	37.5			37.5					± 3.0
	g/l	373	375			375					± 30
	mmol/l	23.16	23.26			23.26					± 1.86
IDR RDW	%	15.5	14.5			14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	72	73			73					± 22
VPM MPV	µm ³ ; fl	8.7	8.7			8.0					± 2.0
LYMPHO	#	1.32	1.32			1.32					± 0.40
	%	63.0	63.0			63.0					± 8.0
MONO	#	0.27	0.27			0.27					± 0.20
	%	13.0	13.0			13.0					± 6.0
GRANULO	#	0.50	0.50			0.50					± 0.40
	%	24.0	24.0			24.0					± 7.0

LOT MX 437 N
Rev 1

CONTROL N

 (Exp.)

2022-12-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.6			7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.60			4.59					± 0.26
HB HGB	g/dl	13.6	13.6			13.6					± 0.7
	g/l	136	136			136					± 7
	mmol/l	8.45	8.45			8.45					± 0.43
HT HCT	%	37.3	37.3			37.2					± 2.7
	l/l	0.373	0.373			0.372					± 0.027
VGM MCV	µm ³ ; fl	81	81			81					± 4
TGMH MCH	pg	29.6	29.6			29.6					± 2.0
	fmol	1.84	1.84			1.84					± 0.12
CCMH MCHC	g/dl	36.5	36.5			36.6					± 3.0
	g/l	365	365			366					± 30
	mmol/l	22.67	22.67			22.72					± 1.86
IDR RDW	%	15.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	269	272			261					± 40
VPM MPV	µm ³ ; fl	8.0	7.9			7.2					± 2.0
LYMPHO	#	2.58	2.70			2.62					± 0.40
	%	33.5	35.5			34.0					± 5.0
MONO	#	0.62	0.61			0.58					± 0.30
	%	8.0	8.0			7.5					± 4.0
GRANULO	#	4.50	4.29			4.50					± 0.60
	%	58.5	56.5			58.5					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104144



LOT

MX 437 H
 Rev 1

CONTROL H

(Exp.)

2022-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.7	20.2			20.7					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.74	5.70			5.79					± 0.32
HB HGB	g/dl	18.3	18.2			18.3					± 0.9
	g/l	183	182			183					± 9
	mmol/l	11.36	11.30			11.36					± 0.56
HT HCT	%	51.1	51.3			52.1					± 3.4
	l/l	0.511	0.513			0.521					± 0.034
VGM MCV	µm ³ ; fl	89	90			90					± 4
TGMH MCH	pg	31.9	31.9			31.6					± 2.0
	fmol	1.98	1.98			1.96					± 0.12
CCMH MCHC	g/dl	35.8	35.5			35.1					± 3.0
	g/l	358	355			351					± 30
	mmol/l	22.25	22.03			21.81					± 1.86
IDR RDW	%	14.0	14.0			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	560	546			550					± 65
VPM MPV	µm ³ ; fl	7.9	7.7			7.3					± 2.0
LYMPHO	#	3.31	3.23			3.31					± 1.00
	%	16.0	16.0			16.0					± 5.0
MONO	#	1.35	1.41			1.14					± 0.80
	%	6.5	7.0			5.5					± 4.0
GRANULO	#	16.04	15.55			16.25					± 1.40
	%	77.5	77.0			78.5					± 7.0