

LOT MX 439 L
 Rev 1

CONTROL L

(Exp.)

2023-04-05
 (YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-----------------|---------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP 200 | | | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.1 | 1.9 | | | 2.0 | | | | | ± 0.4 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.33 | 2.35 | | | 2.31 | | | | | ± 0.17 |
| HB HGB | g/dl | 6.2 | 6.1 | | | 6.2 | | | | | ± 0.5 |
| | g/l | 62 | 61 | | | 62 | | | | | ± 5 |
| | mmol/l | 3.85 | 3.79 | | | 3.85 | | | | | ± 0.31 |
| HT HCT | % | 16.8 | 16.9 | | | 16.6 | | | | | ± 2.0 |
| | l/l | 0.168 | 0.169 | | | 0.166 | | | | | ± 0.020 |
| VGM MCV | µm ³ ; fl | 72 | 72 | | | 72 | | | | | ± 4 |
| TGMH MCH | pg | 26.6 | 26.0 | | | 26.8 | | | | | ± 2.0 |
| | fmol | 1.65 | 1.61 | | | 1.67 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 37.0 | 36.1 | | | 37.3 | | | | | ± 3.0 |
| | g/l | 370 | 361 | | | 373 | | | | | ± 30 |
| | mmol/l | 22.95 | 22.39 | | | 23.15 | | | | | ± 1.86 |
| IDR RDW | % | 15.5 | 14.0 | | | 14.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 78 | 79 | | | 76 | | | | | ± 22 |
| VPM MPV | µm ³ ; fl | 9.2 | 8.5 | | | 8.4 | | | | | ± 2.0 |
| LYMPHO | # | 1.34 | 1.24 | | | 1.28 | | | | | ± 0.40 |
| | % | 64.0 | 65.0 | | | 64.0 | | | | | ± 8.0 |
| MONO | # | 0.19 | 0.17 | | | 0.18 | | | | | ± 0.20 |
| | % | 9.0 | 9.0 | | | 9.0 | | | | | ± 6.0 |
| GRANULO | # | 0.57 | 0.49 | | | 0.54 | | | | | ± 0.40 |
| | % | 27.0 | 26.0 | | | 27.0 | | | | | ± 7.0 |

LOT

MX 439 N
Rev 1

CONTROL N

(Exp.)

2023-04-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-----------------|--------------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP 200 | | | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 7.6 | 7.3 | | | 7.6 | | | | | ± 0.8 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 4.53 | 4.52 | | | 4.49 | | | | | ± 0.26 |
| HB HGB | g/dl | 13.4 | 13.5 | | | 13.4 | | | | | ± 0.7 |
| | g/l | 134 | 135 | | | 134 | | | | | ± 7 |
| | mmol/l | 8.32 | 8.38 | | | 8.32 | | | | | ± 0.43 |
| HT HCT | % | 36.7 | 37.1 | | | 36.8 | | | | | ± 2.7 |
| | l/l | 0.367 | 0.371 | | | 0.368 | | | | | ± 0.027 |
| VGM MCV | µm ³ ; fl | 81 | 82 | | | 82 | | | | | ± 4 |
| TGMH MCH | pg | 29.6 | 29.9 | | | 29.8 | | | | | ± 2.0 |
| | fmol | 1.84 | 1.85 | | | 1.85 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.5 | 36.4 | | | 36.4 | | | | | ± 3.0 |
| | g/l | 365 | 364 | | | 364 | | | | | ± 30 |
| | mmol/l | 22.68 | 22.62 | | | 22.60 | | | | | ± 1.86 |
| IDR RDW | % | 15.0 | 13.5 | | | 13.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 278 | 272 | | | 268 | | | | | ± 40 |
| VPM MPV | µm ³ ; fl | 8.6 | 8.2 | | | 7.9 | | | | | ± 2.0 |
| LYMPHO | # | 2.51 | 2.45 | | | 2.51 | | | | | ± 0.40 |
| | % | 33.0 | 33.5 | | | 33.0 | | | | | ± 5.0 |
| MONO | # | 0.57 | 0.62 | | | 0.53 | | | | | ± 0.30 |
| | % | 7.5 | 8.5 | | | 7.0 | | | | | ± 4.0 |
| GRANULO | # | 4.52 | 4.23 | | | 4.56 | | | | | ± 0.60 |
| | % | 59.5 | 58.0 | | | 60.0 | | | | | ± 6.0 |

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104144



LOT

MX 439 H
 Rev 1

CONTROL H

(Exp.)

2023-04-05
 (YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | ABX Lyse | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-----------------|--------------|--------|--------|----------------------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 Care ST | CRP 200 | | | 60 8P - 60 18P ADVIA 60 | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 20.5 | 19.6 | | | 20.3 | | | | | ± 1.7 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 5.54 | 5.50 | | | 5.50 | | | | | ± 0.32 |
| HB HGB | g/dl | 17.7 | 17.6 | | | 17.7 | | | | | ± 0.9 |
| | g/l | 177 | 176 | | | 177 | | | | | ± 9 |
| | mmol/l | 10.99 | 10.93 | | | 10.99 | | | | | ± 0.56 |
| HT HCT | % | 49.3 | 49.0 | | | 49.5 | | | | | ± 3.4 |
| | l/l | 0.493 | 0.490 | | | 0.495 | | | | | ± 0.034 |
| VGM MCV | µm ³ ; fl | 89 | 89 | | | 90 | | | | | ± 4 |
| TGMH MCH | pg | 31.9 | 32.0 | | | 32.2 | | | | | ± 2.0 |
| | fmol | 1.98 | 1.99 | | | 2.00 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 35.9 | 36.0 | | | 35.8 | | | | | ± 3.0 |
| | g/l | 359 | 360 | | | 358 | | | | | ± 30 |
| | mmol/l | 22.29 | 22.33 | | | 22.21 | | | | | ± 1.86 |
| IDR RDW | % | 14.0 | 13.0 | | | 12.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 530 | 525 | | | 516 | | | | | ± 65 |
| VPM MPV | µm ³ ; fl | 8.5 | 8.2 | | | 7.8 | | | | | ± 2.0 |
| LYMPHO | # | 3.49 | 3.23 | | | 3.25 | | | | | ± 1.00 |
| | % | 17.0 | 16.5 | | | 16.0 | | | | | ± 5.0 |
| MONO | # | 1.13 | 1.18 | | | 1.12 | | | | | ± 0.80 |
| | % | 5.5 | 6.0 | | | 5.5 | | | | | ± 4.0 |
| GRANULO | # | 15.89 | 15.19 | | | 15.94 | | | | | ± 1.40 |
| | % | 77.5 | 77.5 | | | 78.5 | | | | | ± 7.0 |