

LOT MX 437 L
Rev 1

CONTROL L

 (Exp.)

2022-12-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1	2.1			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.27	2.25			2.23					± 0.17
HB HGB	g/dl	5.7	5.8			5.5					± 0.5
	g/l	57	58			55					± 5
	mmol/l	3.54	3.59			3.41					± 0.31
HT HCT	%	15.9	15.8			15.8					± 2.0
	l/l	0.159	0.158			0.158					± 0.020
VGM MCV	µm ³ ; fl	70	70			71					± 4
TGMH MCH	pg	25.1	25.7			24.6					± 2.0
	fmol	1.56	1.60			1.53					± 0.12
CCMH MCHC	g/dl	35.9	36.7			34.7					± 3.0
	g/l	359	367			347					± 30
	mmol/l	22.28	22.80			21.52					± 1.86
IDR RDW	%	15.5	14.5			16.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	71	73			73					± 22
VPM MPV	µm ³ ; fl	8.5	8.7			8.0					± 2.0
LYMPHO	#	1.37	1.34			1.34					± 0.40
	%	65.0	63.6			64.0					± 8.0
MONO	#	0.23	0.27			0.25					± 0.20
	%	11.0	13.0			12.0					± 6.0
GRANULO	#	0.50	0.49			0.50					± 0.40
	%	24.0	23.5			24.0					± 7.0

LOT MX 437 N
 Rev 1

CONTROL N

(Exp.)

2022-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.7	7.6			7.5					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.60			4.55					± 0.26
HB HGB	g/dl	13.5	13.6			13.3					± 0.7
	g/l	135	136			133					± 7
	mmol/l	8.38	8.45			8.28					± 0.43
HT HCT	%	37.3	37.3			37.8					± 2.7
	l/l	0.373	0.373			0.378					± 0.027
VGM MCV	µm ³ ; fl	81	81			83					± 4
TGMH MCH	pg	29.3	29.6			29.3					± 2.0
	fmol	1.82	1.84			1.82					± 0.12
CCMH MCHC	g/dl	36.2	36.5			35.3					± 3.0
	g/l	362	365			353					± 30
	mmol/l	22.50	22.67			21.92					± 1.86
IDR RDW	%	15.0	14.0			15.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	255	272			261					± 40
VPM MPV	µm ³ ; fl	8.1	7.9			7.2					± 2.0
LYMPHO	#	2.70	2.86			2.59					± 0.40
	%	35.0	37.6			34.5					± 5.0
MONO	#	0.54	0.54			0.56					± 0.30
	%	7.0	7.1			7.5					± 4.0
GRANULO	#	4.47	4.21			4.35					± 0.60
	%	58.0	55.4			58.0					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143



LOT MX 437 H
 Rev 1

CONTROL H

(Exp.)

2022-12-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	Minilysebio								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60	CRP 200			60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.4	20.2			20.1					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.78	5.70			5.74					± 0.32
HB HGB	g/dl	18.6	18.2			18.1					± 0.9
	g/l	186	182			181					± 9
	mmol/l	11.55	11.30			11.25					± 0.56
HT HCT	%	51.4	51.3			52.2					± 3.4
	l/l	0.514	0.513			0.522					± 0.034
VGM MCV	µm ³ ; fl	89	90			91					± 4
TGMH MCH	pg	32.2	31.9			31.6					± 2.0
	fmol	2.00	1.98			1.96					± 0.12
CCMH MCHC	g/dl	36.2	35.5			34.7					± 3.0
	g/l	362	355			347					± 30
	mmol/l	22.45	22.03			21.54					± 1.86
IDR RDW	%	14.0	14.0			13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	530	546			550					± 65
VPM MPV	µm ³ ; fl	8.0	7.7			7.3					± 2.0
LYMPHO	#	3.37	3.43			3.32					± 1.00
	%	16.5	17.0			16.5					± 5.0
MONO	#	1.12	1.53			1.11					± 0.80
	%	5.5	7.6			5.5					± 4.0
GRANULO	#	15.91	15.24			15.68					± 1.40
	%	78.0	75.5			78.0					± 7.0