

LOT MX 438 L
 Rev 1

CONTROL L

(Exp.)

2023-02-05
 (YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-------------|---------|--------|--------|----------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 | CRP 200 | | | 60 8P - 60 18P | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.1 | 2.2 | | | 2.1 | | | | | ± 0.4 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.27 | 2.27 | | | 2.25 | | | | | ± 0.17 |
| HB HGB | g/dl | 5.8 | 5.9 | | | 5.6 | | | | | ± 0.5 |
| | g/l | 58 | 59 | | | 56 | | | | | ± 5 |
| | mmol/l | 3.60 | 3.65 | | | 3.47 | | | | | ± 0.31 |
| HT HCT | % | 16.1 | 16.3 | | | 16.4 | | | | | ± 2.0 |
| | l/l | 0.161 | 0.163 | | | 0.164 | | | | | ± 0.020 |
| VGM MCV | µm ³ ; fl | 71 | 72 | | | 73 | | | | | ± 4 |
| TGMH MCH | pg | 25.6 | 25.9 | | | 24.8 | | | | | ± 2.0 |
| | fmol | 1.59 | 1.61 | | | 1.54 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.0 | 36.0 | | | 34.0 | | | | | ± 3.0 |
| | g/l | 360 | 360 | | | 340 | | | | | ± 30 |
| | mmol/l | 22.35 | 22.34 | | | 21.10 | | | | | ± 1.86 |
| IDR RDW | % | 15.5 | 14.0 | | | 15.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 81 | 85 | | | 81 | | | | | ± 22 |
| VPM MPV | µm ³ ; fl | 7.9 | 8.0 | | | 7.5 | | | | | ± 2.0 |
| LYMPHO | # | 1.43 | 1.47 | | | 1.42 | | | | | ± 0.40 |
| | % | 68.0 | 66.7 | | | 67.5 | | | | | ± 8.0 |
| MONO | # | 0.17 | 0.22 | | | 0.17 | | | | | ± 0.20 |
| | % | 8.0 | 10.0 | | | 8.0 | | | | | ± 6.0 |
| GRANULO | # | 0.50 | 0.52 | | | 0.51 | | | | | ± 0.40 |
| | % | 24.0 | 23.5 | | | 24.5 | | | | | ± 7.0 |

LOT MX 438 N
 Rev 1

CONTROL N

(Exp.)

2023-02-05
 (YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|--|-------------|---------|--------|--------|----------------|----------------|--------|--------|--------|-------------------------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | PENTRA | |
| | | ES60 | CRP 200 | | | 60 BP - 60 18P | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 7.6 | 7.6 | | | 7.5 | | | | | ± 0.8 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 4.61 | 4.59 | | | 4.57 | | | | | ± 0.26 |
| HB HGB | g/dl | 13.4 | 13.4 | | | 13.1 | | | | | ± 0.7 |
| | g/l | 134 | 134 | | | 131 | | | | | ± 7 |
| | mmol/l | 8.32 | 8.32 | | | 8.15 | | | | | ± 0.43 |
| HT HCT | % | 37.3 | 37.2 | | | 38.4 | | | | | ± 2.7 |
| | l/l | 0.373 | 0.372 | | | 0.384 | | | | | ± 0.027 |
| VGM MCV | µm ³ ; fl | 81 | 81 | | | 84 | | | | | ± 4 |
| TGMH MCH | pg | 29.1 | 29.2 | | | 28.7 | | | | | ± 2.0 |
| | fmol | 1.81 | 1.81 | | | 1.78 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 35.9 | 36.0 | | | 34.2 | | | | | ± 3.0 |
| | g/l | 359 | 360 | | | 342 | | | | | ± 30 |
| | mmol/l | 22.28 | 22.38 | | | 21.24 | | | | | ± 1.86 |
| IDR RDW | % | 13.5 | 13.0 | | | 14.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 244 | 257 | | | 248 | | | | | ± 40 |
| VPM MPV | µm ³ ; fl | 8.2 | 8.3 | | | 7.6 | | | | | ± 2.0 |
| LYMPHO | # | 2.66 | 2.74 | | | 2.55 | | | | | ± 0.40 |
| | % | 35.0 | 36.0 | | | 34.0 | | | | | ± 5.0 |
| MONO | # | 0.46 | 0.51 | | | 0.49 | | | | | ± 0.30 |
| | % | 6.0 | 6.7 | | | 6.5 | | | | | ± 4.0 |
| GRANULO | # | 4.48 | 4.36 | | | 4.46 | | | | | ± 0.60 |
| | % | 59.0 | 57.3 | | | 59.5 | | | | | ± 6.0 |

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143



LOT

MX 438 H

CONTROL

H

(Exp.)

2023-02-05
(YYYY-MM-DD)

Rev 1

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|--------------|--------------|--------|--------|----------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 | CRP 200 | | | 60 8P - 60 18P | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 20.3 | 20.2 | | | 20.2 | | | | | ± 1.7 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 5.70 | 5.67 | | | 5.68 | | | | | ± 0.32 |
| HB HGB | g/dl | 18.6 | 18.2 | | | 18.0 | | | | | ± 0.9 |
| | g/l | 186 | 182 | | | 180 | | | | | ± 9 |
| | mmol/l | 11.55 | 11.30 | | | 11.19 | | | | | ± 0.56 |
| HT HCT | % | 50.7 | 50.5 | | | 51.7 | | | | | ± 3.4 |
| | l/l | 0.507 | 0.505 | | | 0.517 | | | | | ± 0.034 |
| VGM MCV | µm ³ ; fl | 89 | 89 | | | 91 | | | | | ± 4 |
| TGMH MCH | pg | 32.6 | 32.1 | | | 31.7 | | | | | ± 2.0 |
| | fmol | 2.03 | 1.99 | | | 1.97 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.7 | 36.1 | | | 34.9 | | | | | ± 3.0 |
| | g/l | 367 | 361 | | | 349 | | | | | ± 30 |
| | mmol/l | 22.77 | 22.40 | | | 21.65 | | | | | ± 1.86 |
| IDR RDW | % | 13.5 | 13.0 | | | 14.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 524 | 536 | | | 528 | | | | | ± 65 |
| VPM MPV | µm ³ ; fl | 7.8 | 7.8 | | | 7.2 | | | | | ± 2.0 |
| LYMPHO | # | 3.45 | 3.43 | | | 3.43 | | | | | ± 1.00 |
| | % | 17.0 | 17.0 | | | 17.0 | | | | | ± 5.0 |
| MONO | # | 1.02 | 1.31 | | | 1.01 | | | | | ± 0.80 |
| | % | 5.0 | 6.5 | | | 5.0 | | | | | ± 4.0 |
| GRANULO | # | 15.83 | 15.44 | | | 15.76 | | | | | ± 1.40 |
| | % | 78.0 | 76.4 | | | 78.0 | | | | | ± 7.0 |