

# ABX Minotrol CRP



**LOT**

**MC 424**

**CONTROL**



(Exp.) **2022-10-05**  
(YYYY - MM - DD)

Rev 1

		ABX Lyse														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	2.1	2.1			± 0.4	7.3	7.5			± 1.0	19.0	19.9			± 2.2
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	2.26	2.23			± 0.15	4.47	4.50			± 0.18	5.59	5.65			± 0.25
HB HGB	g/dl	6.0	5.9			± 0.4	13.6	13.6			± 0.5	17.9	17.9			± 0.7
	g/l	60	59			± 4	136	136			± 5	179	179			± 7
	mmol/l	3.73	3.66			± 0.25	8.45	8.45			± 0.31	11.12	11.12			± 0.44
HT HCT	%	16.3	16.1			± 2.0	37.5	37.8			± 2.5	50.9	50.9			± 3.0
	l/l	0.163	0.161			± 0.020	0.375	0.378			± 0.025	0.509	0.509			± 0.030
VGM MCV	µm <sup>3</sup> ; fl	72	72			± 4	84	84			± 4	91	90			± 4
TGMH MCH	pg	26.5	26.5			± 2.4	30.4	30.2			± 2.4	32.0	31.7			± 2.8
	fmol	1.65	1.64			± 0.15	1.89	1.88			± 0.15	1.99	1.97			± 0.18
CCMH MCHC	g/dl	36.9	36.7			± 3.0	36.2	36.0			± 3.0	35.2	35.2			± 3.0
	g/l	369	367			± 30	362	360			± 30	352	352			± 30
	mmol/l	22.90	22.82			± 1.86	22.49	22.34			± 1.86	21.85	21.86			± 1.86
IDR RDW	%	13.5	13.5			± 3.0	13.0	12.5			± 3.0	12.5	12.0			± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	75	80			± 20	262	269			± 40	489	508			± 60
VPM MPV	µm <sup>3</sup> ; fl	8.5	7.5			± 2.0	8.0	7.6			± 2.0	7.7	7.2			± 2.0
LYMPHO	#	1.3	1.3			± 0.2	2.6	2.6			± 0.5	3.2	3.2			± 1.1
	%	64.0	61.0			± 8.0	36.0	35.0			± 6.0	17.0	16.0			± 5.0
MONO	#	0.2	0.2			± 0.2	0.6	0.6			± 0.4	1.1	1.0			± 1.0
	%	10.5	11.0			± 7.0	8.5	8.0			± 5.0	6.0	5.0			± 5.0
GRANULO	#	0.5	0.6			± 0.2	4.1	4.3			± 0.5	14.6	15.7			± 1.5
	%	25.5	28.0			± 7.0	55.5	57.0			± 6.0	77.0	79.0			± 7.0
CRP	mg/dl	0.45	N/A			± 0.30	2.03	N/A			± 0.70	3.50	N/A			± 1.00
	mg/l	4.5	N/A			± 3.0	20.3	N/A			± 7.0	35.0	N/A			± 10.0

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170