

ABX Minotrol CRP



LOT

MC 424

CONTROL



(Exp.) **2022-10-05**
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1		± 0.4	7.3	7.5		± 1.0	18.6	19.9		± 2.2				
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.26	2.21		± 0.15	4.47	4.50		± 0.18	5.59	5.61		± 0.25				
HB HGB	g/dl	5.7	5.7		± 0.4	13.3	13.6		± 0.5	17.9	18.2		± 0.7				
	g/l	57	57		± 4	133	136		± 5	179	182		± 7				
HT HCT	mmol/l	3.54	3.54		± 0.25	8.28	8.45		± 0.31	11.12	11.30		± 0.44				
	%	16.3	16.4		± 2.0	37.5	38.3		± 2.5	50.9	51.1		± 3.0				
VGM MCV	l/l	0.163	0.164		± 0.020	0.375	0.383		± 0.025	0.509	0.511		± 0.030				
	µm ³ ; fl	72	74		± 4	84	85		± 4	91	91		± 4				
TGMH MCH	pg	25.2	25.8		± 2.4	29.8	30.2		± 2.4	32.0	32.4		± 2.8				
	fmol	1.57	1.60		± 0.15	1.85	1.88		± 0.15	1.99	2.01		± 0.18				
CCMH MCHC	g/dl	35.0	34.9		± 3.0	35.5	35.6		± 3.0	35.2	35.7		± 3.0				
	g/l	350	349		± 30	355	356		± 30	352	357		± 30				
	mmol/l	21.75	21.64		± 1.86	22.04	22.08		± 1.86	21.85	22.14		± 1.86				
IDR RDW	%	13.5	14.0		± 3.0	13.0	13.5		± 3.0	12.5	13.0		± 3.0				
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	75	80		± 20	262	269		± 40	489	508		± 60				
VPM MPV	µm ³ ; fl	8.5	9.0		± 2.0	8.0	8.7		± 2.0	7.7	8.3		± 2.0				
LYMPHO	#	1.4	1.3		± 0.2	2.8	2.7		± 0.5	3.4	3.2		± 1.1				
	%	66.6	63.0		± 8.0	37.8	36.0		± 6.0	18.2	16.0		± 5.0				
MONO	#	0.2	0.2		± 0.2	0.6	0.5		± 0.4	1.1	1.1		± 1.0				
	%	8.4	9.0		± 7.0	8.1	7.0		± 5.0	6.0	5.5		± 5.0				
GRANULO	#	0.5	0.6		± 0.2	4.0	4.3		± 0.5	14.2	15.6		± 1.5				
	%	25.0	28.0		± 7.0	54.4	57.0		± 6.0	76.2	78.5		± 7.0				
CRP	mg/dl	0.45	N/A		± 0.30	2.03	N/A		± 0.70	3.50	N/A		± 1.00				
	mg/l	4.5	N/A		± 3.0	20.3	N/A		± 7.0	35.0	N/A		± 10.0				

Ref: TEMP-1070 Rev.42 1300027267-A