

# ABX Minotrol CRP



**LOT**

**MC 425**

**CONTROL**



**(Exp.) 2023-01-05**  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE													
				CONTROL				1				CONTROL						2				CONTROL				3			
				MICROS								MICROS								MICROS									
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P										
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l		<b>2.1</b>	<b>2.1</b>			± 0.4	<b>7.6</b>	<b>7.7</b>			± 1.0	<b>20.0</b>	<b>20.8</b>			± 2.2											
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l		<b>2.24</b>	<b>2.23</b>			± 0.15	<b>4.57</b>	<b>4.64</b>			± 0.18	<b>5.52</b>	<b>5.52</b>			± 0.25											
HB	HGB	g/dl		<b>5.8</b>	<b>5.9</b>			± 0.4	<b>13.2</b>	<b>13.7</b>			± 0.5	<b>18.0</b>	<b>18.7</b>			± 0.7											
		g/l		<b>58</b>	<b>59</b>			± 4	<b>132</b>	<b>137</b>			± 5	<b>180</b>	<b>187</b>			± 7											
		mmol/l		<b>3.60</b>	<b>3.66</b>			± 0.25	<b>8.22</b>	<b>8.51</b>			± 0.31	<b>11.18</b>	<b>11.61</b>			± 0.44											
HT	HCT	%		<b>16.4</b>	<b>16.7</b>			± 2.0	<b>37.0</b>	<b>39.0</b>			± 2.5	<b>50.8</b>	<b>51.9</b>			± 3.0											
		l/l		<b>0.164</b>	<b>0.167</b>			± 0.020	<b>0.370</b>	<b>0.390</b>			± 0.025	<b>0.508</b>	<b>0.519</b>			± 0.030											
VGM	MCV	µm <sup>3</sup> ; fl		<b>73</b>	<b>75</b>			± 4	<b>81</b>	<b>84</b>			± 4	<b>92</b>	<b>94</b>			± 4											
TGMH	MCH	pg		<b>25.9</b>	<b>26.5</b>			± 2.4	<b>28.9</b>	<b>29.5</b>			± 2.4	<b>32.6</b>	<b>33.9</b>			± 2.8											
		fmol		<b>1.61</b>	<b>1.64</b>			± 0.15	<b>1.80</b>	<b>1.83</b>			± 0.15	<b>2.03</b>	<b>2.10</b>			± 0.18											
CCMH	MCHC	g/dl		<b>35.4</b>	<b>35.3</b>			± 3.0	<b>35.7</b>	<b>35.1</b>			± 3.0	<b>35.4</b>	<b>36.0</b>			± 3.0											
		g/l		<b>354</b>	<b>353</b>			± 30	<b>357</b>	<b>351</b>			± 30	<b>354</b>	<b>360</b>			± 30											
		mmol/l		<b>22.01</b>	<b>21.91</b>			± 1.86	<b>22.19</b>	<b>21.83</b>			± 1.86	<b>22.01</b>	<b>22.38</b>			± 1.86											
IDR	RDW	%		<b>13.5</b>	<b>15.0</b>			± 3.0	<b>13.5</b>	<b>14.0</b>			± 3.0	<b>13.0</b>	<b>14.0</b>			± 3.0											
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l		<b>82</b>	<b>80</b>			± 20	<b>271</b>	<b>270</b>			± 40	<b>522</b>	<b>511</b>			± 60											
VPM	MPV	µm <sup>3</sup> ; fl		<b>8.4</b>	<b>8.9</b>			± 2.0	<b>7.8</b>	<b>8.4</b>			± 2.0	<b>8.0</b>	<b>8.6</b>			± 2.0											
LYMPHO		#		<b>1.4</b>	<b>1.4</b>			± 0.2	<b>2.8</b>	<b>2.7</b>			± 0.5	<b>3.8</b>	<b>3.7</b>			± 1.1											
		%		<b>68.1</b>	<b>66.0</b>			± 8.0	<b>37.3</b>	<b>35.5</b>			± 6.0	<b>19.0</b>	<b>18.0</b>			± 5.0											
MONO		#		<b>0.2</b>	<b>0.2</b>			± 0.2	<b>0.5</b>	<b>0.5</b>			± 0.4	<b>1.4</b>	<b>1.2</b>			± 1.0											
		%		<b>7.6</b>	<b>8.5</b>			± 7.0	<b>6.7</b>	<b>6.0</b>			± 5.0	<b>7.0</b>	<b>6.0</b>			± 5.0											
GRANULO		#		<b>0.5</b>	<b>0.5</b>			± 0.2	<b>4.3</b>	<b>4.5</b>			± 0.5	<b>15.0</b>	<b>15.7</b>			± 1.5											
		%		<b>24.5</b>	<b>25.5</b>			± 7.0	<b>56.4</b>	<b>58.5</b>			± 6.0	<b>75.0</b>	<b>76.0</b>			± 7.0											
CRP		mg/dl		<b>0.50</b>	N/A			± 0.30	<b>2.48</b>	N/A			± 0.70	<b>4.01</b>	N/A			± 1.00											
		mg/l		<b>5.0</b>	N/A			± 3.0	<b>24.8</b>	N/A			± 7.0	<b>40.1</b>	N/A			± 10.0											

Ref: TEMP-1070 Rev.42 1300027267-A