

ABX Minotrol CRP



LOT

MC 426

CONTROL



(Exp.) **2023-04-05**
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		Minilysebio												TOLERANCES TOLERANCE							
				CONTROL				1		CONTROL				2				CONTROL				3	
				MICROS				TOLERANCES TOLERANCE		MICROS				TOLERANCES TOLERANCE				MICROS				TOLERANCES TOLERANCE	
				CRP 200	60 8P 60 18P					CRP 200	60 8P 60 18P							CRP 200	60 8P 60 18P				
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l		2.0	2.0			± 0.4		7.4	7.4			± 1.0		19.3	20.3			± 2.2			
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l		2.34	2.32			± 0.15		4.48	4.48			± 0.18		5.48	5.50			± 0.25			
HB	HGB	g/dl		5.9	6.0			± 0.4		12.9	13.4			± 0.5		17.4	18.1			± 0.7			
		g/l		59	60			± 4		129	134			± 5		174	181			± 7			
		mmol/l		3.66	3.73			± 0.25		8.03	8.32			± 0.31		10.81	11.24			± 0.44			
HT	HCT	%		16.8	16.9			± 2.0		36.3	37.6			± 2.5		49.3	50.1			± 3.0			
		l/l		0.168	0.169			± 0.020		0.363	0.376			± 0.025		0.493	0.501			± 0.030			
VGM	MCV	µm ³ ; fl		72	73			± 4		81	84			± 4		90	91			± 4			
TGMH	MCH	pg		25.2	25.9			± 2.4		28.9	29.9			± 2.4		31.8	32.9			± 2.8			
		fmol		1.56	1.61			± 0.15		1.79	1.86			± 0.15		1.97	2.04			± 0.18			
CCMH	MCHC	g/dl		35.0	35.4			± 3.0		35.6	35.6			± 3.0		35.3	36.2			± 3.0			
		g/l		350	354			± 30		356	356			± 30		353	362			± 30			
		mmol/l		21.71	22.00			± 1.86		22.14	22.11			± 1.86		21.91	22.46			± 1.86			
IDR	RDW	%		14.0	15.0			± 3.0		13.5	14.0			± 3.0		13.0	13.5			± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l		77	76			± 20		270	270			± 40		520	512			± 60			
VPM	MPV	µm ³ ; fl		8.8	8.9			± 2.0		8.4	8.8			± 2.0		8.3	9.0			± 2.0			
LYMPHO		#		1.4	1.3			± 0.2		2.6	2.6			± 0.5		3.5	3.5			± 1.1			
		%		68.6	66.0			± 8.0		35.7	34.5			± 6.0		18.2	17.0			± 5.0			
MONO		#		0.1	0.1			± 0.2		0.5	0.5			± 0.4		1.2	1.0			± 1.0			
		%		7.2	6.0			± 7.0		7.1	6.5			± 5.0		6.0	5.0			± 5.0			
GRANULO		#		0.5	0.6			± 0.2		4.2	4.4			± 0.5		14.7	15.8			± 1.5			
		%		24.5	28.0			± 7.0		57.3	59.0			± 6.0		76.2	78.0			± 7.0			
CRP		mg/dl		0.40	N/A			± 0.30		1.93	N/A			± 0.70		4.38	N/A			± 1.00			
		mg/l		4.0	N/A			± 3.0		19.3	N/A			± 7.0		43.8	N/A			± 10.0			

Ref: TEMP-1070 Rev.42 1300027267-A