

ABX Minotrol CRP



LOT

MC 425

CONTROL



(Exp.) **2023-01-05**

(YYYY-MM-DD)

Rev 1

		ABX Lyse														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE
		CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P				CRP 200	60 8P 60 18P			
GB WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1			± 0.4	7.6	7.7			± 1.0	20.4	20.8			± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.24	2.22			± 0.15	4.57	4.61			± 0.18	5.52	5.53			± 0.25
HB HGB	g/dl	6.1	6.1			± 0.4	13.5	13.7			± 0.5	18.0	18.3			± 0.7
	g/l	61	61			± 4	135	137			± 5	180	183			± 7
	mmol/l	3.79	3.79			± 0.25	8.38	8.51			± 0.31	11.18	11.36			± 0.44
HT HCT	%	16.4	16.2			± 2.0	37.0	37.8			± 2.5	50.8	50.9			± 3.0
	l/l	0.164	0.162			± 0.020	0.370	0.378			± 0.025	0.508	0.509			± 0.030
VGM MCV	µm ³ ; fl	73	73			± 4	81	82			± 4	92	92			± 4
TGMH MCH	pg	27.2	27.5			± 2.4	29.5	29.7			± 2.4	32.6	33.1			± 2.8
	fmol	1.69	1.71			± 0.15	1.83	1.85			± 0.15	2.03	2.06			± 0.18
CCMH MCHC	g/dl	37.3	37.6			± 3.0	36.5	36.2			± 3.0	35.4	36.0			± 3.0
	g/l	373	376			± 30	365	362			± 30	354	360			± 30
	mmol/l	23.17	23.37			± 1.86	22.65	22.51			± 1.86	22.01	22.34			± 1.86
IDR RDW	%	13.5	13.5			± 3.0	13.5	13.0			± 3.0	13.0	12.5			± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l	82	78			± 20	271	269			± 40	522	513			± 60
VPM MPV	µm ³ ; fl	8.4	7.9			± 2.0	7.8	7.1			± 2.0	8.0	7.4			± 2.0
LYMPHO	#	1.4	1.4			± 0.2	2.7	2.6			± 0.5	3.6	3.5			± 1.1
	%	65.5	64.5			± 8.0	35.5	34.0			± 6.0	17.5	17.0			± 5.0
MONO	#	0.2	0.2			± 0.2	0.5	0.6			± 0.4	1.4	1.4			± 1.0
	%	9.5	9.5			± 7.0	7.0	7.5			± 5.0	7.0	6.5			± 5.0
GRANULO	#	0.5	0.5			± 0.2	4.4	4.5			± 0.5	15.4	15.9			± 1.5
	%	25.0	26.0			± 7.0	57.5	58.5			± 6.0	75.5	76.5			± 7.0
CRP	mg/dl	0.50	N/A			± 0.30	2.48	N/A			± 0.70	4.01	N/A			± 1.00
	mg/l	5.0	N/A			± 3.0	24.8	N/A			± 7.0	40.1	N/A			± 10.0

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170