

ABX Minotrol CRP



LOT

MC 426

CONTROL



(Exp.) **2023-04-05**
(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lyse												TOLERANCES TOLERANCE	
				CONTROL			1	CONTROL			2	CONTROL			3		
				MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE	MICROS			TOLERANCES TOLERANCE		
				CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P			CRP 200	60 8P 60 18P				
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0		± 0.4	7.4	7.5		± 1.0	19.7	20.4		± 2.2			
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.34	2.33		± 0.15	4.48	4.47		± 0.18	5.48	5.46		± 0.25			
HB	HGB	g/dl	6.2	6.2		± 0.4	13.2	13.4		± 0.5	17.4	17.6		± 0.7			
		g/l	62	62		± 4	132	134		± 5	174	176		± 7			
		mmol/l	3.85	3.85		± 0.25	8.20	8.32		± 0.31	10.81	10.93		± 0.44			
HT	HCT	%	16.8	16.8		± 2.0	36.3	36.7		± 2.5	49.3	49.1		± 3.0			
		l/l	0.168	0.168		± 0.020	0.363	0.367		± 0.025	0.493	0.491		± 0.030			
VGM	MCV	µm ³ ; fl	72	72		± 4	81	82		± 4	90	90		± 4			
TGMH	MCH	pg	26.5	26.6		± 2.4	29.5	30.0		± 2.4	31.8	32.2		± 2.8			
		fmol	1.65	1.65		± 0.15	1.83	1.86		± 0.15	1.97	2.00		± 0.18			
CCMH	MCHC	g/dl	36.8	37.0		± 3.0	36.4	36.6		± 3.0	35.3	35.8		± 3.0			
		g/l	368	370		± 30	364	366		± 30	353	358		± 30			
		mmol/l	22.85	22.95		± 1.86	22.59	22.70		± 1.86	21.91	22.24		± 1.86			
IDR	RDW	%	14.0	14.0		± 3.0	13.5	13.0		± 3.0	13.0	12.5		± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	77	76		± 20	270	269		± 40	520	514		± 60			
VPM	MPV	µm ³ ; fl	8.8	7.8		± 2.0	8.4	7.7		± 2.0	8.3	7.7		± 2.0			
LYMPHO		#	1.3	1.3		± 0.2	2.5	2.5		± 0.5	3.3	3.3		± 1.1			
		%	66.0	64.0		± 8.0	34.0	33.5		± 6.0	17.0	16.0		± 5.0			
MONO		#	0.2	0.2		± 0.2	0.6	0.5		± 0.4	1.2	1.1		± 1.0			
		%	9.0	8.0		± 7.0	7.5	7.0		± 5.0	6.0	5.5		± 5.0			
GRANULO		#	0.5	0.6		± 0.2	4.3	4.5		± 0.5	15.2	16.0		± 1.5			
		%	25.0	28.0		± 7.0	58.5	59.5		± 6.0	77.0	78.5		± 7.0			
CRP		mg/dl	0.40	N/A		± 0.30	1.93	N/A		± 0.70	4.38	N/A		± 1.00			
		mg/l	4.0	N/A		± 3.0	19.3	N/A		± 7.0	43.8	N/A		± 10.0			

Ref: TEMP-1070 Rev.42 FRONT/RECTO 1300092170