



**LOT** MX 437 L  
 Rev 1

**CONTROL** L



**2022-12-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.1	2.1			2.1					± 0.4
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.26	2.25			2.25					± 0.17
HB HGB	g/dl	5.9	5.9			5.9					± 0.5
	g/l	59	59			59					± 5
	mmol/l	3.66	3.66			3.66					± 0.31
HT HCT	%	15.8	15.8			15.8					± 2.0
	l/l	0.158	0.158			0.158					± 0.020
VGM MCV	µm <sup>3</sup> ; fl	70	70			70					± 4
TGMH MCH	pg	26.1	26.2			26.2					± 2.0
	fmol	1.62	1.63			1.63					± 0.12
CCMH MCHC	g/dl	37.3	37.5			37.5					± 3.0
	g/l	373	375			375					± 30
	mmol/l	23.16	23.26			23.26					± 1.86
IDR RDW	%	15.5	14.5			14.5					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	72	73			73					± 22
VPM MPV	µm <sup>3</sup> ; fl	8.7	8.7			8.0					± 2.0
LYMPHO	#	1.32	1.32			1.32					± 0.40
	%	63.0	63.0			63.0					± 8.0
MONO	#	0.27	0.27			0.27					± 0.20
	%	13.0	13.0			13.0					± 6.0
GRANULO	#	0.50	0.50			0.50					± 0.40
	%	24.0	24.0			24.0					± 7.0



**LOT** MX 437 N  
 Rev 1

**CONTROL** N

(Exp.)

**2022-12-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.7	7.6			7.7					± 0.8
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	4.60	4.60			4.59					± 0.26
HB HGB	g/dl	13.6	13.6			13.6					± 0.7
	g/l	136	136			136					± 7
	mmol/l	8.45	8.45			8.45					± 0.43
HT HCT	%	37.3	37.3			37.2					± 2.7
	l/l	0.373	0.373			0.372					± 0.027
VGM MCV	µm <sup>3</sup> ; fl	81	81			81					± 4
TGMH MCH	pg	29.6	29.6			29.6					± 2.0
	fmol	1.84	1.84			1.84					± 0.12
CCMH MCHC	g/dl	36.5	36.5			36.6					± 3.0
	g/l	365	365			366					± 30
	mmol/l	22.67	22.67			22.72					± 1.86
IDR RDW	%	15.0	14.0			14.0					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	269	272			261					± 40
VPM MPV	µm <sup>3</sup> ; fl	8.0	7.9			7.2					± 2.0
LYMPHO	#	2.58	2.70			2.62					± 0.40
	%	33.5	35.5			34.0					± 5.0
MONO	#	0.62	0.61			0.58					± 0.30
	%	8.0	8.0			7.5					± 4.0
GRANULO	#	4.50	4.29			4.50					± 0.60
	%	58.5	56.5			58.5					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104144



**LOT** MX 437 H  
 Rev 1

**CONTROL** H



**2022-12-05**  
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>20.7</b>	<b>20.2</b>			<b>20.7</b>					± 1.7
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>5.74</b>	<b>5.70</b>			<b>5.79</b>					± 0.32
HB HGB	g/dl	<b>18.3</b>	<b>18.2</b>			<b>18.3</b>					± 0.9
	g/l	<b>183</b>	<b>182</b>			<b>183</b>					± 9
	mmol/l	<b>11.36</b>	<b>11.30</b>			<b>11.36</b>					± 0.56
HT HCT	%	<b>51.1</b>	<b>51.3</b>			<b>52.1</b>					± 3.4
	l/l	<b>0.511</b>	<b>0.513</b>			<b>0.521</b>					± 0.034
VGM MCV	µm <sup>3</sup> ; fl	<b>89</b>	<b>90</b>			<b>90</b>					± 4
TGMH MCH	pg	<b>31.9</b>	<b>31.9</b>			<b>31.6</b>					± 2.0
	fmol	<b>1.98</b>	<b>1.98</b>			<b>1.96</b>					± 0.12
CCMH MCHC	g/dl	<b>35.8</b>	<b>35.5</b>			<b>35.1</b>					± 3.0
	g/l	<b>358</b>	<b>355</b>			<b>351</b>					± 30
	mmol/l	<b>22.25</b>	<b>22.03</b>			<b>21.81</b>					± 1.86
IDR RDW	%	<b>14.0</b>	<b>14.0</b>			<b>13.0</b>					± 3.0
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>560</b>	<b>546</b>			<b>550</b>					± 65
VPM MPV	µm <sup>3</sup> ; fl	<b>7.9</b>	<b>7.7</b>			<b>7.3</b>					± 2.0
LYMPHO	#	<b>3.31</b>	<b>3.23</b>			<b>3.31</b>					± 1.00
	%	<b>16.0</b>	<b>16.0</b>			<b>16.0</b>					± 5.0
MONO	#	<b>1.35</b>	<b>1.41</b>			<b>1.14</b>					± 0.80
	%	<b>6.5</b>	<b>7.0</b>			<b>5.5</b>					± 4.0
GRANULO	#	<b>16.04</b>	<b>15.55</b>			<b>16.25</b>					± 1.40
	%	<b>77.5</b>	<b>77.0</b>			<b>78.5</b>					± 7.0