

LOT MX 438 L
 Rev 1

CONTROL L

(Exp.)

2023-02-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.2	2.2			2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.25	2.27			2.22					± 0.17
HB HGB	g/dl	6.0	6.0			6.0					± 0.5
	g/l	60	60			60					± 5
	mmol/l	3.73	3.73			3.73					± 0.31
HT HCT	%	16.2	16.3			16.0					± 2.0
	l/l	0.162	0.163			0.160					± 0.020
VGM MCV	µm ³ ; fl	72	72			72					± 4
TGMH MCH	pg	26.7	26.4			27.0					± 2.0
	fmol	1.66	1.64			1.68					± 0.12
CCMH MCHC	g/dl	37.0	36.7			37.5					± 3.0
	g/l	370	367			375					± 30
	mmol/l	23.00	22.80			23.31					± 1.86
IDR RDW	%	15.5	14.0			14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	81	85			81					± 22
VPM MPV	µm ³ ; fl	8.0	8.0			7.5					± 2.0
LYMPHO	#	1.44	1.45			1.37					± 0.40
	%	65.5	66.0			65.0					± 8.0
MONO	#	0.21	0.22			0.21					± 0.20
	%	9.5	10.0			10.0					± 6.0
GRANULO	#	0.55	0.53			0.53					± 0.40
	%	25.0	24.0			25.0					± 7.0

LOT		MX 438 N		CONTROL N						(Exp.) 2023-02-05 (YYYY - MM - DD)	
Rev 1		ABX Lyse									
PARAMETRES PARAMETERS	UNITES UNITS	MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	TOLERANCES TOLERANCE
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	7.6	7.6			7.7					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.60	4.59			4.57					± 0.26
HB HGB	g/dl	13.4	13.4			13.4					± 0.7
	g/l	134	134			134					± 7
	mmol/l	8.32	8.32			8.32					± 0.43
HT HCT	%	37.3	37.2			37.5					± 2.7
	l/l	0.373	0.372			0.375					± 0.027
VGM MCV	µm ³ ; fl	81	81			82					± 4
TGMH MCH	pg	29.1	29.2			29.3					± 2.0
	fmol	1.81	1.81			1.82					± 0.12
CCMH MCHC	g/dl	36.0	36.0			35.8					± 3.0
	g/l	360	360			358					± 30
	mmol/l	22.33	22.38			22.21					± 1.86
IDR RDW	%	14.5	13.0			13.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	259	257			248					± 40
VPM MPV	µm ³ ; fl	8.1	8.3			7.6					± 2.0
LYMPHO	#	2.58	2.58			2.66					± 0.40
	%	34.0	34.0			34.5					± 5.0
MONO	#	0.53	0.57			0.54					± 0.30
	%	7.0	7.5			7.0					± 4.0
GRANULO	#	4.48	4.45			4.50					± 0.60
	%	59.0	58.5			58.5					± 6.0

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104144

LOT

MX 438 H
 Rev 1

CONTROL H

(Exp.)

2023-02-05
 (YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST	CRP 200			60 8P - 60 18P ADVIA 60					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.6	20.2			20.8					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.70	5.67			5.66					± 0.32
HB HGB	g/dl	18.1	18.2			18.2					± 0.9
	g/l	181	182			182					± 9
	mmol/l	11.24	11.30			11.30					± 0.56
HT HCT	%	50.7	50.5			50.9					± 3.4
	l/l	0.507	0.505			0.509					± 0.034
VGM MCV	µm ³ ; fl	89	89			90					± 4
TGMH MCH	pg	31.8	32.1			32.2					± 2.0
	fmol	1.97	1.99			2.00					± 0.12
CCMH MCHC	g/dl	35.7	36.1			35.7					± 3.0
	g/l	357	361			357					± 30
	mmol/l	22.16	22.40			22.19					± 1.86
IDR RDW	%	14.0	13.0			12.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	543	536			528					± 65
VPM MPV	µm ³ ; fl	7.7	7.8			7.2					± 2.0
LYMPHO	#	3.30	3.23			3.33					± 1.00
	%	16.0	16.0			16.0					± 5.0
MONO	#	1.03	1.21			1.04					± 0.80
	%	5.0	6.0			5.0					± 4.0
GRANULO	#	16.27	15.76			16.43					± 1.40
	%	79.0	78.0			79.0					± 7.0