

LOT MX 436 L
Rev 1

CONTROL L

 (Exp.)

2022-10-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-------------|---------|--------|--------|----------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 | CRP 200 | | | 60 8P - 60 18P | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 2.1 | 2.0 | | | 2.1 | | | | | ± 0.4 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.25 | 2.28 | | | 2.20 | | | | | ± 0.17 |
| HB HGB | g/dl | 5.8 | 6.0 | | | 5.6 | | | | | ± 0.5 |
| | g/l | 58 | 60 | | | 56 | | | | | ± 5 |
| | mmol/l | 3.60 | 3.71 | | | 3.47 | | | | | ± 0.31 |
| HT HCT | % | 16.2 | 16.4 | | | 16.1 | | | | | ± 2.0 |
| | l/l | 0.162 | 0.164 | | | 0.161 | | | | | ± 0.020 |
| VGM MCV | µm ³ ; fl | 72 | 72 | | | 73 | | | | | ± 4 |
| TGMH MCH | pg | 25.8 | 26.2 | | | 25.4 | | | | | ± 2.0 |
| | fmol | 1.60 | 1.63 | | | 1.58 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 35.8 | 36.4 | | | 34.7 | | | | | ± 3.0 |
| | g/l | 358 | 364 | | | 347 | | | | | ± 30 |
| | mmol/l | 22.23 | 22.61 | | | 21.58 | | | | | ± 1.86 |
| IDR RDW | % | 15.5 | 14.0 | | | 15.0 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 75 | 74 | | | 75 | | | | | ± 22 |
| VPM MPV | µm ³ ; fl | 8.3 | 8.1 | | | 7.5 | | | | | ± 2.0 |
| LYMPHO | # | 1.30 | 1.31 | | | 1.30 | | | | | ± 0.40 |
| | % | 64.0 | 65.7 | | | 63.5 | | | | | ± 8.0 |
| MONO | # | 0.20 | 0.20 | | | 0.20 | | | | | ± 0.20 |
| | % | 8.0 | 10.0 | | | 8.5 | | | | | ± 6.0 |
| GRANULO | # | 0.60 | 0.49 | | | 0.60 | | | | | ± 0.40 |
| | % | 28.0 | 24.5 | | | 28.0 | | | | | ± 7.0 |

LOT

MX 436 N
Rev 1

CONTROL N

(Exp.)

2022-10-05
(YYYY-MM-DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|--------------|--------------|--------|--------|----------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 | CRP 200 | | | 60 8P - 60 18P | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 7.5 | 7.4 | | | 7.3 | | | | | ± 0.8 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 4.51 | 4.52 | | | 4.49 | | | | | ± 0.26 |
| HB HGB | g/dl | 13.5 | 13.7 | | | 13.2 | | | | | ± 0.7 |
| | g/l | 135 | 137 | | | 132 | | | | | ± 7 |
| | mmol/l | 8.38 | 8.51 | | | 8.22 | | | | | ± 0.43 |
| HT HCT | % | 37.4 | 38.0 | | | 38.2 | | | | | ± 2.7 |
| | l/l | 0.374 | 0.380 | | | 0.382 | | | | | ± 0.027 |
| VGM MCV | µm ³ ; fl | 83 | 84 | | | 85 | | | | | ± 4 |
| TGMH MCH | pg | 29.9 | 30.3 | | | 29.5 | | | | | ± 2.0 |
| | fmol | 1.86 | 1.88 | | | 1.83 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.1 | 36.1 | | | 34.7 | | | | | ± 3.0 |
| | g/l | 361 | 361 | | | 347 | | | | | ± 30 |
| | mmol/l | 22.40 | 22.41 | | | 21.53 | | | | | ± 1.86 |
| IDR RDW | % | 14.5 | 13.5 | | | 14.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 263 | 269 | | | 269 | | | | | ± 40 |
| VPM MPV | µm ³ ; fl | 8.3 | 8.0 | | | 8.5 | | | | | ± 2.0 |
| LYMPHO | # | 2.80 | 2.82 | | | 2.60 | | | | | ± 0.40 |
| | % | 37.0 | 38.2 | | | 36.0 | | | | | ± 5.0 |
| MONO | # | 0.50 | 0.53 | | | 0.50 | | | | | ± 0.30 |
| | % | 6.5 | 7.1 | | | 7.0 | | | | | ± 4.0 |
| GRANULO | # | 4.20 | 4.06 | | | 4.20 | | | | | ± 0.60 |
| | % | 56.5 | 54.9 | | | 57.0 | | | | | ± 6.0 |

Ref: TEMP-0830 Rev.42 FRONT / RECTO 1300104143

LOT MX 436 H
Rev 1

CONTROL H

 (Exp.)

2022-10-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | UNITES UNITS | Minilysebio | | | | | | | | TOLERANCES TOLERANCE | |
|--------------------------|--|-------------|---------|--------|--------|----------------|----------------|--------|--------|-------------------------|---------|
| | | MICROS | MICROS | MICROS | MICROS | MICROS | ARGOS - HELIOS | PENTRA | PENTRA | | PENTRA |
| | | ES60 | CRP 200 | | | 60 8P - 60 18P | | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 19.5 | 19.4 | | | 19.3 | | | | | ± 1.7 |
| GR RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 5.63 | 5.66 | | | 5.58 | | | | | ± 0.32 |
| HB HGB | g/dl | 18.2 | 18.0 | | | 17.8 | | | | | ± 0.9 |
| | g/l | 182 | 180 | | | 178 | | | | | ± 9 |
| | mmol/l | 11.30 | 11.18 | | | 11.07 | | | | | ± 0.56 |
| HT HCT | % | 50.1 | 50.9 | | | 50.8 | | | | | ± 3.4 |
| | l/l | 0.501 | 0.509 | | | 0.508 | | | | | ± 0.034 |
| VGM MCV | µm ³ ; fl | 89 | 90 | | | 91 | | | | | ± 4 |
| TGMH MCH | pg | 32.3 | 31.8 | | | 31.9 | | | | | ± 2.0 |
| | fmol | 2.01 | 1.97 | | | 1.98 | | | | | ± 0.12 |
| CCMH MCHC | g/dl | 36.3 | 35.3 | | | 35.1 | | | | | ± 3.0 |
| | g/l | 363 | 353 | | | 351 | | | | | ± 30 |
| | mmol/l | 22.56 | 21.94 | | | 21.79 | | | | | ± 1.86 |
| IDR RDW | % | 13.5 | 13.0 | | | 13.5 | | | | | ± 3.0 |
| PLAQ. PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 490 | 505 | | | 505 | | | | | ± 65 |
| VPM MPV | µm ³ ; fl | 8.0 | 7.6 | | | 7.2 | | | | | ± 2.0 |
| LYMPHO | # | 3.30 | 3.29 | | | 3.30 | | | | | ± 1.00 |
| | % | 17.0 | 17.0 | | | 17.0 | | | | | ± 5.0 |
| MONO | # | 1.00 | 1.26 | | | 0.80 | | | | | ± 0.80 |
| | % | 5.0 | 6.5 | | | 4.0 | | | | | ± 4.0 |
| GRANULO | # | 15.20 | 14.83 | | | 15.20 | | | | | ± 1.40 |
| | % | 78.0 | 76.4 | | | 79.0 | | | | | ± 7.0 |