

# ABX Difftrol



**LOT** PX 436  
Rev 1

**CONTROL**

(Exp.) 2022-09-05  
(YYYY-MM-DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
			DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		<b>2.9</b>	<b>2.9</b>		± 0.4		<b>8.8</b>	<b>8.8</b>			± 1.0		<b>19.0</b>	<b>19.0</b>			± 2.2	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l		<b>2.39</b>	<b>2.39</b>		± 0.16		<b>4.74</b>	<b>4.74</b>			± 0.20		<b>5.26</b>	<b>5.26</b>			± 0.25	
HB HGB	g/dl		<b>6.4</b>	<b>6.4</b>		± 0.4		<b>13.6</b>	<b>13.6</b>			± 0.5		<b>16.1</b>	<b>16.1</b>			± 0.6	
	g/l		<b>64</b>	<b>64</b>		± 4		<b>136</b>	<b>136</b>			± 5		<b>161</b>	<b>161</b>			± 6	
	mmol/l		<b>3.97</b>	<b>3.97</b>		± 0.25		<b>8.45</b>	<b>8.45</b>			± 0.31		<b>10.00</b>	<b>10.00</b>			± 0.37	
HT HCT	%		<b>19.4</b>	<b>19.4</b>		± 1.5		<b>40.3</b>	<b>40.3</b>			± 2.0		<b>47.3</b>	<b>47.3</b>			± 2.5	
	l/l		<b>0.194</b>	<b>0.194</b>		± 0.015		<b>0.403</b>	<b>0.403</b>			± 0.020		<b>0.473</b>	<b>0.473</b>			± 0.025	
VGM MCV	µm <sup>3</sup> ; fl		<b>81</b>	<b>81</b>		± 5		<b>85</b>	<b>85</b>			± 5		<b>90</b>	<b>90</b>			± 5	
TGMH MCH	pg		<b>26.8</b>	<b>26.8</b>		± 2.0		<b>28.7</b>	<b>28.7</b>			± 2.0		<b>30.6</b>	<b>30.6</b>			± 2.5	
	fmoi		<b>1.66</b>	<b>1.66</b>		± 0.12		<b>1.78</b>	<b>1.78</b>			± 0.12		<b>1.90</b>	<b>1.90</b>			± 0.16	
CCMH MCHC	g/dl		<b>33.1</b>	<b>33.1</b>		± 3.0		<b>33.8</b>	<b>33.8</b>			± 3.0		<b>34.0</b>	<b>34.0</b>			± 3.0	
	g/l		<b>331</b>	<b>331</b>		± 30		<b>338</b>	<b>338</b>			± 30		<b>340</b>	<b>340</b>			± 30	
	mmol/l		<b>20.53</b>	<b>20.53</b>		± 1.86		<b>20.96</b>	<b>20.96</b>			± 1.86		<b>21.12</b>	<b>21.12</b>			± 1.86	
IDR RDW	%		<b>15.5</b>	<b>15.5</b>		± 4.0		<b>15.5</b>	<b>15.5</b>			± 4.0		<b>15.5</b>	<b>15.5</b>			± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l		<b>72</b>	<b>72</b>		± 20		<b>240</b>	<b>240</b>			± 30		<b>480</b>	<b>480</b>			± 50	
VPM MPV	µm <sup>3</sup> ; fl		<b>10.3</b>	<b>10.3</b>		± 2.0		<b>12.0</b>	<b>12.0</b>			± 2.0		<b>11.2</b>	<b>11.2</b>			± 2.0	
NEUT	#		<b>1.45</b>	<b>1.45</b>		± 0.35		<b>4.58</b>	<b>4.58</b>			± 0.90		<b>13.70</b>	<b>13.70</b>			± 1.90	
	%		<b>49.9</b>	<b>49.9</b>		± 10.0		<b>52.1</b>	<b>52.1</b>			± 10.0		<b>71.9</b>	<b>71.9</b>			± 10.0	
LYMPHO	#		<b>0.83</b>	<b>0.83</b>		± 0.33		<b>3.01</b>	<b>3.01</b>			± 0.70		<b>3.00</b>	<b>3.00</b>			± 1.50	
	%		<b>28.7</b>	<b>28.7</b>		± 12.0		<b>34.2</b>	<b>34.2</b>			± 8.0		<b>15.8</b>	<b>15.8</b>			± 8.0	
MONO	#		<b>0.21</b>	<b>0.21</b>		± 0.21		<b>0.35</b>	<b>0.35</b>			± 0.35		<b>0.53</b>	<b>0.53</b>			± 0.53	
	%		<b>7.3</b>	<b>7.3</b>		± 7.3		<b>4.0</b>	<b>4.0</b>			± 4.0		<b>2.8</b>	<b>2.8</b>			± 2.8	
EOS	#		<b>0.26</b>	<b>0.26</b>		± 0.26		<b>0.39</b>	<b>0.39</b>			± 0.39		<b>1.05</b>	<b>1.05</b>			± 1.05	
	%		<b>8.9</b>	<b>8.9</b>		± 8.9		<b>4.4</b>	<b>4.4</b>			± 4.4		<b>5.5</b>	<b>5.5</b>			± 5.5	
BASO	#		<b>0.15</b>	<b>0.15</b>		± 0.15		<b>0.47</b>	<b>0.47</b>			± 0.47		<b>0.76</b>	<b>0.76</b>			± 0.76	
	%		<b>5.2</b>	<b>5.2</b>		± 5.2		<b>5.3</b>	<b>5.3</b>			± 5.3		<b>4.0</b>	<b>4.0</b>			± 4.0	

Ref: TEMP-0821 Rev.46 BACK / VERSO 1300090306