


**LOT** PX 437

**CONTROL**

 (Exp.) 2022-11-05  
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE	
		CONTROL				L	CONTROL				N	CONTROL					H
		PENTRA				TOLERANCES TOLERANCE	PENTRA				TOLERANCES TOLERANCE	PENTRA					TOLERANCES TOLERANCE
			DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.2	3.2			± 0.4	8.7	8.7			± 1.0	19.0	19.0			± 2.2	
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.38			± 0.16	4.70	4.70			± 0.20	5.33	5.33			± 0.25	
HB HGB	g/dl	6.2	6.2			± 0.4	13.7	13.7			± 0.5	16.8	16.8			± 0.6	
	g/l	62	62			± 4	137	137			± 5	168	168			± 6	
	mmol/l	3.85	3.85			± 0.25	8.51	8.51			± 0.31	10.43	10.43			± 0.37	
HT HCT	%	18.6	18.6			± 1.5	40.4	40.4			± 2.0	49.0	49.0			± 2.5	
	l/l	0.186	0.186			± 0.015	0.404	0.404			± 0.020	0.490	0.490			± 0.025	
VGM MCV	µm <sup>3</sup> ; fl	78	78			± 5	86	86			± 5	92	92			± 5	
TGMH MCH	pg	26.1	26.1			± 2.0	29.1	29.1			± 2.0	31.5	31.5			± 2.5	
	fmol	1.62	1.62			± 0.12	1.81	1.81			± 0.12	1.96	1.96			± 0.16	
CCMH MCHC	g/dl	33.4	33.4			± 3.0	33.9	33.9			± 3.0	34.3	34.3			± 3.0	
	g/l	334	334			± 30	339	339			± 30	343	343			± 30	
	mmol/l	20.74	20.74			± 1.86	21.05	21.05			± 1.86	21.28	21.28			± 1.86	
IDR RDW	%	17.0	17.0			± 4.0	15.5	15.5			± 4.0	14.5	14.5			± 4.0	
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	67	67			± 20	234	234			± 30	471	471			± 50	
VPM MPV	µm <sup>3</sup> ; fl	10.6	10.6			± 2.0	9.9	9.9			± 2.0	10.9	10.9			± 2.0	
NEUT	#	1.72	1.72			± 0.35	4.85	4.85			± 0.90	14.25	14.25			± 1.90	
	%	53.6	53.6			± 10.0	55.7	55.7			± 10.0	75.0	75.0			± 10.0	
LYMPHO	#	1.04	1.04			± 0.33	2.91	2.91			± 0.70	2.64	2.64			± 1.50	
	%	32.4	32.4			± 12.0	33.5	33.5			± 8.0	13.9	13.9			± 8.0	
MONO	#	0.08	0.08			± 0.08	0.17	0.17			± 0.17	0.38	0.38			± 0.38	
	%	2.6	2.6			± 2.6	2.0	2.0			± 2.0	2.0	2.0			± 2.0	
EOS	#	0.21	0.21			± 0.21	0.34	0.34			± 0.34	0.89	0.89			± 0.89	
	%	6.6	6.6			± 6.6	3.9	3.9			± 3.9	4.7	4.7			± 4.7	
BASO	#	0.15	0.15			± 0.15	0.43	0.43			± 0.43	0.84	0.84			± 0.84	
	%	4.8	4.8			± 4.8	4.9	4.9			± 4.9	4.4	4.4			± 4.4	

Ref: TEMP-0821 Rev.47 BACK / VERSO 1300090306