

 (Exp.) **2023-03-05**  
(YYYY-MM-DD)

**LOT** PX 439

**CONTROL**

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE				
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE	
		PENTRA						PENTRA						PENTRA						
			DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS						
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.9	2.9			± 0.4	8.7	8.7			± 1.0	19.0	19.0			± 2.2				
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.39	2.39			± 0.16	4.63	4.63			± 0.20	5.25	5.25			± 0.25				
HB HGB	g/dl	6.3	6.3			± 0.4	13.8	13.8			± 0.5	16.5	16.5			± 0.6				
	g/l	63	63			± 4	138	138			± 5	165	165			± 6				
	mmol/l	3.91	3.91			± 0.25	8.57	8.57			± 0.31	10.25	10.25			± 0.37				
HT HCT	%	18.9	18.9			± 1.5	40.3	40.3			± 2.0	47.8	47.8			± 2.5				
	l/l	0.189	0.189			± 0.015	0.403	0.403			± 0.020	0.478	0.478			± 0.025				
VGM MCV	µm <sup>3</sup> ; fl	79	79			± 5	87	87			± 5	91	91			± 5				
TGMH MCH	pg	26.4	26.4			± 2.0	29.8	29.8			± 2.0	31.4	31.4			± 2.5				
	fmol	1.64	1.64			± 0.12	1.85	1.85			± 0.12	1.95	1.95			± 0.16				
CCMH MCHC	g/dl	33.4	33.4			± 3.0	34.3	34.3			± 3.0	34.5	34.5			± 3.0				
	g/l	334	334			± 30	343	343			± 30	345	345			± 30				
	mmol/l	20.72	20.72			± 1.86	21.28	21.28			± 1.86	21.45	21.45			± 1.86				
IDR RDW	%	17.0	17.0			± 4.0	14.5	14.5			± 4.0	14.0	14.0			± 4.0				
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	74	74			± 20	263	263			± 30	499	499			± 50				
VPM MPV	µm <sup>3</sup> ; fl	9.0	9.0			± 2.0	8.5	8.5			± 2.0	8.8	8.8			± 2.0				
NEUT	#	1.55	1.55			± 0.35	4.58	4.58			± 0.90	13.79	13.79			± 1.90				
	%	53.4	53.4			± 10.0	52.6	52.6			± 10.0	72.6	72.6			± 10.0				
LYMPHO	#	0.92	0.92			± 0.33	3.23	3.23			± 0.70	3.17	3.17			± 1.50				
	%	31.7	31.7			± 12.0	37.1	37.1			± 8.0	16.7	16.7			± 8.0				
MONO	#	0.09	0.09			± 0.09	0.19	0.19			± 0.19	0.40	0.40			± 0.40				
	%	3.1	3.1			± 3.1	2.2	2.2			± 2.2	2.1	2.1			± 2.1				
EOS	#	0.19	0.19			± 0.19	0.29	0.29			± 0.29	0.76	0.76			± 0.76				
	%	6.7	6.7			± 6.7	3.3	3.3			± 3.3	4.0	4.0			± 4.0				
BASO	#	0.15	0.15			± 0.15	0.42	0.42			± 0.42	0.87	0.87			± 0.87				
	%	5.1	5.1			± 5.1	4.8	4.8			± 4.8	4.6	4.6			± 4.6				

1300090306  
BACK / VERSO  
Ref: TEMP-0821 Rev.47