

ABX Erytrol



LOT

NX 436

CONTROL



(Exp.) 2022-09-05

(YYYY - MM - DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA			PENTRA	PENTRA			PENTRA	PENTRA		
		DX120	DX NEXUS	DX120	DX NEXUS	DX120	DX NEXUS						
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	7.3	7.3		± 0.8	18.1	18.1		± 1.5	2.3	2.3		± 0.4
GBC CWBC	10 ³ /mm ³ ; 10 ⁹ /l	7.2	7.2		± 0.8	16.9	16.9		± 1.4	1.4	1.4		± 0.4
ERB	%	2.0	2.0		± 1.7	6.5	6.5		± 3.3	62.0	62.0		± 50.0
ERB	#	0.144	0.144		± 0.100	1.100	1.100		± 0.700	0.868	0.868		± 0.500

1219930076-C

Ref: TEMP-0822 Rev.35

GBC / CWBC

* les valeurs sont données à titre d'indication et doivent être utilisées comme RuO (Utilisation en Recherche Seulement) Indicative expected values, must be used as RuO (Research used only)