

ABX Erytrol



LOT

NX 437

Rev 1

CONTROL



(Exp.) 2022-11-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA	DX120		DX NEXUS	DX120	DX NEXUS		DX120	DX NEXUS		
		DX120	DX NEXUS										
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	7.7	7.7		± 0.8	18.1	18.1		± 1.5	2.7	2.7		± 0.4
GBC CWBC	10 ³ /mm ³ ; 10 ⁹ /l	7.3	7.3		± 0.8	16.6	16.6		± 1.4	1.6	1.6		± 0.4
ERB	%	4.6	4.6		± 1.7	9.1	9.1		± 3.3	66.2	66.2		± 22.0
ERB	#	0.336	0.336		± 0.100	1.511	1.511		± 0.700	1.059	1.059		± 0.300

1219930076-C

Ref: TEMP-0822 Rev.35

GBC / CWBC

* les valeurs sont données à titre d'indication et doivent être utilisées comme RuO (Utilisation en Recherche Seulement) Indicative expected values, must be used as RuO (Research used only)