

ABX Minotrol Retic



LOT RX 438
Rev 1

CONTROL

(Exp.) 2023-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	± 0.8	5.8	5.8	± 2.4	11.2	11.2	± 3.4			
RETIC	%	1.64	1.64	± 0.82	5.29	5.29	± 1.32	11.31	11.31	± 2.26			
RETIC	10 ⁶ /µl	0.057	0.057	± 0.029	0.183	0.183	± 0.046	0.394	0.394	± 0.079			
MFI MRI *	%	28.4	28.4	± 8.0	43.3	43.3	± 10.0	50.0	50.0	± 12.0			
GR RBC	10 ⁶ /µl	3.47	3.47	± 0.30	3.45	3.45	± 0.30	3.48	3.48	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.42 FRONT / RECTO 1300090305

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 438
Rev 1

CONTROL

(Exp.) 2023-01-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITS UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	1.5			± 0.8	5.8			± 2.4	11.2			± 3.4
RETIC	%	1.64			± 0.82	5.49			± 1.37	11.97			± 2.39
RETIC	10 ⁶ /µl	0.056			± 0.028	0.186			± 0.047	0.407			± 0.081
MFI MRI *	%	26.9			± 8.0	40.1			± 10.0	43.1			± 12.0
GR RBC	10 ⁶ /µl	3.40			± 0.30	3.38			± 0.30	3.40			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.42 BACK / VERSO 1300090305

* For research use only in Canada and USA