

ABX Minotrol Retic



LOT RX 439
Rev 1

CONTROL

(Exp.) 2023-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.4	1.4	± 0.7	5.9	5.9	± 2.4	11.0	11.0	± 3.3			
RETIC	%	1.51	1.51	± 0.76	5.43	5.43	± 1.36	9.69	9.69	± 1.94			
RETIC	10 ⁶ /µl	0.052	0.052	± 0.026	0.186	0.186	± 0.047	0.333	0.333	± 0.067			
MFI MRI *	%	29.7	29.7	± 8.0	46.1	46.1	± 10.0	50.7	50.7	± 12.0			
GR RBC	10 ⁶ /µl	3.43	3.43	± 0.30	3.43	3.43	± 0.30	3.44	3.44	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.42 FRONT / RECTO 1300090305

* For research use only in Canada and USA

ABX Minotrol Retic



LOT RX 439
Rev 1

CONTROL

(Exp.) 2023-03-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITS UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA				PENTRA				PENTRA			
XLR			XLR			XLR			XLR				
Méthode Manuelle / Manual RETIC	%	1.4			± 0.7	5.9			± 2.4	11.0			± 3.3
RETIC	%	1.65			± 0.83	6.04			± 1.51	11.56			± 2.31
RETIC	10 ⁶ /µl	0.056			± 0.028	0.204			± 0.051	0.390			± 0.078
MFI MRI *	%	27.6			± 8.0	38.7			± 10.0	37.5			± 12.0
GR RBC	10 ⁶ /µl	3.39			± 0.30	3.38			± 0.30	3.37			± 0.30
PIC RANGE	-	8			± 4	8			± 4	8			± 4

Ref: TEMP-0382 Rev.42 BACK / VERSO 1300090305

* For research use only in Canada and USA