


LOT PX 436
Rev 1

CONTROL

 (Exp.) **2022-09-05**
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE					
			CONTROL				L	CONTROL					N	CONTROL					H				
			PENTRA					PENTRA						PENTRA									
			XL80	XLR			XL80	XLR				XL80	XLR										
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0					± 0.4	8.6	8.6					± 1.0	18.7	18.7				± 2.2	
GR	RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.33					± 0.16	4.64	4.64					± 0.20	5.09	5.09				± 0.25	
HB	HGB	g/dl	6.2	6.2					± 0.4	13.6	13.6					± 0.5	16.3	16.3				± 0.6	
		g/l	62	62						± 4	136	136					± 5	163	163				± 6
		mmol/l	3.85	3.85						± 0.25	8.45	8.45					± 0.31	10.12	10.12				± 0.37
HT	HCT	%	18.9	18.9					± 1.5	39.4	39.4					± 2.0	46.3	46.3				± 2.5	
		l/l	0.189	0.189						± 0.015	0.394	0.394					± 0.020	0.463	0.463				± 0.025
VGM	MCV	µm ³ ; fl	81	81					± 5	85	85					± 5	91	91				± 5	
TGMH	MCH	pg	26.6	26.6					± 2.0	29.3	29.3					± 2.0	32.0	32.0				± 2.5	
		fmol	1.65	1.65						± 0.12	1.82	1.82					± 0.12	1.99	1.99				± 0.16
CCMH	MCHC	g/dl	32.9	32.9					± 3.0	34.5	34.5					± 3.0	35.2	35.2				± 3.0	
		g/l	329	329						± 30	345	345					± 30	352	352				± 30
		mmol/l	20.40	20.40						± 1.86	21.41	21.41					± 1.86	21.85	21.85				± 1.86
IDR	RDW	%	15.0	15.0					± 4.0	13.0	13.0					± 4.0	12.5	12.5				± 4.0	
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	69	69					± 20	220	220					± 30	449	449				± 50	
VPM	MPV	µm ³ ; fl	9.8	9.8					± 2.0	12.0	12.0					± 2.0	11.1	11.1				± 2.0	
NEUT		#	1.43	1.43					± 0.35	4.34	4.34					± 0.90	13.13	13.13				± 1.90	
		%	47.5	47.5						± 10.0	50.5	50.5					± 10.0	70.2	70.2				± 10.0
LYMPHO		#	1.10	1.10					± 0.33	3.40	3.40					± 0.70	3.48	3.48				± 1.50	
		%	36.7	36.7						± 12.0	39.5	39.5					± 8.0	18.6	18.6				± 8.0
MONO		#	0.19	0.19					± 0.19	0.30	0.30					± 0.30	0.43	0.43				± 0.43	
		%	6.4	6.4						± 6.4	3.5	3.5					± 3.5	2.3	2.3				± 2.3
EOS		#	0.20	0.20					± 0.20	0.32	0.32					± 0.32	0.92	0.92				± 0.92	
		%	6.8	6.8						± 6.8	3.7	3.7					± 3.7	4.9	4.9				± 4.9
BASO		#	0.08	0.08					± 0.08	0.24	0.24					± 0.24	0.75	0.75				± 0.75	
		%	2.6	2.6						± 2.6	2.8	2.8					± 2.8	4.0	4.0				± 4.0

ABX Difftrol



LOT PX 436
Rev 1

CONTROL

(Exp.) **2022-09-05**
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE					
			CONTROL			L	CONTROL				N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE						
			PENTRA				PENTRA						PENTRA										
			60 60C+ ES60	MS60	MSCRIP	60 60C+ ES60	MS60	MSCRIP			60 60C+ ES60	MS60	MSCRIP			60 60C+ ES60	MS60		MSCRIP				
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	2.9	3.0				± 0.4	8.7	8.7	8.6				± 1.0	19.1	18.7	18.5				± 2.2
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.41	2.38	2.34				± 0.16	4.68	4.63	4.60				± 0.20	5.11	5.11	5.01				± 0.25
HB	HGB	g/dl	6.3	6.2	6.2				± 0.4	13.6	13.5	13.5				± 0.5	16.3	16.3	16.1				± 0.6
		g/l	63	62	62				± 4	136	135	135				± 5	163	163	161				± 6
HT	HCT	mmol/l	3.91	3.85	3.85				± 0.25	8.45	8.38	8.38				± 0.31	10.12	10.12	10.00				± 0.37
		%	18.6	18.3	18.1				± 1.5	39.3	38.4	38.6				± 2.0	46.5	45.5	45.1				± 2.5
		l/l	0.186	0.183	0.181				± 0.015	0.393	0.384	0.386				± 0.020	0.465	0.455	0.451				± 0.025
VGM	MCV	µm ³ ; fl	77	77	77.5				± 5	84	83	84.0				± 5	91	90	90.0				± 5
TGMH	MCH	pg	26.1	26.1	26.5				± 2.0	29.1	29.2	29.3				± 2.0	31.9	31.9	32.1				± 2.5
		fmol	1.62	1.62	1.65				± 0.12	1.80	1.81	1.82				± 0.12	1.98	1.98	2.00				± 0.16
CCMH	MCHC	g/dl	33.9	33.8	34.2				± 3.0	34.6	35.1	34.9				± 3.0	35.1	35.4	35.7				± 3.0
		g/l	339	338	342				± 30	346	351	349				± 30	351	354	357				± 30
		mmol/l	21.08	21.01	21.23				± 1.86	21.48	21.82	21.70				± 1.86	21.77	22.01	22.17				± 1.86
IDR	RDW	%	15.0	15.0	14.0				± 4.0	13.0	12.5	12.0				± 4.0	12.0	12.0	11.0				± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	72	67				± 20	234	231	212				± 30	472	471	432				± 50
VPM	MPV	µm ³ ; fl	9.5	9.6	8.8				± 2.0	11.3	11.4	10.4				± 2.0	10.4	10.4	9.6				± 2.0
NEUT		#	1.41	1.33	1.41				± 0.35	4.25	4.21	4.26				± 0.90	13.24	12.90	12.95				± 1.90
		%	46.9	45.9	47.0				± 10.0	48.8	48.4	49.5				± 10.0	69.3	69.0	70.0				± 10.0
LYMPHO		#	1.10	1.10	1.14				± 0.33	3.56	3.58	3.53				± 0.70	3.63	3.59	3.52				± 1.50
		%	36.7	37.8	38.0				± 12.0	40.9	41.1	41.0				± 8.0	19.0	19.2	19.0				± 8.0
MONO		#	0.23	0.22	0.21				± 0.21	0.34	0.37	0.30				± 0.30	0.48	0.52	0.37				± 0.37
		%	7.7	7.7	7.0				± 7.0	3.9	4.3	3.5				± 3.5	2.5	2.8	2.0				± 2.0
EOS		#	0.18	0.17	0.15				± 0.15	0.32	0.30	0.26				± 0.26	0.99	0.94	0.93				± 0.93
		%	6.1	6.0	5.0				± 5.0	3.7	3.5	3.0				± 3.0	5.2	5.0	5.0				± 5.0
BASO		#	0.08	0.08	0.09				± 0.08	0.23	0.23	0.26				± 0.23	0.76	0.75	0.74				± 0.74
		%	2.6	2.6	3.0				± 2.6	2.7	2.7	3.0				± 2.7	4.0	4.0	4.0				± 4.0

Ref: TEMP-0821 Rev.46 FRONT / RECTO 1300090306