


**LOT** PX 438  
Rev 1


**CONTROL**

 (Exp.) 2023-01-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.1	3.1			± 0.4	8.5	8.5			± 1.0	18.1	18.1			± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.30	2.30			± 0.16	4.58	4.58			± 0.20	5.13	5.13			± 0.25			
HB HGB	g/dl	6.3	6.3			± 0.4	14.2	14.2			± 0.5	16.5	16.5			± 0.6			
	g/l	63	63			± 4	142	142			± 5	165	165			± 6			
	mmol/l	3.91	3.91			± 0.25	8.82	8.82			± 0.31	10.25	10.25			± 0.37			
HT HCT	%	18.6	18.6			± 1.5	40.8	40.8			± 2.0	46.2	46.2			± 2.5			
	l/l	0.186	0.186			± 0.015	0.408	0.408			± 0.020	0.462	0.462			± 0.025			
VGM MCV	µm <sup>3</sup> fl	81	81			± 5	89	89			± 5	90	90			± 5			
TGMH MCH	pg	27.4	27.4			± 2.0	31.0	31.0			± 2.0	32.2	32.2			± 2.5			
	fmol	1.70	1.70			± 0.12	1.93	1.93			± 0.12	2.00	2.00			± 0.16			
CCMH MCHC	g/dl	33.8	33.8			± 3.0	34.8	34.8			± 3.0	35.7	35.7			± 3.0			
	g/l	338	338			± 30	348	348			± 30	357	357			± 30			
	mmol/l	21.00	21.00			± 1.86	21.63	21.63			± 1.86	22.19	22.19			± 1.86			
IDR RDW	%	14.5	14.5			± 4.0	13.0	13.0			± 4.0	12.5	12.5			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	67	67			± 20	236	236			± 30	500	500			± 50			
VPM MPV	µm <sup>3</sup> fl	9.0	9.0			± 2.0	9.2	9.2			± 2.0	8.3	8.3			± 2.0			
NEUT	#	1.55	1.55			± 0.35	4.47	4.47			± 0.90	12.81	12.81			± 1.90			
	%	50.0	50.0			± 10.0	52.6	52.6			± 10.0	70.8	70.8			± 10.0			
LYMPHO	#	1.15	1.15			± 0.33	3.30	3.30			± 0.70	3.48	3.48			± 1.50			
	%	37.2	37.2			± 12.0	38.8	38.8			± 8.0	19.2	19.2			± 8.0			
MONO	#	0.10	0.10			± 0.10	0.24	0.24			± 0.24	0.36	0.36			± 0.36			
	%	3.3	3.3			± 3.3	2.8	2.8			± 2.8	2.0	2.0			± 2.0			
EOS	#	0.21	0.21			± 0.21	0.25	0.25			± 0.25	0.72	0.72			± 0.72			
	%	6.7	6.7			± 6.7	2.9	2.9			± 2.9	4.0	4.0			± 4.0			
BASO	#	0.09	0.09			± 0.09	0.25	0.25			± 0.25	0.72	0.72			± 0.72			
	%	2.8	2.8			± 2.8	2.9	2.9			± 2.9	4.0	4.0			± 4.0			

**LOT** PX 438  
Rev 1

**CONTROL**

 (Exp.) 2023-01-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE		
			CONTROL			L	CONTROL			N	CONTROL			H	TOLERANCES TOLERANCE					
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE						
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP							
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.0	3.0	3.1			± 0.4	8.5	8.6	8.4			± 1.0	18.6	18.5	18.4			± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.37	2.34	2.33			± 0.16	4.62	4.58	4.56			± 0.20	5.13	5.09	5.07			± 0.25
		g/dl	6.3	6.2	6.3			± 0.4	14.2	14.1	14.1			± 0.5	16.6	16.4	16.4			± 0.6
HB	HGB	g/l	63	62	63			± 0.4	142	141	141			± 5	166	164	164			± 6
		mmol/l	3.91	3.85	3.91			± 0.25	8.82	8.76	8.76			± 0.31	10.31	10.18	10.18			± 0.37
HT	HCT	%	18.7	18.3	17.9			± 1.5	41.1	40.3	40.1			± 2.0	47.2	46.3	46.1			± 2.5
		l/l	0.187	0.183	0.179			± 0.015	0.411	0.403	0.401			± 0.020	0.472	0.463	0.461			± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	79	78	77.0			± 5	89	88	88.0			± 5	92	91	91.0			± 5
TGMH	MCH	pg	26.6	26.5	27.0			± 2.0	30.7	30.8	30.9			± 2.0	32.4	32.2	32.3			± 2.5
		fmol	1.65	1.65	1.68			± 0.12	1.91	1.91	1.92			± 0.12	2.01	2.00	2.01			± 0.16
		g/dl	33.6	34.0	35.1			± 3.0	34.5	35.0	35.1			± 3.0	35.2	35.4	35.5			± 3.0
CCMH	MCHC	g/l	336	340	351			± 30	345	350	351			± 30	352	354	355			± 30
		mmol/l	20.90	21.09	21.81			± 1.86	21.45	21.73	21.82			± 1.86	21.84	21.99	22.07			± 1.86
IDR	RDW	%	15.5	14.5	14.0			± 4.0	12.5	12.0	11.5			± 4.0	12.0	11.5	11.0			± 4.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	69	65	67			± 20	240	234	230			± 30	493	489	480			± 50
VPM	MPV	µm <sup>3</sup> ; fl	8.4	8.8	8.3			± 2.0	8.6	8.7	8.4			± 2.0	7.8	7.8	7.5			± 2.0
		#	1.48	1.46	1.57			± 0.35	4.35	4.41	4.37			± 0.90	13.02	13.02	13.06			± 1.90
		%	49.3	48.7	50.5			± 10.0	51.2	51.3	52.0			± 10.0	70.0	70.4	71.0			± 10.0
		#	1.13	1.13	1.16			± 0.33	3.40	3.41	3.36			± 0.70	3.68	3.63	3.59			± 1.50
		%	37.8	37.8	37.5			± 12.0	40.0	39.7	40.0			± 8.0	19.8	19.6	19.5			± 8.0
		#	0.12	0.12	0.11			± 0.11	0.26	0.26	0.21			± 0.21	0.41	0.37	0.28			± 0.28
		%	4.0	3.9	3.5			± 3.5	3.1	3.0	2.5			± 2.5	2.2	2.0	1.5			± 1.5
		#	0.18	0.20	0.17			± 0.17	0.24	0.27	0.21			± 0.21	0.74	0.74	0.74			± 0.74
		%	6.1	6.7	5.5			± 5.5	2.8	3.1	2.5			± 2.5	4.0	4.0	4.0			± 4.0
		#	0.08	0.09	0.09			± 0.08	0.25	0.25	0.25			± 0.25	0.74	0.74	0.74			± 0.74
		%	2.8	2.9	3.0			± 2.8	2.9	2.9	3.0			± 2.9	4.0	4.0	4.0			± 4.0

1300090306 FRONT / RECTO Ref: TEMP-0821 Rev.47