


**LOT** PX 439  
Rev 1


**CONTROL**

 (Exp.) 2023-03-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		XL80	XLR					XL80	XLR					XL80	XLR				
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.0	3.0			± 0.4	8.5	8.5			± 1.0	18.6	18.6			± 2.2			
GR RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.34	2.34			± 0.16	4.53	4.53			± 0.20	5.10	5.10			± 0.25			
HB HGB	g/dl	6.1	6.1			± 0.4	13.7	13.7			± 0.5	16.4	16.4			± 0.6			
	g/l	61	61			± 4	137	137			± 5	164	164			± 6			
	mmol/l	3.79	3.79			± 0.25	8.51	8.51			± 0.31	10.18	10.18			± 0.37			
HT HCT	%	18.3	18.3			± 1.5	39.4	39.4			± 2.0	45.9	45.9			± 2.5			
	l/l	0.183	0.183			± 0.015	0.394	0.394			± 0.020	0.459	0.459			± 0.025			
VGM MCV	µm <sup>3</sup> ; fl	78	78			± 5	87	87			± 5	90	90			± 5			
TGMH MCH	pg	26.1	26.1			± 2.0	30.2	30.2			± 2.0	32.2	32.2			± 2.5			
	fmol	1.62	1.62			± 0.12	1.88	1.88			± 0.12	2.00	2.00			± 0.16			
CCMH MCHC	g/dl	33.4	33.4			± 3.0	34.8	34.8			± 3.0	35.7	35.7			± 3.0			
	g/l	334	334			± 30	348	348			± 30	357	357			± 30			
	mmol/l	20.75	20.75			± 1.86	21.59	21.59			± 1.86	22.19	22.19			± 1.86			
IDR RDW	%	15.5	15.5			± 4.0	13.0	13.0			± 4.0	13.0	13.0			± 4.0			
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	69	69			± 20	257	257			± 30	481	481			± 50			
VPM MPV	µm <sup>3</sup> ; fl	8.8	8.8			± 2.0	8.5	8.5			± 2.0	9.0	9.0			± 2.0			
NEUT	#	1.52	1.52			± 0.35	4.37	4.37			± 0.90	13.19	13.19			± 1.90			
	%	50.7	50.7			± 10.0	51.4	51.4			± 10.0	70.9	70.9			± 10.0			
LYMPHO	#	1.13	1.13			± 0.33	3.45	3.45			± 0.70	3.46	3.46			± 1.50			
	%	37.8	37.8			± 12.0	40.6	40.6			± 8.0	18.6	18.6			± 8.0			
MONO	#	0.07	0.07			± 0.07	0.17	0.17			± 0.17	0.37	0.37			± 0.37			
	%	2.4	2.4			± 2.4	2.0	2.0			± 2.0	2.0	2.0			± 2.0			
EOS	#	0.19	0.19			± 0.19	0.27	0.27			± 0.27	0.84	0.84			± 0.84			
	%	6.4	6.4			± 6.4	3.2	3.2			± 3.2	4.5	4.5			± 4.5			
BASO	#	0.08	0.08			± 0.08	0.24	0.24			± 0.24	0.74	0.74			± 0.74			
	%	2.7	2.7			± 2.7	2.8	2.8			± 2.8	4.0	4.0			± 4.0			

**LOT** PX 439  
Rev 1

**CONTROL**

 (Exp.) 2023-03-05  
(YYYY - MM - DD)

		ABX Lysebio																	
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE
		PENTRA						PENTRA						PENTRA					
		60 60C+ ES60	MS60	MSCR		60 60C+ ES60		MS60	MSCR		60 60C+ ES60	MS60		MSCR		60 60C+ ES60	MS60	MSCR	
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	2.9	2.9	3.0		± 0.4	8.5	8.5	8.5		± 1.0	19.1	18.7	18.8		± 2.2			
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.38	2.35	2.37		± 0.16	4.55	4.52	4.55		± 0.20	5.11	5.08	5.09		± 0.25			
HB HGB	g/dl	6.1	6.2	6.1		± 0.4	13.7	13.7	13.7		± 0.5	16.5	16.4	16.4		± 0.6			
	g/l	61	62	61		± 4	137	137	137		± 5	165	164	164		± 6			
	mmol/l	3.79	3.85	3.79		± 0.25	8.51	8.51	8.51		± 0.31	10.25	10.18	10.18		± 0.37			
HT HCT	%	18.1	17.9	17.7		± 1.5	39.1	38.9	38.9		± 2.0	47.0	46.2	46.3		± 2.5			
	l/l	0.181	0.179	0.177		± 0.015	0.391	0.389	0.389		± 0.020	0.470	0.462	0.463		± 0.025			
VGM MCV	µm <sup>3</sup> ·fl	76	76	74.5		± 5	86	86	85.5		± 5	92	91	91.0		± 5			
TGMH MCH	pg	25.6	26.4	25.7		± 2.0	30.1	30.3	30.1		± 2.0	32.3	32.3	32.2		± 2.5			
	fmol	1.59	1.64	1.60		± 0.12	1.87	1.88	1.87		± 0.12	2.01	2.00	2.00		± 0.16			
CCMH MCHC	g/dl	33.7	34.7	34.5		± 3.0	35.0	35.2	35.2		± 3.0	35.1	35.5	35.4		± 3.0			
	g/l	337	347	345		± 30	350	352	352		± 30	351	355	354		± 30			
	mmol/l	20.94	21.56	21.45		± 1.86	21.74	21.89	21.87		± 1.86	21.80	22.03	21.99		± 1.86			
IDR RDW	%	15.0	14.5	13.5		± 4.0	12.5	12.0	11.5		± 4.0	12.0	12.0	11.0		± 4.0			
PLAQ. PLTS	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	73	73	72		± 20	253	254	252		± 30	484	496	471		± 50			
VPM MPV	µm <sup>3</sup> ·fl	8.6	8.7	7.9		± 2.0	7.9	8.0	7.7		± 2.0	8.2	8.2	7.8		± 2.0			
NEUT	#	1.44	1.43	1.52		± 0.35	4.21	4.24	4.21		± 0.90	13.39	13.28	13.25		± 1.90			
	%	49.5	49.2	50.5		± 10.0	49.5	49.9	49.5		± 10.0	70.1	71.0	70.5		± 10.0			
LYMPHO	#	1.13	1.10	1.17		± 0.33	3.60	3.52	3.61		± 0.70	3.72	3.53	3.76		± 1.50			
	%	38.8	37.8	39.0		± 12.0	42.4	41.4	42.5		± 8.0	19.5	18.9	20.0		± 8.0			
MONO	#	0.09	0.09	0.08		± 0.08	0.20	0.19	0.17		± 0.17	0.42	0.37	0.28		± 0.28			
	%	3.1	3.0	2.5		± 2.5	2.3	2.2	2.0		± 2.0	2.2	2.0	1.5		± 1.5			
EOS	#	0.17	0.21	0.15		± 0.15	0.26	0.31	0.26		± 0.26	0.78	0.77	0.75		± 0.75			
	%	5.9	7.3	5.0		± 5.0	3.0	3.7	3.0		± 3.0	4.1	4.1	4.0		± 4.0			
BASO	#	0.08	0.08	0.09		± 0.08	0.24	0.24	0.26		± 0.24	0.78	0.75	0.75		± 0.75			
	%	2.7	2.7	3.0		± 2.7	2.8	2.8	3.0		± 2.8	4.1	4.0	4.0		± 4.0			

1300090306 FRONT / RECTO Ref: TEMP-0821 Rev.47