



LOT

MX 436 L
Rev 1

CONTROL

L

(Exp.)

2022-10-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE						
		MICROS														
		EMI CRP														
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l									2.1						± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l									2.26						± 0.15
HB HGB	g/dl									6.0						± 0.4
	g/l									60						± 4
	mmol/l									3.73						± 0.25
HT HCT	%									16.0						± 2.0
	l/l									0.160						± 0.020
VGM MCV	µm ³ ; fl									71						± 4
TGMH MCH	pg									26.5						± 2.0
	fmol									1.65						± 0.12
CCMH MCHC	g/dl									37.4						± 3.0
	g/l									374						± 30
	mmol/l									23.22						± 1.86
IDR RDW	%									14.5						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l									70						± 20
VPM MPV	µm ³ ; fl									7.7						± 2.0
LYMPHO	#									1.30						± 0.40
	%									62.5						± 8.0
MONO	#									0.20						± 0.20
	%									8.5						± 6.0
GRANULO	#									0.60						± 0.40
	%									29.0						± 7.0

LOT

MX 436 N
Rev 1

CONTROL

N

 (Exp.)

2022-10-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE									
		MICROS																	
		EMI CRP																	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l									7.7									± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l									4.44									± 0.18
HB HGB	g/dl									13.3									± 0.5
	g/l									133									± 5
	mmol/l									8.26									± 0.31
HT HCT	%									36.4									± 2.5
	l/l									0.364									± 0.025
VGM MCV	µm ³ ; fl									82									± 4
TGMH MCH	pg									30.0									± 2.0
	fmol									1.86									± 0.12
CCMH MCHC	g/dl									36.5									± 3.0
	g/l									365									± 30
	mmol/l									22.69									± 1.86
IDR RDW	%									13.0									± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l									271									± 40
VPM MPV	µm ³ ; fl									8.3									± 2.0
LYMPHO	#									2.80									± 0.40
	%									36.5									± 5.0
MONO	#									0.50									± 0.30
	%									7.0									± 4.0
GRANULO	#									4.40									± 0.60
	%									56.5									± 6.0

Ret: TEMP-0830 Rev.42 BACK / VERSO 1300104144

LOT MX 436 H
Rev 1

CONTROL H

 (Exp.)

2022-10-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			20.3						± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			5.52						± 0.20
HB HGB	g/dl			17.7						± 0.6
	g/l			177						± 6
	mmol/l			10.99						± 0.37
HT HCT	%			49.1						± 3.0
	l/l			0.491						± 0.030
VGM MCV	µm ³ ; fl			89						± 4
TGMH MCH	pg			32.1						± 2.0
	fmol			1.99						± 0.12
CCMH MCHC	g/dl			36.0						± 3.0
	g/l			360						± 30
	mmol/l			22.37						± 1.86
IDR RDW	%			12.5						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			499						± 45
VPM MPV	µm ³ ; fl			8.1						± 2.0
LYMPHO	#			3.70						± 1.00
	%			18.0						± 5.0
MONO	#			1.10						± 0.80
	%			5.5						± 4.0
GRANULO	#			15.50						± 1.40
	%			76.5						± 7.0