

LOT MX 438 L
Rev 1

CONTROL L

 (Exp.)

2023-02-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE									
		MICROS																	
		EMI CRP																	
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l									2.2									± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l									2.27									± 0.15
HB HGB	g/dl									6.0									± 0.4
	g/l									60									± 4
	mmol/l									3.73									± 0.25
HT HCT	%									16.1									± 2.0
	l/l									0.161									± 0.020
VGM MCV	µm ³ ; fl									71									± 4
TGMH MCH	pg									26.4									± 2.0
	fmol									1.64									± 0.12
CCMH MCHC	g/dl									37.2									± 3.0
	g/l									372									± 30
	mmol/l									23.12									± 1.86
IDR RDW	%									14.5									± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l									77									± 20
VPM MPV	µm ³ ; fl									7.7									± 2.0
LYMPHO	#									1.42									± 0.40
	%									64.5									± 8.0
MONO	#									0.19									± 0.20
	%									8.5									± 6.0
GRANULO	#									0.59									± 0.40
	%									27.0									± 7.0

LOT

MX 438 N
Rev 1

CONTROL

N

 (Exp.)

2023-02-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			7.9						± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			4.54						± 0.18
HB HGB	g/dl			13.3						± 0.5
	g/l			133						± 5
	mmol/l			8.26						± 0.31
HT HCT	%			36.8						± 2.5
	l/l			0.368						± 0.025
VGM MCV	µm ³ ; fl			81						± 4
TGMH MCH	pg			29.3						± 2.0
	fmol			1.82						± 0.12
CCMH MCHC	g/dl			36.2						± 3.0
	g/l			362						± 30
	mmol/l			22.46						± 1.86
IDR RDW	%			13.0						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			250						± 40
VPM MPV	µm ³ ; fl			8.2						± 2.0
LYMPHO	#			2.77						± 0.40
	%			35.0						± 5.0
MONO	#			0.51						± 0.30
	%			6.5						± 4.0
GRANULO	#			4.62						± 0.60
	%			58.5						± 6.0

Ret: TEMP-0830 Rev.42 BACK / VERSO 1300104144

LOT MX 438 H
Rev 1

CONTROL H

 (Exp.)

2023-02-05
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			21.1						± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			5.54						± 0.20
HB HGB	g/dl			17.9						± 0.6
	g/l			179						± 6
	mmol/l			11.12						± 0.37
HT HCT	%			49.3						± 3.0
	l/l			0.493						± 0.030
VGM MCV	µm ³ ; fl			89						± 4
TGMH MCH	pg			32.3						± 2.0
	fmol			2.01						± 0.12
CCMH MCHC	g/dl			36.3						± 3.0
	g/l			363						± 30
	mmol/l			22.54						± 1.86
IDR RDW	%			13.0						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			525						± 45
VPM MPV	µm ³ ; fl			8.2						± 2.0
LYMPHO	#			3.69						± 1.00
	%			17.5						± 5.0
MONO	#			1.16						± 0.80
	%			5.5						± 4.0
GRANULO	#			16.25						± 1.40
	%			77.0						± 7.0