


LOT		MX 439 L		CONTROL		L		 (Exp.)		2023-04-05 (YYYY-MM-DD)							
Rev 1		ABX Lysebio															
PARAMETRES PARAMETERS	UNITES UNITS	MICROS								TOLERANCES TOLERANCE							
		EMI CRP															
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l									2.0							± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l									2.34							± 0.15
HB HGB	g/dl									6.3							± 0.4
	g/l									63							± 4
	mmol/l									3.91							± 0.25
HT HCT	%									16.8							± 2.0
	l/l									0.168							± 0.020
VGM MCV	µm ³ ; fl									72							± 4
TGMH MCH	pg									26.9							± 2.0
	fmol									1.67							± 0.12
CCMH MCHC	g/dl									37.4							± 3.0
	g/l									374							± 30
	mmol/l									23.22							± 1.86
IDR RDW	%									14.5							± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l									73							± 20
VPM MPV	µm ³ ; fl									8.5							± 2.0
LYMPHO	#									1.27							± 0.40
	%									63.5							± 8.0
MONO	#									0.15							± 0.20
	%									7.5							± 6.0
GRANULO	#									0.58							± 0.40
	%									29.0							± 7.0

LOT

MX 439 N
Rev 1

CONTROL N

(Exp.)

2023-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE							
		MICROS															
		EMI CRP															
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l									7.7							± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l									4.46							± 0.18
HB HGB	g/dl									13.3							± 0.5
	g/l									133							± 5
	mmol/l									8.26							± 0.31
HT HCT	%									36.6							± 2.5
	l/l									0.366							± 0.025
VGM MCV	µm ³ ; fl									82							± 4
TGMH MCH	pg									29.8							± 2.0
	fmol									1.85							± 0.12
CCMH MCHC	g/dl									36.4							± 3.0
	g/l									364							± 30
	mmol/l									22.58							± 1.86
IDR RDW	%									13.5							± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l									274							± 40
VPM MPV	µm ³ ; fl									8.5							± 2.0
LYMPHO	#									2.62							± 0.40
	%									34.0							± 5.0
MONO	#									0.58							± 0.30
	%									7.5							± 4.0
GRANULO	#									4.50							± 0.60
	%									58.5							± 6.0

Ret: TEMP-0830 Rev.42 BACK / VERSO 1300104144

LOT

MX 439 H
Rev 1

CONTROL H

 (Exp.)

2023-04-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio								TOLERANCES TOLERANCE
				MICROS EMI CRP						
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l			20.4						± 1.6
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l			5.39						± 0.20
HB HGB	g/dl			17.4						± 0.6
	g/l			174						± 6
	mmol/l			10.81						± 0.37
HT HCT	%			48.0						± 3.0
	l/l			0.480						± 0.030
VGM MCV	µm ³ ; fl			89						± 4
TGMH MCH	pg			32.3						± 2.0
	fmol			2.00						± 0.12
CCMH MCHC	g/dl			36.3						± 3.0
	g/l			363						± 30
	mmol/l			22.52						± 1.86
IDR RDW	%			13.0						± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l			510						± 45
VPM MPV	µm ³ ; fl			8.5						± 2.0
LYMPHO	#			3.67						± 1.00
	%			18.0						± 5.0
MONO	#			1.12						± 0.80
	%			5.5						± 4.0
GRANULO	#			15.61						± 1.40
	%			76.5						± 7.0