

# ABX Minotrol CRP



**LOT**

**MC 425**

**CONTROL**



**(Exp.) 2023-01-05**

(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio															
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA			
				EMI CRP	LC-767G	MS CRP			EMI CRP	LC-767G	MS CRP			EMI CRP	LC-767G	MS CRP			
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.1</b>	<b>2.1</b>	<b>2.2</b>		± 0.4		<b>7.9</b>	<b>7.9</b>	<b>8.2</b>			± 1.0	<b>20.8</b>	<b>21.6</b>	<b>21.8</b>		
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.27</b>	<b>2.26</b>	<b>2.24</b>		± 0.15	<b>4.59</b>	<b>4.59</b>	<b>4.60</b>		± 0.18	<b>5.39</b>	<b>5.43</b>	<b>5.49</b>		± 0.25		
HB	HGB	g/dl	<b>6.2</b>	<b>6.2</b>	<b>5.9</b>		± 0.4	<b>13.7</b>	<b>13.6</b>	<b>13.6</b>		± 0.5	<b>18.0</b>	<b>18.1</b>	<b>18.3</b>		± 0.7		
		g/l	<b>62</b>	<b>62</b>	<b>59</b>		± 4	<b>137</b>	<b>136</b>	<b>136</b>		± 5	<b>180</b>	<b>181</b>	<b>183</b>		± 7		
		mmol/l	<b>3.85</b>	<b>3.85</b>	<b>3.66</b>		± 0.25	<b>8.51</b>	<b>8.45</b>	<b>8.45</b>		± 0.31	<b>11.18</b>	<b>11.24</b>	<b>11.36</b>		± 0.44		
HT	HCT	%	<b>16.3</b>	<b>16.5</b>	<b>16.4</b>		± 2.0	<b>37.6</b>	<b>37.2</b>	<b>36.8</b>		± 2.5	<b>49.6</b>	<b>49.4</b>	<b>48.9</b>		± 3.0		
		l/l	<b>0.163</b>	<b>0.165</b>	<b>0.164</b>		± 0.020	<b>0.376</b>	<b>0.372</b>	<b>0.368</b>		± 0.025	<b>0.496</b>	<b>0.494</b>	<b>0.489</b>		± 0.030		
VGM	MCV	µm <sup>3</sup> ; fl	<b>72</b>	<b>73</b>	<b>73.0</b>		± 4	<b>82</b>	<b>81</b>	<b>80.0</b>		± 4	<b>92</b>	<b>91</b>	<b>89.0</b>		± 4		
TGMH	MCH	pg	<b>27.3</b>	<b>27.4</b>	<b>26.3</b>		± 2.4	<b>29.8</b>	<b>29.6</b>	<b>29.6</b>		± 2.4	<b>33.4</b>	<b>33.3</b>	<b>33.3</b>		± 2.8		
		fmol	<b>1.70</b>	<b>1.70</b>	<b>1.64</b>		± 0.15	<b>1.85</b>	<b>1.84</b>	<b>1.84</b>		± 0.15	<b>2.07</b>	<b>2.07</b>	<b>2.07</b>		± 0.18		
CCMH	MCHC	g/dl	<b>37.9</b>	<b>37.6</b>	<b>36.1</b>		± 3.0	<b>36.4</b>	<b>36.6</b>	<b>37.0</b>		± 3.0	<b>36.3</b>	<b>36.6</b>	<b>37.5</b>		± 3.0		
		g/l	<b>379</b>	<b>376</b>	<b>361</b>		± 30	<b>364</b>	<b>366</b>	<b>370</b>		± 30	<b>363</b>	<b>366</b>	<b>375</b>		± 30		
		mmol/l	<b>23.56</b>	<b>23.34</b>	<b>22.41</b>		± 1.86	<b>22.60</b>	<b>22.72</b>	<b>22.95</b>		± 1.86	<b>22.54</b>	<b>22.75</b>	<b>23.26</b>		± 1.86		
IDR	RDW	%	<b>13.5</b>	<b>13.0</b>	<b>14.5</b>		± 3.0	<b>13.0</b>	<b>12.0</b>	<b>13.0</b>		± 3.0	<b>12.5</b>	<b>12.0</b>	<b>12.5</b>		± 3.0		
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>76</b>	<b>79</b>	<b>68</b>		± 20	<b>271</b>	<b>271</b>	<b>263</b>		± 40	<b>511</b>	<b>514</b>	<b>506</b>		± 60		
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.1</b>	<b>8.9</b>	<b>9.4</b>		± 2.0	<b>8.4</b>	<b>8.2</b>	<b>8.9</b>		± 2.0	<b>8.7</b>	<b>8.6</b>	<b>8.9</b>		± 2.0		
LYMPHO		#	<b>1.3</b>	<b>1.3</b>	N/A		± 0.2	<b>2.7</b>	<b>2.7</b>	N/A		± 0.5	<b>3.9</b>	<b>3.8</b>	N/A		± 1.1		
		%	<b>64.0</b>	<b>63.0</b>	N/A		± 8.0	<b>34.5</b>	<b>34.0</b>	N/A		± 6.0	<b>18.5</b>	<b>17.5</b>	N/A		± 5.0		
MONO		#	<b>0.2</b>	<b>0.2</b>	N/A		± 0.2	<b>0.6</b>	<b>0.6</b>	N/A		± 0.4	<b>1.4</b>	<b>1.4</b>	N/A		± 1.0		
		%	<b>8.5</b>	<b>8.5</b>	N/A		± 7.0	<b>7.0</b>	<b>7.0</b>	N/A		± 5.0	<b>6.5</b>	<b>6.5</b>	N/A		± 5.0		
GRANULO		#	<b>0.6</b>	<b>0.6</b>	N/A		± 0.2	<b>4.6</b>	<b>4.7</b>	N/A		± 0.5	<b>15.7</b>	<b>16.4</b>	N/A		± 1.5		
		%	<b>27.5</b>	<b>28.5</b>	N/A		± 7.0	<b>58.5</b>	<b>59.0</b>	N/A		± 6.0	<b>75.0</b>	<b>76.0</b>	N/A		± 7.0		
CRP		mg/dl	<b>0.54</b>	<b>0.57</b>	<b>0.54</b>		± 0.40	<b>2.77</b>	<b>2.95</b>	<b>2.48</b>		± 0.80	<b>4.52</b>	<b>4.67</b>	<b>4.63</b>		± 1.50		
		mg/l	<b>5.4</b>	<b>5.7</b>	<b>5.4</b>		± 4.0	<b>27.7</b>	<b>29.5</b>	<b>24.8</b>		± 8.0	<b>45.2</b>	<b>46.7</b>	<b>46.3</b>		± 15.0		

Ref: TEMP-1070 Rev.42 BACKVERSO 1300092170