

ABX Difftrol



LOT PX 436
Rev 1

CONTROL

(Exp.) 2022-09-05
(YYYY-MM-DD)

		ABX Lysebio														
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL			L	TOLERANCES TOLERANCE	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE
		PENTRA					PENTRA					PENTRA				
		DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS			
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	2.9	2.9		± 0.4	8.8	8.8			± 1.0	19.0	19.0			± 2.2	
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.39	2.39		± 0.16	4.74	4.74			± 0.20	5.26	5.26			± 0.25	
HB HGB	g/dl	6.4	6.4		± 0.4	13.6	13.6			± 0.5	16.1	16.1			± 0.6	
	g/l	64	64		± 4	136	136			± 5	161	161			± 6	
	mmol/l	3.97	3.97		± 0.25	8.45	8.45			± 0.31	10.00	10.00			± 0.37	
HT HCT	%	19.4	19.4		± 1.5	40.3	40.3			± 2.0	47.3	47.3			± 2.5	
	l/l	0.194	0.194		± 0.015	0.403	0.403			± 0.020	0.473	0.473			± 0.025	
VGM MCV	µm ³ ; fl	81	81		± 5	85	85			± 5	90	90			± 5	
TGMH MCH	pg	26.8	26.8		± 2.0	28.7	28.7			± 2.0	30.6	30.6			± 2.5	
	fmoi	1.66	1.66		± 0.12	1.78	1.78			± 0.12	1.90	1.90			± 0.16	
CCMH MCHC	g/dl	33.1	33.1		± 3.0	33.8	33.8			± 3.0	34.0	34.0			± 3.0	
	g/l	331	331		± 30	338	338			± 30	340	340			± 30	
	mmol/l	20.53	20.53		± 1.86	20.96	20.96			± 1.86	21.12	21.12			± 1.86	
IDR RDW	%	15.5	15.5		± 4.0	15.5	15.5			± 4.0	15.5	15.5			± 4.0	
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	72	72		± 20	240	240			± 30	480	480			± 50	
VPM MPV	µm ³ ; fl	10.3	10.3		± 2.0	12.0	12.0			± 2.0	11.2	11.2			± 2.0	
NEUT	#	1.45	1.45		± 0.35	4.58	4.58			± 0.90	13.70	13.70			± 1.90	
	%	49.9	49.9		± 10.0	52.1	52.1			± 10.0	71.9	71.9			± 10.0	
LYMPHO	#	0.83	0.83		± 0.33	3.01	3.01			± 0.70	3.00	3.00			± 1.50	
	%	28.7	28.7		± 12.0	34.2	34.2			± 8.0	15.8	15.8			± 8.0	
MONO	#	0.21	0.21		± 0.21	0.35	0.35			± 0.35	0.53	0.53			± 0.53	
	%	7.3	7.3		± 7.3	4.0	4.0			± 4.0	2.8	2.8			± 2.8	
EOS	#	0.26	0.26		± 0.26	0.39	0.39			± 0.39	1.05	1.05			± 1.05	
	%	8.9	8.9		± 8.9	4.4	4.4			± 4.4	5.5	5.5			± 5.5	
BASO	#	0.15	0.15		± 0.15	0.47	0.47			± 0.47	0.76	0.76			± 0.76	
	%	5.2	5.2		± 5.2	5.3	5.3			± 5.3	4.0	4.0			± 4.0	

Ref: TEMP-0821 Rev.46 BACK / VERSO 1300090306