


**LOT** PX 438

**CONTROL**

 (Exp.) 2023-01-05  
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio														TOLERANCES TOLERANCE				
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE	
		PENTRA						PENTRA						PENTRA						
			DX120 DF120	DX NEXUS DF NEXUS				DX120 DF120	DX NEXUS DF NEXUS					DX120 DF120	DX NEXUS DF NEXUS					
GB WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	3.0	3.0			± 0.4	8.9	8.9			± 1.0	19.0	19.0			± 2.2				
GR RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	2.35	2.35			± 0.16	4.67	4.67			± 0.20	5.26	5.26			± 0.25				
HB HGB	g/dl	6.4	6.4			± 0.4	14.3	14.3			± 0.5	16.5	16.5			± 0.6				
	g/l	64	64			± 4	143	143			± 5	165	165			± 6				
	mmol/l	3.97	3.97			± 0.25	8.88	8.88			± 0.31	10.25	10.25			± 0.37				
HT HCT	%	19.0	19.0			± 1.5	41.6	41.6			± 2.0	47.3	47.3			± 2.5				
	l/l	0.190	0.190			± 0.015	0.416	0.416			± 0.020	0.473	0.473			± 0.025				
VGM MCV	µm <sup>3</sup> ; fl	81	81			± 5	89	89			± 5	90	90			± 5				
TGMH MCH	pg	27.2	27.2			± 2.0	30.6	30.6			± 2.0	31.4	31.4			± 2.5				
	fmol	1.69	1.69			± 0.12	1.90	1.90			± 0.12	1.95	1.95			± 0.16				
CCMH MCHC	g/dl	33.6	33.6			± 3.0	34.4	34.4			± 3.0	34.9	34.9			± 3.0				
	g/l	336	336			± 30	344	344			± 30	349	349			± 30				
	mmol/l	20.88	20.88			± 1.86	21.37	21.37			± 1.86	21.64	21.64			± 1.86				
IDR RDW	%	17.5	17.5			± 4.0	14.5	14.5			± 4.0	14.0	14.0			± 4.0				
PLAQ. PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	70	70			± 20	243	243			± 30	494	494			± 50				
VPM MPV	µm <sup>3</sup> ; fl	9.3	9.3			± 2.0	9.1	9.1			± 2.0	8.1	8.1			± 2.0				
NEUT	#	1.49	1.49			± 0.35	4.73	4.73			± 0.90	13.60	13.60			± 1.90				
	%	49.6	49.6			± 10.0	53.1	53.1			± 10.0	71.8	71.8			± 10.0				
LYMPHO	#	1.05	1.05			± 0.33	3.29	3.29			± 0.70	3.27	3.27			± 1.50				
	%	35.1	35.1			± 12.0	37.0	37.0			± 8.0	17.2	17.2			± 8.0				
MONO	#	0.11	0.11			± 0.11	0.28	0.28			± 0.28	0.42	0.42			± 0.42				
	%	3.8	3.8			± 3.8	3.2	3.2			± 3.2	2.2	2.2			± 2.2				
EOS	#	0.20	0.20			± 0.20	0.24	0.24			± 0.24	0.74	0.74			± 0.74				
	%	6.7	6.7			± 6.7	2.7	2.7			± 2.7	3.9	3.9			± 3.9				
BASO	#	0.14	0.14			± 0.14	0.36	0.36			± 0.36	0.93	0.93			± 0.93				
	%	4.8	4.8			± 4.8	4.0	4.0			± 4.0	4.9	4.9			± 4.9				

Ref: TEMP-0821 Rev.47 BACK / VERSO 1300090306