

# ABX Erytrol



**LOT**

**NX 438**

Rev 1

**CONTROL**



**(Exp.) 2023-01-05**

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA			PENTRA	PENTRA			PENTRA	PENTRA		
		DX120	DX NEXUS		DX120	DX NEXUS		DX120	DX NEXUS				
GB WBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.5	7.5		± 0.8	18.0	18.0		± 1.5	2.5	2.5		± 0.4
GBC CWBC	10 <sup>3</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	7.1	7.1		± 0.8	16.5	16.5		± 1.4	1.4	1.4		± 0.4
ERB	%	5.0	5.0		± 1.7	9.0	9.0		± 3.3	74.1	74.1		± 24.0
ERB	#	0.355	0.355		± 0.100	1.485	1.485		± 0.700	1.037	1.037		± 0.300

1219930076-C

Ref: TEMP-0822 Rev.35

GBC / CWBC

\* les valeurs sont données à titre d'indication et doivent être utilisées comme RuO (Utilisation en Recherche Seulement) Indicative expected values, must be used as RuO ( Research used only)