

ABX Erytrol



LOT

NX 439

Rev 1

CONTROL



(Exp.) 2023-03-05

(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL		1	TOLERANCES TOLERANCE	CONTROL		2	TOLERANCES TOLERANCE	CONTROL		3	TOLERANCES TOLERANCE
		PENTRA	PENTRA	DX120		DX NEXUS	DX120	DX NEXUS		DX120	DX NEXUS		
		DX120	DX NEXUS										
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	7.7	7.7		± 0.8	17.8	17.8		± 1.5	2.8	2.8		± 0.4
GBC CWBC	10 ³ /mm ³ ; 10 ⁹ /l	7.4	7.4		± 0.8	16.4	16.4		± 1.4	1.7	1.7		± 0.4
ERB	%	3.2	3.2		± 1.7	8.3	8.3		± 3.3	65.1	65.1		± 22.0
ERB	#	0.237	0.237		± 0.100	1.361	1.361		± 0.700	1.107	1.107		± 0.300

1219930076-C

Ref: TEMP-0822 Rev.35

GBC / CWBC

* les valeurs sont données à titre d'indication et doivent être utilisées comme RuO (Utilisation en Recherche Seulement) Indicative expected values, must be used as RuO (Research used only)