

# ABX Minotrol Retic



**LOT** RX 436  
Rev 1

**CONTROL**

(Exp.) 2022-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL 1			TOLERANCES TOLERANCE	CONTROL 2			TOLERANCES TOLERANCE	CONTROL 3			TOLERANCES TOLERANCE
		PENTRA	PENTRA	PENTRA		PENTRA	PENTRA	PENTRA					
		DX120	DX Nexus	DX120		DX Nexus	DX120	DX Nexus					
Méthode Manuelle / Manual RETIC	%	1.5	1.5	± 0.8	5.5	5.5	± 2.2	10.4	10.4	± 3.1			
RETIC	%	1.63	1.63	± 0.82	5.02	5.02	± 1.26	11.45	11.45	± 2.29			
RETIC	10 <sup>6</sup> /µl	0.056	0.056	± 0.028	0.174	0.174	± 0.044	0.395	0.395	± 0.079			
MFI MRI *	%	30.3	30.3	± 8.0	44.9	44.9	± 10.0	52.5	52.5	± 12.0			
GR RBC	10 <sup>6</sup> /µl	3.45	3.45	± 0.30	3.46	3.46	± 0.30	3.45	3.45	± 0.30			
PIC RANGE	-	27.0	27.0	± 5.0	27.0	27.0	± 5.0	27.0	27.0	± 5.0			

Ref: TEMP-0382 Rev.42 FRONT / RECTO 1300090305

\* For research use only in Canada and USA

# ABX Minotrol Retic



**LOT** RX 436  
Rev 1

**CONTROL**

(Exp.) 2022-09-05  
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX											
		CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE	CONTROL			TOLERANCES TOLERANCE
		PENTRA		1		PENTRA		2		PENTRA		3	
		XLR				XLR				XLR			
Méthode Manuelle / Manual RETIC	%	<b>1.5</b>			± 0.8	<b>5.5</b>			± 2.2	<b>10.4</b>			± 3.1
RETIC	%	<b>1.70</b>			± 0.85	<b>5.26</b>			± 1.32	<b>11.99</b>			± 2.40
RETIC	10 <sup>6</sup> /µl	<b>0.058</b>			± 0.029	<b>0.180</b>			± 0.045	<b>0.408</b>			± 0.082
MFI MRI *	%	<b>29.3</b>			± 8.0	<b>41.7</b>			± 10.0	<b>46.7</b>			± 12.0
GR RBC	10 <sup>6</sup> /µl	<b>3.43</b>			± 0.30	<b>3.43</b>			± 0.30	<b>3.40</b>			± 0.30
PIC RANGE	-	<b>8</b>			± 4	<b>8</b>			± 4	<b>8</b>			± 4

Ref: TEMP-0382 Rev.42 BACK / VERSO 1300090305

\* For research use only in Canada and USA