

**LOT** PX 436  
Rev 1

**CONTROL**

 (Exp.) **2022-09-05**  
(YYYY - MM - DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE				
			CONTROL					L	CONTROL					N	CONTROL					H	TOLERANCES TOLERANCE	
			PENTRA						PENTRA						PENTRA							
			XL80	XLR					XL80	XLR					XL80	XLR						
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>3.0</b>	<b>3.0</b>				± 0.4	<b>8.6</b>	<b>8.6</b>				± 1.0	<b>18.7</b>	<b>18.7</b>				± 2.2		
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.33</b>	<b>2.33</b>				± 0.16	<b>4.64</b>	<b>4.64</b>				± 0.20	<b>5.09</b>	<b>5.09</b>				± 0.25		
HB	HGB	g/dl	<b>6.2</b>	<b>6.2</b>				± 0.4	<b>13.6</b>	<b>13.6</b>				± 0.5	<b>16.3</b>	<b>16.3</b>				± 0.6		
		g/l	<b>62</b>	<b>62</b>				± 4	<b>136</b>	<b>136</b>				± 5	<b>163</b>	<b>163</b>				± 6		
		mmol/l	<b>3.85</b>	<b>3.85</b>				± 0.25	<b>8.45</b>	<b>8.45</b>				± 0.31	<b>10.12</b>	<b>10.12</b>				± 0.37		
HT	HCT	%	<b>18.9</b>	<b>18.9</b>				± 1.5	<b>39.4</b>	<b>39.4</b>				± 2.0	<b>46.3</b>	<b>46.3</b>				± 2.5		
		l/l	<b>0.189</b>	<b>0.189</b>				± 0.015	<b>0.394</b>	<b>0.394</b>				± 0.020	<b>0.463</b>	<b>0.463</b>				± 0.025		
VGM	MCV	µm <sup>3</sup> ; fl	<b>81</b>	<b>81</b>				± 5	<b>85</b>	<b>85</b>				± 5	<b>91</b>	<b>91</b>				± 5		
TGMH	MCH	pg	<b>26.6</b>	<b>26.6</b>				± 2.0	<b>29.3</b>	<b>29.3</b>				± 2.0	<b>32.0</b>	<b>32.0</b>				± 2.5		
		fmol	<b>1.65</b>	<b>1.65</b>				± 0.12	<b>1.82</b>	<b>1.82</b>				± 0.12	<b>1.99</b>	<b>1.99</b>				± 0.16		
CCMH	MCHC	g/dl	<b>32.9</b>	<b>32.9</b>				± 3.0	<b>34.5</b>	<b>34.5</b>				± 3.0	<b>35.2</b>	<b>35.2</b>				± 3.0		
		g/l	<b>329</b>	<b>329</b>				± 30	<b>345</b>	<b>345</b>				± 30	<b>352</b>	<b>352</b>				± 30		
		mmol/l	<b>20.40</b>	<b>20.40</b>				± 1.86	<b>21.41</b>	<b>21.41</b>				± 1.86	<b>21.85</b>	<b>21.85</b>				± 1.86		
IDR	RDW	%	<b>15.0</b>	<b>15.0</b>				± 4.0	<b>13.0</b>	<b>13.0</b>				± 4.0	<b>12.5</b>	<b>12.5</b>				± 4.0		
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>69</b>	<b>69</b>				± 20	<b>220</b>	<b>220</b>				± 30	<b>449</b>	<b>449</b>				± 50		
VPM	MPV	µm <sup>3</sup> ; fl	<b>9.8</b>	<b>9.8</b>				± 2.0	<b>12.0</b>	<b>12.0</b>				± 2.0	<b>11.1</b>	<b>11.1</b>				± 2.0		
NEUT		#	<b>1.43</b>	<b>1.43</b>				± 0.35	<b>4.34</b>	<b>4.34</b>				± 0.90	<b>13.13</b>	<b>13.13</b>				± 1.90		
		%	<b>47.5</b>	<b>47.5</b>				± 10.0	<b>50.5</b>	<b>50.5</b>				± 10.0	<b>70.2</b>	<b>70.2</b>				± 10.0		
LYMPHO		#	<b>1.10</b>	<b>1.10</b>				± 0.33	<b>3.40</b>	<b>3.40</b>				± 0.70	<b>3.48</b>	<b>3.48</b>				± 1.50		
		%	<b>36.7</b>	<b>36.7</b>				± 12.0	<b>39.5</b>	<b>39.5</b>				± 8.0	<b>18.6</b>	<b>18.6</b>				± 8.0		
MONO		#	<b>0.19</b>	<b>0.19</b>				± 0.19	<b>0.30</b>	<b>0.30</b>				± 0.30	<b>0.43</b>	<b>0.43</b>				± 0.43		
		%	<b>6.4</b>	<b>6.4</b>				± 6.4	<b>3.5</b>	<b>3.5</b>				± 3.5	<b>2.3</b>	<b>2.3</b>				± 2.3		
EOS		#	<b>0.20</b>	<b>0.20</b>				± 0.20	<b>0.32</b>	<b>0.32</b>				± 0.32	<b>0.92</b>	<b>0.92</b>				± 0.92		
		%	<b>6.8</b>	<b>6.8</b>				± 6.8	<b>3.7</b>	<b>3.7</b>				± 3.7	<b>4.9</b>	<b>4.9</b>				± 4.9		
BASO		#	<b>0.08</b>	<b>0.08</b>				± 0.08	<b>0.24</b>	<b>0.24</b>				± 0.24	<b>0.75</b>	<b>0.75</b>				± 0.75		
		%	<b>2.6</b>	<b>2.6</b>				± 2.6	<b>2.8</b>	<b>2.8</b>				± 2.8	<b>4.0</b>	<b>4.0</b>				± 4.0		

# ABX Difftrol



**LOT** PX 436  
Rev 1

**CONTROL**

(Exp.) **2022-09-05**  
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE	
			CONTROL			L	CONTROL				N	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE	PENTRA			TOLERANCES TOLERANCE					
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP						
GB	WBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>3.0</b>	<b>2.9</b>	<b>3.0</b>		± 0.4	<b>8.7</b>	<b>8.7</b>	<b>8.6</b>			± 1.0	<b>19.1</b>	<b>18.7</b>	<b>18.5</b>			± 2.2
GR	RBC	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>12</sup> /l	<b>2.41</b>	<b>2.38</b>	<b>2.34</b>		± 0.16	<b>4.68</b>	<b>4.63</b>	<b>4.60</b>			± 0.20	<b>5.11</b>	<b>5.11</b>	<b>5.01</b>			± 0.25
HB	HGB	g/dl	<b>6.3</b>	<b>6.2</b>	<b>6.2</b>		± 0.4	<b>13.6</b>	<b>13.5</b>	<b>13.5</b>			± 0.5	<b>16.3</b>	<b>16.3</b>	<b>16.1</b>			± 0.6
		g/l	<b>63</b>	<b>62</b>	<b>62</b>		± 4	<b>136</b>	<b>135</b>	<b>135</b>			± 5	<b>163</b>	<b>163</b>	<b>161</b>			± 6
HT	HCT	mmol/l	<b>3.91</b>	<b>3.85</b>	<b>3.85</b>		± 0.25	<b>8.45</b>	<b>8.38</b>	<b>8.38</b>			± 0.31	<b>10.12</b>	<b>10.12</b>	<b>10.00</b>			± 0.37
		%	<b>18.6</b>	<b>18.3</b>	<b>18.1</b>		± 1.5	<b>39.3</b>	<b>38.4</b>	<b>38.6</b>			± 2.0	<b>46.5</b>	<b>45.5</b>	<b>45.1</b>			± 2.5
		l/l	<b>0.186</b>	<b>0.183</b>	<b>0.181</b>		± 0.015	<b>0.393</b>	<b>0.384</b>	<b>0.386</b>			± 0.020	<b>0.465</b>	<b>0.455</b>	<b>0.451</b>			± 0.025
VGM	MCV	µm <sup>3</sup> ; fl	<b>77</b>	<b>77</b>	<b>77.5</b>		± 5	<b>84</b>	<b>83</b>	<b>84.0</b>			± 5	<b>91</b>	<b>90</b>	<b>90.0</b>			± 5
TGMH	MCH	pg	<b>26.1</b>	<b>26.1</b>	<b>26.5</b>		± 2.0	<b>29.1</b>	<b>29.2</b>	<b>29.3</b>			± 2.0	<b>31.9</b>	<b>31.9</b>	<b>32.1</b>			± 2.5
		fmol	<b>1.62</b>	<b>1.62</b>	<b>1.65</b>		± 0.12	<b>1.80</b>	<b>1.81</b>	<b>1.82</b>			± 0.12	<b>1.98</b>	<b>1.98</b>	<b>2.00</b>			± 0.16
CCMH	MCHC	g/dl	<b>33.9</b>	<b>33.8</b>	<b>34.2</b>		± 3.0	<b>34.6</b>	<b>35.1</b>	<b>34.9</b>			± 3.0	<b>35.1</b>	<b>35.4</b>	<b>35.7</b>			± 3.0
		g/l	<b>339</b>	<b>338</b>	<b>342</b>		± 30	<b>346</b>	<b>351</b>	<b>349</b>			± 30	<b>351</b>	<b>354</b>	<b>357</b>			± 30
		mmol/l	<b>21.08</b>	<b>21.01</b>	<b>21.23</b>		± 1.86	<b>21.48</b>	<b>21.82</b>	<b>21.70</b>			± 1.86	<b>21.77</b>	<b>22.01</b>	<b>22.17</b>			± 1.86
IDR	RDW	%	<b>15.0</b>	<b>15.0</b>	<b>14.0</b>		± 4.0	<b>13.0</b>	<b>12.5</b>	<b>12.0</b>			± 4.0	<b>12.0</b>	<b>12.0</b>	<b>11.0</b>			± 4.0
PLAQ.	PLTS	10 <sup>9</sup> /mm <sup>3</sup> ; 10 <sup>9</sup> /l	<b>75</b>	<b>72</b>	<b>67</b>		± 20	<b>234</b>	<b>231</b>	<b>212</b>			± 30	<b>472</b>	<b>471</b>	<b>432</b>			± 50
VPM	MPV	µm <sup>3</sup> ; fl	<b>9.5</b>	<b>9.6</b>	<b>8.8</b>		± 2.0	<b>11.3</b>	<b>11.4</b>	<b>10.4</b>			± 2.0	<b>10.4</b>	<b>10.4</b>	<b>9.6</b>			± 2.0
NEUT		#	<b>1.41</b>	<b>1.33</b>	<b>1.41</b>		± 0.35	<b>4.25</b>	<b>4.21</b>	<b>4.26</b>			± 0.90	<b>13.24</b>	<b>12.90</b>	<b>12.95</b>			± 1.90
		%	<b>46.9</b>	<b>45.9</b>	<b>47.0</b>		± 10.0	<b>48.8</b>	<b>48.4</b>	<b>49.5</b>			± 10.0	<b>69.3</b>	<b>69.0</b>	<b>70.0</b>			± 10.0
LYMPHO		#	<b>1.10</b>	<b>1.10</b>	<b>1.14</b>		± 0.33	<b>3.56</b>	<b>3.58</b>	<b>3.53</b>			± 0.70	<b>3.63</b>	<b>3.59</b>	<b>3.52</b>			± 1.50
		%	<b>36.7</b>	<b>37.8</b>	<b>38.0</b>		± 12.0	<b>40.9</b>	<b>41.1</b>	<b>41.0</b>			± 8.0	<b>19.0</b>	<b>19.2</b>	<b>19.0</b>			± 8.0
MONO		#	<b>0.23</b>	<b>0.22</b>	<b>0.21</b>		± 0.21	<b>0.34</b>	<b>0.37</b>	<b>0.30</b>			± 0.30	<b>0.48</b>	<b>0.52</b>	<b>0.37</b>			± 0.37
		%	<b>7.7</b>	<b>7.7</b>	<b>7.0</b>		± 7.0	<b>3.9</b>	<b>4.3</b>	<b>3.5</b>			± 3.5	<b>2.5</b>	<b>2.8</b>	<b>2.0</b>			± 2.0
EOS		#	<b>0.18</b>	<b>0.17</b>	<b>0.15</b>		± 0.15	<b>0.32</b>	<b>0.30</b>	<b>0.26</b>			± 0.26	<b>0.99</b>	<b>0.94</b>	<b>0.93</b>			± 0.93
		%	<b>6.1</b>	<b>6.0</b>	<b>5.0</b>		± 5.0	<b>3.7</b>	<b>3.5</b>	<b>3.0</b>			± 3.0	<b>5.2</b>	<b>5.0</b>	<b>5.0</b>			± 5.0
BASO		#	<b>0.08</b>	<b>0.08</b>	<b>0.09</b>		± 0.08	<b>0.23</b>	<b>0.23</b>	<b>0.26</b>			± 0.23	<b>0.76</b>	<b>0.75</b>	<b>0.74</b>			± 0.74
		%	<b>2.6</b>	<b>2.6</b>	<b>3.0</b>		± 2.6	<b>2.7</b>	<b>2.7</b>	<b>3.0</b>			± 2.7	<b>4.0</b>	<b>4.0</b>	<b>4.0</b>			± 4.0

Ref: TEMP-0821 Rev.46 FRONT / RECTO 1300090306