

LOT PX 438

Rev 1

CONTROL



(Exp.)

2023-01-05

(YYYY - MM - DD)

| PARAMETRES PARAMETERS | | UNITES UNITS | ABX Lysebio | | | | | | | | | | | | | | | | | | | | |
|--------------------------|------|--|-------------|-------|--|--|---|-------------------------|---------|-------|-----|--|---|-------------------------|---------|-------|------|-----|---|-------------------------|--|--|--|
| | | | CONTROL | | | | L | TOLERANCES TOLERANCE | CONTROL | | | | N | TOLERANCES TOLERANCE | CONTROL | | | | H | TOLERANCES TOLERANCE | | | |
| | | | PENTRA | | | | | | PENTRA | | | | | | PENTRA | | | | | | | | |
| | | | XL80 | XLR | | | | | | XL80 | XLR | | | | | | XL80 | XLR | | | | | |
| GB | WBC | 10 ³ /mm ³ ; 10 ⁹ /l | 3.1 | 3.1 | | | | ± 0.4 | 8.5 | 8.5 | | | | ± 1.0 | 18.1 | 18.1 | | | | ± 2.2 | | | |
| GR | RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.30 | 2.30 | | | | ± 0.16 | 4.58 | 4.58 | | | | ± 0.20 | 5.13 | 5.13 | | | | ± 0.25 | | | |
| HB | HGB | g/dl | 6.3 | 6.3 | | | | ± 0.4 | 14.2 | 14.2 | | | | ± 0.5 | 16.5 | 16.5 | | | | ± 0.6 | | | |
| | | g/l | 63 | 63 | | | | ± 4 | 142 | 142 | | | | ± 5 | 165 | 165 | | | | ± 6 | | | |
| | | mmol/l | 3.91 | 3.91 | | | | ± 0.25 | 8.82 | 8.82 | | | | ± 0.31 | 10.25 | 10.25 | | | | ± 0.37 | | | |
| HT | HCT | % | 18.6 | 18.6 | | | | ± 1.5 | 40.8 | 40.8 | | | | ± 2.0 | 46.2 | 46.2 | | | | ± 2.5 | | | |
| | | l/l | 0.186 | 0.186 | | | | ± 0.015 | 0.408 | 0.408 | | | | ± 0.020 | 0.462 | 0.462 | | | | ± 0.025 | | | |
| VGM | MCV | µm ³ ·fl | 81 | 81 | | | | ± 5 | 89 | 89 | | | | ± 5 | 90 | 90 | | | | ± 5 | | | |
| TGMH | MCH | pg | 27.4 | 27.4 | | | | ± 2.0 | 31.0 | 31.0 | | | | ± 2.0 | 32.2 | 32.2 | | | | ± 2.5 | | | |
| | | fmol | 1.70 | 1.70 | | | | ± 0.12 | 1.93 | 1.93 | | | | ± 0.12 | 2.00 | 2.00 | | | | ± 0.16 | | | |
| CCMH | MCHC | g/dl | 33.8 | 33.8 | | | | ± 3.0 | 34.8 | 34.8 | | | | ± 3.0 | 35.7 | 35.7 | | | | ± 3.0 | | | |
| | | g/l | 338 | 338 | | | | ± 30 | 348 | 348 | | | | ± 30 | 357 | 357 | | | | ± 30 | | | |
| | | mmol/l | 21.00 | 21.00 | | | | ± 1.86 | 21.63 | 21.63 | | | | ± 1.86 | 22.19 | 22.19 | | | | ± 1.86 | | | |
| IDR | RDW | % | 14.5 | 14.5 | | | | ± 4.0 | 13.0 | 13.0 | | | | ± 4.0 | 12.5 | 12.5 | | | | ± 4.0 | | | |
| PLAQ. | PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 67 | 67 | | | | ± 20 | 236 | 236 | | | | ± 30 | 500 | 500 | | | | ± 50 | | | |
| VPM | MPV | µm ³ ·fl | 9.0 | 9.0 | | | | ± 2.0 | 9.2 | 9.2 | | | | ± 2.0 | 8.3 | 8.3 | | | | ± 2.0 | | | |
| NEUT | # | | 1.55 | 1.55 | | | | ± 0.35 | 4.47 | 4.47 | | | | ± 0.90 | 12.81 | 12.81 | | | | ± 1.90 | | | |
| | | % | 50.0 | 50.0 | | | | ± 10.0 | 52.6 | 52.6 | | | | ± 10.0 | 70.8 | 70.8 | | | | ± 10.0 | | | |
| LYMPHO | # | | 1.15 | 1.15 | | | | ± 0.33 | 3.30 | 3.30 | | | | ± 0.70 | 3.48 | 3.48 | | | | ± 1.50 | | | |
| | | % | 37.2 | 37.2 | | | | ± 12.0 | 38.8 | 38.8 | | | | ± 8.0 | 19.2 | 19.2 | | | | ± 8.0 | | | |
| MONO | # | | 0.10 | 0.10 | | | | ± 0.10 | 0.24 | 0.24 | | | | ± 0.24 | 0.36 | 0.36 | | | | ± 0.36 | | | |
| | | % | 3.3 | 3.3 | | | | ± 3.3 | 2.8 | 2.8 | | | | ± 2.8 | 2.0 | 2.0 | | | | ± 2.0 | | | |
| EOS | # | | 0.21 | 0.21 | | | | ± 0.21 | 0.25 | 0.25 | | | | ± 0.25 | 0.72 | 0.72 | | | | ± 0.72 | | | |
| | | % | 6.7 | 6.7 | | | | ± 6.7 | 2.9 | 2.9 | | | | ± 2.9 | 4.0 | 4.0 | | | | ± 4.0 | | | |
| BASO | # | | 0.09 | 0.09 | | | | ± 0.09 | 0.25 | 0.25 | | | | ± 0.25 | 0.72 | 0.72 | | | | ± 0.72 | | | |
| | | % | 2.8 | 2.8 | | | | ± 2.8 | 2.9 | 2.9 | | | | ± 2.9 | 4.0 | 4.0 | | | | ± 4.0 | | | |

