


LOT PX 437
Rev 2


CONTROL

 (Exp.) 2022-11-05
(YYYY - MM - DD)

| | | ABX Lysebio | | | | | | | | | | | | | | | |
|--------------------------|--|-------------|-------|--|--|---------|---------|-------|--|--|---------|---------|-------|--|--|---------|-------------------------|
| PARAMETRES PARAMETERS | UNITES UNITS | CONTROL | | | | L | CONTROL | | | | N | CONTROL | | | | H | TOLERANCES TOLERANCE |
| | | PENTRA | | | | | PENTRA | | | | | PENTRA | | | | | |
| | | XL80 | XLR | | | | XL80 | XLR | | | | XL80 | XLR | | | | |
| GB WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 3.2 | 3.2 | | | ± 0.4 | 8.5 | 8.5 | | | ± 1.0 | 18.4 | 18.4 | | | ± 2.2 | |
| GR RBC | 10 ⁹ /mm ³ ; 10 ¹² /l | 2.29 | 2.29 | | | ± 0.16 | 4.59 | 4.59 | | | ± 0.20 | 5.14 | 5.14 | | | ± 0.25 | |
| HB HGB | g/dl | 6.0 | 6.0 | | | ± 0.4 | 13.6 | 13.6 | | | ± 0.5 | 16.8 | 16.8 | | | ± 0.6 | |
| | g/l | 60 | 60 | | | ± 4 | 136 | 136 | | | ± 5 | 168 | 168 | | | ± 6 | |
| | mmol/l | 3.73 | 3.73 | | | ± 0.25 | 8.45 | 8.45 | | | ± 0.31 | 10.43 | 10.43 | | | ± 0.37 | |
| HT HCT | % | 17.9 | 17.9 | | | ± 1.5 | 39.5 | 39.5 | | | ± 2.0 | 47.8 | 47.8 | | | ± 2.5 | |
| | l/l | 0.179 | 0.179 | | | ± 0.015 | 0.395 | 0.395 | | | ± 0.020 | 0.478 | 0.478 | | | ± 0.025 | |
| VGM MCV | µm ³ fl | 78 | 78 | | | ± 5 | 86 | 86 | | | ± 5 | 93 | 93 | | | ± 5 | |
| TGMH MCH | pg | 26.2 | 26.2 | | | ± 2.0 | 29.6 | 29.6 | | | ± 2.0 | 32.7 | 32.7 | | | ± 2.5 | |
| | fmol | 1.63 | 1.63 | | | ± 0.12 | 1.84 | 1.84 | | | ± 0.12 | 2.03 | 2.03 | | | ± 0.16 | |
| CCMH MCHC | g/dl | 33.6 | 33.6 | | | ± 3.0 | 34.5 | 34.5 | | | ± 3.0 | 35.1 | 35.1 | | | ± 3.0 | |
| | g/l | 336 | 336 | | | ± 30 | 345 | 345 | | | ± 30 | 351 | 351 | | | ± 30 | |
| | mmol/l | 20.86 | 20.86 | | | ± 1.86 | 21.40 | 21.40 | | | ± 1.86 | 21.83 | 21.83 | | | ± 1.86 | |
| IDR RDW | % | 15.0 | 15.0 | | | ± 4.0 | 13.0 | 13.0 | | | ± 4.0 | 12.0 | 12.0 | | | ± 4.0 | |
| PLAQ. PLTS | 10 ⁹ /mm ³ ; 10 ⁹ /l | 63 | 63 | | | ± 20 | 220 | 220 | | | ± 30 | 440 | 440 | | | ± 50 | |
| VPM MPV | µm ³ fl | 10.4 | 10.4 | | | ± 2.0 | 9.9 | 9.9 | | | ± 2.0 | 11.0 | 11.0 | | | ± 2.0 | |
| NEUT | # | 1.64 | 1.64 | | | ± 0.35 | 4.61 | 4.61 | | | ± 0.90 | 13.58 | 13.58 | | | ± 1.90 | |
| | % | 51.1 | 51.1 | | | ± 10.0 | 54.2 | 54.2 | | | ± 10.0 | 73.8 | 73.8 | | | ± 10.0 | |
| LYMPHO | # | 1.21 | 1.21 | | | ± 0.33 | 3.17 | 3.17 | | | ± 0.70 | 2.87 | 2.87 | | | ± 1.50 | |
| | % | 37.7 | 37.7 | | | ± 12.0 | 37.3 | 37.3 | | | ± 8.0 | 15.6 | 15.6 | | | ± 8.0 | |
| MONO | # | 0.08 | 0.08 | | | ± 0.08 | 0.17 | 0.17 | | | ± 0.17 | 0.35 | 0.35 | | | ± 0.35 | |
| | % | 2.4 | 2.4 | | | ± 2.4 | 2.0 | 2.0 | | | ± 2.0 | 1.9 | 1.9 | | | ± 1.9 | |
| EOS | # | 0.19 | 0.19 | | | ± 0.19 | 0.30 | 0.30 | | | ± 0.30 | 0.83 | 0.83 | | | ± 0.83 | |
| | % | 6.0 | 6.0 | | | ± 6.0 | 3.5 | 3.5 | | | ± 3.5 | 4.5 | 4.5 | | | ± 4.5 | |
| BASO | # | 0.09 | 0.09 | | | ± 0.09 | 0.26 | 0.26 | | | ± 0.26 | 0.77 | 0.77 | | | ± 0.77 | |
| | % | 2.8 | 2.8 | | | ± 2.8 | 3.0 | 3.0 | | | ± 3.0 | 4.2 | 4.2 | | | ± 4.2 | |

LOT PX 437
Rev 2

CONTROL

 (Exp.) 2022-11-05
(YYYY - MM - DD)

| PARAMETRES PARAMETERS | | UNITES UNITS | ABX Lysebio | | | | | | | | | | | | | | | TOLERANCES TOLERANCE |
|--------------------------|------|--|--------------------|-------|--------|-------------------------|--------------------|-------|--------|-------------------------|--------------------|---------|--------|-------------------------|-------------------------|--|---------|-------------------------|
| | | | CONTROL | | | L | CONTROL | | | N | CONTROL | | | H | TOLERANCES TOLERANCE | | | |
| | | | PENTRA | | | TOLERANCES TOLERANCE | PENTRA | | | TOLERANCES TOLERANCE | PENTRA | | | TOLERANCES TOLERANCE | | | | |
| | | | 60 60C+ ES60 | MS60 | MSCRIP | | 60 60C+ ES60 | MS60 | MSCRIP | | 60 60C+ ES60 | MS60 | MSCRIP | | | | | |
| GB | WBC | 10 ⁹ /mm ³ ; 10 ⁹ /l | 3.2 | 3.2 | 3.2 | | ± 0.4 | 8.5 | 8.6 | 8.6 | | ± 1.0 | 18.9 | 18.8 | 18.4 | | ± 2.2 | |
| GR | RBC | 10 ⁶ /mm ³ ; 10 ¹² /l | 2.37 | 2.36 | 2.35 | | ± 0.16 | 4.61 | 4.61 | 4.60 | | ± 0.20 | 5.15 | 5.15 | 5.16 | | ± 0.25 | |
| | | g/dl | 6.1 | 6.0 | 6.0 | | ± 0.4 | 13.7 | 13.6 | 13.6 | | ± 0.5 | 16.9 | 16.9 | 16.8 | | ± 0.6 | |
| HB | HGB | g/l | 61 | 60 | 60 | | ± 4 | 137 | 136 | 136 | | ± 5 | 169 | 169 | 168 | | ± 6 | |
| | | mmol/l | 3.79 | 3.73 | 3.73 | | ± 0.25 | 8.51 | 8.45 | 8.45 | | ± 0.31 | 10.49 | 10.49 | 10.43 | | ± 0.37 | |
| HT | HCT | % | 17.8 | 17.7 | 17.4 | | ± 1.5 | 39.2 | 38.7 | 38.4 | | ± 2.0 | 47.9 | 47.4 | 47.2 | | ± 2.5 | |
| | | l/l | 0.178 | 0.177 | 0.174 | | ± 0.015 | 0.392 | 0.387 | 0.384 | | ± 0.020 | 0.479 | 0.474 | 0.472 | | ± 0.025 | |
| VGM | MCV | µm ³ ; fl | 75 | 75 | 74.0 | | ± 5 | 85 | 84 | 83.5 | | ± 5 | 93 | 92 | 91.5 | | ± 5 | |
| TGMH | MCH | pg | 25.7 | 25.4 | 25.5 | | ± 2.0 | 29.7 | 29.5 | 29.6 | | ± 2.0 | 32.8 | 32.8 | 32.6 | | ± 2.5 | |
| | | fmol | 1.60 | 1.58 | 1.59 | | ± 0.12 | 1.85 | 1.83 | 1.84 | | ± 0.12 | 2.04 | 2.04 | 2.02 | | ± 0.16 | |
| | | g/dl | 34.3 | 33.9 | 34.5 | | ± 3.0 | 35.0 | 35.1 | 35.4 | | ± 3.0 | 35.3 | 35.7 | 35.6 | | ± 3.0 | |
| CCMH | MCHC | g/l | 343 | 339 | 345 | | ± 30 | 350 | 351 | 354 | | ± 30 | 353 | 357 | 356 | | ± 30 | |
| | | mmol/l | 21.31 | 21.05 | 21.43 | | ± 1.86 | 21.71 | 21.81 | 21.99 | | ± 1.86 | 21.91 | 22.15 | 22.10 | | ± 1.86 | |
| IDR | RDW | % | 15.0 | 14.5 | 14.0 | | ± 4.0 | 13.0 | 12.5 | 12.0 | | ± 4.0 | 12.0 | 11.5 | 11.0 | | ± 4.0 | |
| PLAQ. | PLTS | 10 ³ /mm ³ ; 10 ⁹ /l | 68 | 64 | 66 | | ± 20 | 228 | 226 | 216 | | ± 30 | 457 | 463 | 438 | | ± 50 | |
| VPM | MPV | µm ³ ; fl | 9.8 | 10.0 | 9.3 | | ± 2.0 | 9.3 | 9.3 | 8.7 | | ± 2.0 | 10.1 | 10.1 | 9.3 | | ± 2.0 | |
| | | # | 1.64 | 1.63 | 1.70 | | ± 0.35 | 4.53 | 4.57 | 4.69 | | ± 0.90 | 13.82 | 13.82 | 13.71 | | ± 1.90 | |
| | | % | 51.2 | 50.8 | 53.0 | | ± 10.0 | 53.3 | 53.1 | 54.5 | | ± 10.0 | 73.1 | 73.5 | 74.5 | | ± 10.0 | |
| | | # | 1.22 | 1.17 | 1.18 | | ± 0.33 | 3.24 | 3.23 | 3.23 | | ± 0.70 | 3.02 | 2.97 | 2.85 | | ± 1.50 | |
| | | % | 38.0 | 36.7 | 37.0 | | ± 12.0 | 38.1 | 37.6 | 37.5 | | ± 8.0 | 16.0 | 15.8 | 15.5 | | ± 8.0 | |
| | | # | 0.09 | 0.09 | 0.08 | | ± 0.08 | 0.19 | 0.17 | 0.17 | | ± 0.17 | 0.40 | 0.34 | 0.28 | | ± 0.28 | |
| | | % | 2.8 | 2.7 | 2.5 | | ± 2.5 | 2.2 | 2.0 | 2.0 | | ± 2.0 | 2.1 | 1.8 | 1.5 | | ± 1.5 | |
| | | # | 0.17 | 0.22 | 0.14 | | ± 0.14 | 0.29 | 0.37 | 0.26 | | ± 0.26 | 0.87 | 0.86 | 0.83 | | ± 0.83 | |
| | | % | 5.2 | 7.0 | 4.5 | | ± 4.5 | 3.4 | 4.3 | 3.0 | | ± 3.0 | 4.6 | 4.6 | 4.5 | | ± 4.5 | |
| | | # | 0.09 | 0.09 | 0.10 | | ± 0.09 | 0.26 | 0.26 | 0.26 | | ± 0.26 | 0.79 | 0.81 | 0.74 | | ± 0.74 | |
| | | % | 2.8 | 2.8 | 3.0 | | ± 2.8 | 3.0 | 3.0 | 3.0 | | ± 3.0 | 4.2 | 4.3 | 4.0 | | ± 4.0 | |

1300090306
FRONT / RECTO
Ref: TEMP-0821 Rev.47