

ABX Minotrol CRP



LOT

MC 424

CONTROL



(Exp.) 2022-10-05
(YYYY-MM-DD)

Rev 1

| PARAMETRES PARAMETERS | | UNITES UNITS | | ABX Lysebio | | | | | | | | | | | | | | | | |
|--------------------------|------|---|--------------|--------------|---------|--------------|--------|-------------------------|--------------|--------------|---------|--------------|-------------------------|---------|--------------|--------------|---------|-------------------------|-------------------------|---------|
| | | | | CONTROL | | | | 1 | CONTROL | | | | 2 | CONTROL | | | | 3 | TOLERANCES TOLERANCE | |
| | | | | MICROS | | PENTRA | | TOLERANCES TOLERANCE | MICROS | | PENTRA | | TOLERANCES TOLERANCE | MICROS | | PENTRA | | TOLERANCES TOLERANCE | | |
| | | | | EMI CRP | LC-767G | | MS CRP | | | EMI CRP | LC-767G | | | MS CRP | | EMI CRP | LC-767G | | | |
| GB | WBC | 10 ³ /mm ³ ;10 ⁹ /l | 2.1 | 2.1 | | 2.1 | | ± 0.4 | 7.6 | 7.7 | | 7.7 | | ± 1.0 | 20.0 | 20.4 | | 20.6 | | ± 2.2 |
| GR | RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 2.26 | 2.28 | | 2.22 | | ± 0.15 | 4.46 | 4.48 | | 4.47 | | ± 0.18 | 5.50 | 5.53 | | 5.59 | | ± 0.25 |
| HB | HGB | g/dl | 6.0 | 6.0 | | 5.7 | | ± 0.4 | 13.6 | 13.5 | | 13.5 | | ± 0.5 | 17.6 | 17.9 | | 18.1 | | ± 0.7 |
| | | g/l | 60 | 60 | | 57 | | ± 4 | 136 | 135 | | 135 | | ± 5 | 176 | 179 | | 181 | | ± 7 |
| | | mmol/l | 3.73 | 3.73 | | 3.54 | | ± 0.25 | 8.45 | 8.38 | | 8.38 | | ± 0.31 | 10.93 | 11.12 | | 11.24 | | ± 0.44 |
| HT | HCT | % | 16.3 | 16.4 | | 15.8 | | ± 2.0 | 37.0 | 37.6 | | 36.7 | | ± 2.5 | 49.0 | 49.8 | | 49.2 | | ± 3.0 |
| | | l/l | 0.163 | 0.164 | | 0.158 | | ± 0.020 | 0.370 | 0.376 | | 0.367 | | ± 0.025 | 0.490 | 0.498 | | 0.492 | | ± 0.030 |
| VGM | MCV | µm ³ ; fl | 72 | 72 | | 71.0 | | ± 4 | 83 | 84 | | 82.0 | | ± 4 | 89 | 90 | | 88.0 | | ± 4 |
| TGMH | MCH | pg | 26.5 | 26.3 | | 25.7 | | ± 2.4 | 30.5 | 30.1 | | 30.2 | | ± 2.4 | 32.0 | 32.4 | | 32.4 | | ± 2.8 |
| | | fmol | 1.65 | 1.63 | | 1.59 | | ± 0.15 | 1.89 | 1.87 | | 1.88 | | ± 0.15 | 1.99 | 2.01 | | 2.01 | | ± 0.18 |
| CCMH | MCHC | g/dl | 36.9 | 36.5 | | 36.2 | | ± 3.0 | 36.7 | 35.9 | | 36.8 | | ± 3.0 | 36.0 | 36.0 | | 36.8 | | ± 3.0 |
| | | g/l | 369 | 365 | | 362 | | ± 30 | 367 | 359 | | 368 | | ± 30 | 360 | 360 | | 368 | | ± 30 |
| | | mmol/l | 22.90 | 22.70 | | 22.46 | | ± 1.86 | 22.81 | 22.28 | | 22.87 | | ± 1.86 | 22.33 | 22.33 | | 22.85 | | ± 1.86 |
| IDR | RDW | % | 13.5 | 13.0 | | 15.0 | | ± 3.0 | 12.5 | 12.5 | | 12.5 | | ± 3.0 | 12.0 | 12.0 | | 12.0 | | ± 3.0 |
| PLAQ. | PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 77 | 79 | | 68 | | ± 20 | 271 | 263 | | 258 | | ± 40 | 490 | 492 | | 500 | | ± 60 |
| VPM | MPV | µm ³ ; fl | 7.7 | 8.4 | | 9.0 | | ± 2.0 | 8.3 | 8.3 | | 9.2 | | ± 2.0 | 8.1 | 8.3 | | 8.7 | | ± 2.0 |
| LYMPHO | | # | 1.3 | 1.3 | | N/A | | ± 0.2 | 2.7 | 2.7 | | N/A | | ± 0.5 | 3.4 | 3.3 | | N/A | | ± 1.1 |
| | | % | 62.0 | 61.5 | | N/A | | ± 8.0 | 35.0 | 34.5 | | N/A | | ± 6.0 | 17.0 | 16.0 | | N/A | | ± 5.0 |
| MONO | | # | 0.2 | 0.2 | | N/A | | ± 0.2 | 0.6 | 0.6 | | N/A | | ± 0.4 | 1.3 | 1.2 | | N/A | | ± 1.0 |
| | | % | 9.0 | 9.0 | | N/A | | ± 7.0 | 8.0 | 8.0 | | N/A | | ± 5.0 | 6.5 | 6.0 | | N/A | | ± 5.0 |
| GRANULO | | # | 0.6 | 0.6 | | N/A | | ± 0.2 | 4.3 | 4.4 | | N/A | | ± 0.5 | 15.3 | 15.9 | | N/A | | ± 1.5 |
| | | % | 29.0 | 29.5 | | N/A | | ± 7.0 | 57.0 | 57.5 | | N/A | | ± 6.0 | 76.5 | 78.0 | | N/A | | ± 7.0 |
| CRP | | mg/dl | 0.46 | 0.51 | | 0.49 | | ± 0.40 | 1.88 | 2.07 | | 1.96 | | ± 0.80 | 3.34 | 3.66 | | 3.57 | | ± 1.50 |
| | | mg/l | 4.6 | 5.1 | | 4.9 | | ± 4.0 | 18.8 | 20.7 | | 19.6 | | ± 8.0 | 33.4 | 36.6 | | 35.7 | | ± 15.0 |

Ref: TEMP-1070 Rev.42 BACK/VERSO 1300092170