

ABX Minotrol CRP



LOT

MC 425

CONTROL



(Exp.) 2023-01-05
(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE	
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				
				EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G			MS CRP		EMI CRP	LC-767G			MS CRP
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.1	2.1		2.2		± 0.4	7.9	7.9		8.2		± 1.0	20.8	21.6		21.8		± 2.2
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.27	2.26		2.24		± 0.15	4.59	4.59		4.60		± 0.18	5.39	5.43		5.49		± 0.25
HB	HGB	g/dl	6.2	6.2		5.9		± 0.4	13.7	13.6		13.6		± 0.5	18.0	18.1		18.3		± 0.7
		g/l	62	62		59		± 4	137	136		136		± 5	180	181		183		± 7
		mmol/l	3.85	3.85		3.66		± 0.25	8.51	8.45		8.45		± 0.31	11.18	11.24		11.36		± 0.44
HT	HCT	%	16.3	16.5		16.4		± 2.0	37.6	37.2		36.8		± 2.5	49.6	49.4		48.9		± 3.0
		l/l	0.163	0.165		0.164		± 0.020	0.376	0.372		0.368		± 0.025	0.496	0.494		0.489		± 0.030
VGM	MCV	µm ³ ; fl	72	73		73.0		± 4	82	81		80.0		± 4	92	91		89.0		± 4
TGMH	MCH	pg	27.3	27.4		26.3		± 2.4	29.8	29.6		29.6		± 2.4	33.4	33.3		33.3		± 2.8
		fmol	1.70	1.70		1.64		± 0.15	1.85	1.84		1.84		± 0.15	2.07	2.07		2.07		± 0.18
CCMH	MCHC	g/dl	37.9	37.6		36.1		± 3.0	36.4	36.6		37.0		± 3.0	36.3	36.6		37.5		± 3.0
		g/l	379	376		361		± 30	364	366		370		± 30	363	366		375		± 30
		mmol/l	23.56	23.34		22.41		± 1.86	22.60	22.72		22.95		± 1.86	22.54	22.75		23.26		± 1.86
IDR	RDW	%	13.5	13.0		14.5		± 3.0	13.0	12.0		13.0		± 3.0	12.5	12.0		12.5		± 3.0
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	76	79		68		± 20	271	271		263		± 40	511	514		506		± 60
VPM	MPV	µm ³ ; fl	8.1	8.9		9.4		± 2.0	8.4	8.2		8.9		± 2.0	8.7	8.6		8.9		± 2.0
LYMPHO		#	1.3	1.3		N/A		± 0.2	2.7	2.7		N/A		± 0.5	3.9	3.8		N/A		± 1.1
		%	64.0	63.0		N/A		± 8.0	34.5	34.0		N/A		± 6.0	18.5	17.5		N/A		± 5.0
MONO		#	0.2	0.2		N/A		± 0.2	0.6	0.6		N/A		± 0.4	1.4	1.4		N/A		± 1.0
		%	8.5	8.5		N/A		± 7.0	7.0	7.0		N/A		± 5.0	6.5	6.5		N/A		± 5.0
GRANULO		#	0.6	0.6		N/A		± 0.2	4.6	4.7		N/A		± 0.5	15.7	16.4		N/A		± 1.5
		%	27.5	28.5		N/A		± 7.0	58.5	59.0		N/A		± 6.0	75.0	76.0		N/A		± 7.0
CRP		mg/dl	0.54	0.57		0.54		± 0.40	2.77	2.95		2.48		± 0.80	4.52	4.67		4.63		± 1.50
		mg/l	5.4	5.7		5.4		± 4.0	27.7	29.5		24.8		± 8.0	45.2	46.7		46.3		± 15.0

Ref: TEMP-1070 Rev.42 BACKVERSO 1300092170