

# ABX Minotrol CRP



**LOT**

**MC 426**

**CONTROL**



**(Exp.) 2023-04-05**

(YYYY-MM-DD)

Rev 1

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio																
				CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE	
				MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA		TOLERANCES TOLERANCE	MICROS		PENTRA				
				EMI CRP	LC-767G		MS CRP			EMI CRP	LC-767G			MS CRP		EMI CRP	LC-767G			MS CRP
GB	WBC	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>2.0</b>	<b>2.0</b>		<b>2.1</b>		± 0.4	<b>7.7</b>	<b>7.7</b>		<b>8.0</b>		± 1.0	<b>20.3</b>	<b>20.8</b>		<b>21.6</b>		± 2.2
GR	RBC	10 <sup>6</sup> /mm <sup>3</sup> ;10 <sup>12</sup> /l	<b>2.35</b>	<b>2.37</b>		<b>2.35</b>		± 0.15	<b>4.45</b>	<b>4.46</b>		<b>4.50</b>		± 0.18	<b>5.37</b>	<b>5.38</b>		<b>5.49</b>		± 0.25
HB	HGB	g/dl	<b>6.3</b>	<b>6.3</b>		<b>6.0</b>		± 0.4	<b>13.3</b>	<b>13.4</b>		<b>13.4</b>		± 0.5	<b>17.3</b>	<b>17.4</b>		<b>17.8</b>		± 0.7
		g/l	<b>63</b>	<b>63</b>		<b>60</b>		± 4	<b>133</b>	<b>134</b>		<b>134</b>		± 5	<b>173</b>	<b>174</b>		<b>178</b>		± 7
		mmol/l	<b>3.91</b>	<b>3.91</b>		<b>3.73</b>		± 0.25	<b>8.26</b>	<b>8.32</b>		<b>8.32</b>		± 0.31	<b>10.74</b>	<b>10.81</b>		<b>11.05</b>		± 0.44
HT	HCT	%	<b>16.9</b>	<b>17.1</b>		<b>16.8</b>		± 2.0	<b>36.5</b>	<b>36.6</b>		<b>36.5</b>		± 2.5	<b>47.8</b>	<b>47.9</b>		<b>48.3</b>		± 3.0
		l/l	<b>0.169</b>	<b>0.171</b>		<b>0.168</b>		± 0.020	<b>0.365</b>	<b>0.366</b>		<b>0.365</b>		± 0.025	<b>0.478</b>	<b>0.479</b>		<b>0.483</b>		± 0.030
VGM	MCV	µm <sup>3</sup> ; fl	<b>72</b>	<b>72</b>		<b>71.5</b>		± 4	<b>82</b>	<b>82</b>		<b>81.0</b>		± 4	<b>89</b>	<b>89</b>		<b>88.0</b>		± 4
TGMH	MCH	pg	<b>26.8</b>	<b>26.6</b>		<b>25.5</b>		± 2.4	<b>29.9</b>	<b>30.0</b>		<b>29.8</b>		± 2.4	<b>32.2</b>	<b>32.3</b>		<b>32.4</b>		± 2.8
		fmol	<b>1.66</b>	<b>1.65</b>		<b>1.59</b>		± 0.15	<b>1.86</b>	<b>1.87</b>		<b>1.85</b>		± 0.15	<b>2.00</b>	<b>2.01</b>		<b>2.01</b>		± 0.18
CCMH	MCHC	g/dl	<b>37.2</b>	<b>36.9</b>		<b>35.7</b>		± 3.0	<b>36.4</b>	<b>36.6</b>		<b>36.8</b>		± 3.0	<b>36.2</b>	<b>36.3</b>		<b>36.8</b>		± 3.0
		g/l	<b>372</b>	<b>369</b>		<b>357</b>		± 30	<b>364</b>	<b>366</b>		<b>368</b>		± 30	<b>362</b>	<b>363</b>		<b>368</b>		± 30
		mmol/l	<b>23.12</b>	<b>22.93</b>		<b>22.18</b>		± 1.86	<b>22.63</b>	<b>22.75</b>		<b>22.83</b>		± 1.86	<b>22.48</b>	<b>22.57</b>		<b>22.88</b>		± 1.86
IDR	RDW	%	<b>14.0</b>	<b>13.5</b>		<b>15.0</b>		± 3.0	<b>13.0</b>	<b>13.0</b>		<b>13.0</b>		± 3.0	<b>12.5</b>	<b>12.5</b>		<b>12.5</b>		± 3.0
PLAQ.	PLTS	10 <sup>3</sup> /mm <sup>3</sup> ;10 <sup>9</sup> /l	<b>72</b>	<b>71</b>		<b>64</b>		± 20	<b>270</b>	<b>270</b>		<b>267</b>		± 40	<b>513</b>	<b>511</b>		<b>522</b>		± 60
VPM	MPV	µm <sup>3</sup> ; fl	<b>8.0</b>	<b>8.8</b>		<b>9.7</b>		± 2.0	<b>8.4</b>	<b>8.6</b>		<b>9.1</b>		± 2.0	<b>8.6</b>	<b>8.6</b>		<b>9.1</b>		± 2.0
LYMPHO		#	<b>1.3</b>	<b>1.3</b>		N/A		± 0.2	<b>2.6</b>	<b>2.5</b>		N/A		± 0.5	<b>3.5</b>	<b>3.3</b>		N/A		± 1.1
		%	<b>63.5</b>	<b>63.0</b>		N/A		± 8.0	<b>33.5</b>	<b>33.0</b>		N/A		± 6.0	<b>17.0</b>	<b>16.0</b>		N/A		± 5.0
MONO		#	<b>0.1</b>	<b>0.2</b>		N/A		± 0.2	<b>0.5</b>	<b>0.6</b>		N/A		± 0.4	<b>1.2</b>	<b>1.2</b>		N/A		± 1.0
		%	<b>7.0</b>	<b>7.5</b>		N/A		± 7.0	<b>7.0</b>	<b>7.5</b>		N/A		± 5.0	<b>6.0</b>	<b>6.0</b>		N/A		± 5.0
GRANULO		#	<b>0.6</b>	<b>0.6</b>		N/A		± 0.2	<b>4.6</b>	<b>4.6</b>		N/A		± 0.5	<b>15.6</b>	<b>16.2</b>		N/A		± 1.5
		%	<b>29.5</b>	<b>29.5</b>		N/A		± 7.0	<b>59.5</b>	<b>59.5</b>		N/A		± 6.0	<b>77.0</b>	<b>78.0</b>		N/A		± 7.0
CRP		mg/dl	<b>0.47</b>	<b>0.53</b>		<b>0.52</b>		± 0.40	<b>2.17</b>	<b>2.32</b>		<b>2.13</b>		± 0.80	<b>4.52</b>	<b>4.72</b>		<b>4.61</b>		± 1.50
		mg/l	<b>4.7</b>	<b>5.3</b>		<b>5.2</b>		± 4.0	<b>21.7</b>	<b>23.2</b>		<b>21.3</b>		± 8.0	<b>45.2</b>	<b>47.2</b>		<b>46.1</b>		± 15.0

Ref: TEMP-1070 Rev.42 BACK/VERSO 1300092170