

LOT PX 436
Rev 1

CONTROL

 (Exp.) **2022-09-05**
(YYYY - MM - DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE			
		CONTROL				L	TOLERANCES TOLERANCE	CONTROL				N	TOLERANCES TOLERANCE	CONTROL				H	TOLERANCES TOLERANCE	
		PENTRA						PENTRA						PENTRA						
		XL80	XLR			XL80	XLR			XL80	XLR			XL80	XLR					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	3.0			± 0.4	8.6	8.6			± 1.0	18.7	18.7			± 2.2				
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.33	2.33			± 0.16	4.64	4.64			± 0.20	5.09	5.09			± 0.25				
HB HGB	g/dl	6.2	6.2			± 0.4	13.6	13.6			± 0.5	16.3	16.3			± 0.6				
	g/l	62	62			± 4	136	136			± 5	163	163			± 6				
	mmol/l	3.85	3.85			± 0.25	8.45	8.45			± 0.31	10.12	10.12			± 0.37				
HT HCT	%	18.9	18.9			± 1.5	39.4	39.4			± 2.0	46.3	46.3			± 2.5				
	l/l	0.189	0.189			± 0.015	0.394	0.394			± 0.020	0.463	0.463			± 0.025				
VGM MCV	µm ³ ; fl	81	81			± 5	85	85			± 5	91	91			± 5				
TGMH MCH	pg	26.6	26.6			± 2.0	29.3	29.3			± 2.0	32.0	32.0			± 2.5				
	fmol	1.65	1.65			± 0.12	1.82	1.82			± 0.12	1.99	1.99			± 0.16				
CCMH MCHC	g/dl	32.9	32.9			± 3.0	34.5	34.5			± 3.0	35.2	35.2			± 3.0				
	g/l	329	329			± 30	345	345			± 30	352	352			± 30				
	mmol/l	20.40	20.40			± 1.86	21.41	21.41			± 1.86	21.85	21.85			± 1.86				
IDR RDW	%	15.0	15.0			± 4.0	13.0	13.0			± 4.0	12.5	12.5			± 4.0				
PLAQ. PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	69	69			± 20	220	220			± 30	449	449			± 50				
VPM MPV	µm ³ ; fl	9.8	9.8			± 2.0	12.0	12.0			± 2.0	11.1	11.1			± 2.0				
NEUT	#	1.43	1.43			± 0.35	4.34	4.34			± 0.90	13.13	13.13			± 1.90				
	%	47.5	47.5			± 10.0	50.5	50.5			± 10.0	70.2	70.2			± 10.0				
LYMPHO	#	1.10	1.10			± 0.33	3.40	3.40			± 0.70	3.48	3.48			± 1.50				
	%	36.7	36.7			± 12.0	39.5	39.5			± 8.0	18.6	18.6			± 8.0				
MONO	#	0.19	0.19			± 0.19	0.30	0.30			± 0.30	0.43	0.43			± 0.43				
	%	6.4	6.4			± 6.4	3.5	3.5			± 3.5	2.3	2.3			± 2.3				
EOS	#	0.20	0.20			± 0.20	0.32	0.32			± 0.32	0.92	0.92			± 0.92				
	%	6.8	6.8			± 6.8	3.7	3.7			± 3.7	4.9	4.9			± 4.9				
BASO	#	0.08	0.08			± 0.08	0.24	0.24			± 0.24	0.75	0.75			± 0.75				
	%	2.6	2.6			± 2.6	2.8	2.8			± 2.8	4.0	4.0			± 4.0				

LOT PX 436
Rev 1

CONTROL

 (Exp.) **2022-09-05**
(YYYY-MM-DD)

PARAMETRES PARAMETERS		UNITES UNITS	ABX Lysebio															TOLERANCES TOLERANCE	
			CONTROL			L	CONTROL			N	TOLERANCES TOLERANCE	CONTROL			H	TOLERANCES TOLERANCE			
			PENTRA				PENTRA					PENTRA							
			60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP		60 60C+ ES60	MS60	MSCRIP						
GB	WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	3.0	2.9	3.0		± 0.4	8.7	8.7	8.6			± 1.0	19.1	18.7	18.5			± 2.2
GR	RBC	10 ⁹ /mm ³ ; 10 ¹² /l	2.41	2.38	2.34		± 0.16	4.68	4.63	4.60			± 0.20	5.11	5.11	5.01			± 0.25
HB	HGB	g/dl	6.3	6.2	6.2		± 0.4	13.6	13.5	13.5			± 0.5	16.3	16.3	16.1			± 0.6
		g/l	63	62	62		± 4	136	135	135			± 5	163	163	161			± 6
HT	HCT	mmol/l	3.91	3.85	3.85		± 0.25	8.45	8.38	8.38			± 0.31	10.12	10.12	10.00			± 0.37
		%	18.6	18.3	18.1		± 1.5	39.3	38.4	38.6			± 2.0	46.5	45.5	45.1			± 2.5
VGM	MCV	l/l	0.186	0.183	0.181		± 0.015	0.393	0.384	0.386			± 0.020	0.465	0.455	0.451			± 0.025
		µm ³ ; fl	77	77	77.5		± 5	84	83	84.0			± 5	91	90	90.0			± 5
TGMH	MCH	pg	26.1	26.1	26.5		± 2.0	29.1	29.2	29.3			± 2.0	31.9	31.9	32.1			± 2.5
		fmol	1.62	1.62	1.65		± 0.12	1.80	1.81	1.82			± 0.12	1.98	1.98	2.00			± 0.16
CCMH	MCHC	g/dl	33.9	33.8	34.2		± 3.0	34.6	35.1	34.9			± 3.0	35.1	35.4	35.7			± 3.0
		g/l	339	338	342		± 30	346	351	349			± 30	351	354	357			± 30
		mmol/l	21.08	21.01	21.23		± 1.86	21.48	21.82	21.70			± 1.86	21.77	22.01	22.17			± 1.86
IDR	RDW	%	15.0	15.0	14.0		± 4.0	13.0	12.5	12.0			± 4.0	12.0	12.0	11.0			± 4.0
PLAQ.	PLTS	10 ⁹ /mm ³ ; 10 ⁹ /l	75	72	67		± 20	234	231	212			± 30	472	471	432			± 50
VPM	MPV	µm ³ ; fl	9.5	9.6	8.8		± 2.0	11.3	11.4	10.4			± 2.0	10.4	10.4	9.6			± 2.0
NEUT		#	1.41	1.33	1.41		± 0.35	4.25	4.21	4.26			± 0.90	13.24	12.90	12.95			± 1.90
		%	46.9	45.9	47.0		± 10.0	48.8	48.4	49.5			± 10.0	69.3	69.0	70.0			± 10.0
LYMPHO		#	1.10	1.10	1.14		± 0.33	3.56	3.58	3.53			± 0.70	3.63	3.59	3.52			± 1.50
		%	36.7	37.8	38.0		± 12.0	40.9	41.1	41.0			± 8.0	19.0	19.2	19.0			± 8.0
MONO		#	0.23	0.22	0.21		± 0.21	0.34	0.37	0.30			± 0.30	0.48	0.52	0.37			± 0.37
		%	7.7	7.7	7.0		± 7.0	3.9	4.3	3.5			± 3.5	2.5	2.8	2.0			± 2.0
EOS		#	0.18	0.17	0.15		± 0.15	0.32	0.30	0.26			± 0.26	0.99	0.94	0.93			± 0.93
		%	6.1	6.0	5.0		± 5.0	3.7	3.5	3.0			± 3.0	5.2	5.0	5.0			± 5.0
BASO		#	0.08	0.08	0.09		± 0.08	0.23	0.23	0.26			± 0.23	0.76	0.75	0.74			± 0.74
		%	2.6	2.6	3.0		± 2.6	2.7	2.7	3.0			± 2.7	4.0	4.0	4.0			± 4.0

Ref: TEMP-0821 Rev.46 FRONT / RECTO 1300090306